

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE041221\
 Data File : BE101285.D
 Acq On : 12 Apr 2021 17:27
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE041221

Quant Time: Apr 12 18:30:14 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE041221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 12 18:02:55 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	92	0.00
2	1,4-Dioxane	0.400	0.427	-6.7	93	0.00
3	n-Nitrosodimethylamine	0.400	0.397	0.8	86	0.00
4 S	2-Fluorophenol	0.400	0.398	0.5	90	0.00
5 S	Phenol-d6	0.400	0.386	3.5	89	0.00
6	bis(2-Chloroethyl)ether	0.400	0.407	-1.7	88	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	88	0.00
8 S	Nitrobenzene-d5	0.400	0.391	2.3	89	0.00
9	Naphthalene	0.400	0.410	-2.5	88	0.00
10	Hexachlorobutadiene	0.400	0.420	-5.0	90	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.392	2.0	86	0.00
12	2-Methylnaphthalene	0.400	0.385	3.8	84	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	79	0.00
14 S	2,4,6-Tribromophenol	0.400	0.286	28.5#	68	0.00
15 S	2-Fluorobiphenyl	0.400	0.424	-6.0	85	0.00
16	Acenaphthylene	0.400	0.376	6.0	81	0.00
17	Acenaphthene	0.400	0.399	0.3	81	0.00
18	Fluorene	0.400	0.371	7.3	75	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	73	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.333	16.8	61	0.00
21	4-Bromophenyl-phenylether	0.400	0.398	0.5	74	0.00
22	Hexachlorobenzene	0.400	0.411	-2.7	77	0.00
23	Atrazine	0.400	0.335	16.3	68	0.00
24	Pentachlorophenol	0.400	0.333	16.8	32	0.00
25	Phenanthrene	0.400	0.398	0.5	73	0.00
26	Anthracene	0.400	0.372	7.0	71	0.00
27 SURR	Fluoranthene-d10	0.400	0.360	10.0	66	0.00
28	Fluoranthene	0.400	0.358	10.5	65	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	78	0.00
30	Pyrene	0.400	0.355	11.3	66	0.00
31 S	Terphenyl-d14	0.400	0.355	11.3	65	0.00
32	Benzo(a)anthracene	0.400	0.386	3.5	78	0.00
33	Chrysene	0.400	0.403	-0.8	77	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.371	7.3	83	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.461	-15.3	98	0.00
36 I	Perylene-d12	0.400	0.400	0.0	105	0.00
37	Benzo(b)fluoranthene	0.400	0.367	8.3	100	0.00
38	Benzo(k)fluoranthene	0.400	0.392	2.0	107	0.00
39 C	Benzo(a)pyrene	0.400	0.377	5.8	103	0.00
40	Dibenzo(a,h)anthracene	0.400	0.395	1.3	100	0.00
41	Benzo(g,h,i)perylene	0.400	0.392	2.0	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_E

ClientSampleId :

ICVBE041221