

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE041316\
 Data File : BE091654.D
 Acq On : 13 Apr 2016 13:08
 Operator : UM/SJ
 Sample : H2090-05
 Misc : Four Times Internal standard added
 ALS Vial : 5 Sample Multiplier: 4

Instrument :
 BNA_E
 ClientSampleId :
 KY038PS027-160330

Quant Time: Apr 14 04:59:32 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE040816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 08 16:48:14 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.80	152	130038	5.00	ng	-0.02
7) Naphthalene-d8	10.57	136	615873	5.00	ng	-0.02
13) Acenaphthene-d10	14.42	164	373227	5.00	ng	0.00
19) Phenanthrene-d10	17.15	188	994314	5.00	ng	-0.01
26) Chrysene-d12	21.31	240	1126505	5.00	ng	0.00
35) Perylene-d12	23.76	264	1047337	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.41	112	15752	2.03	ng	0.00
5) Phenol-d6	6.97	99	12855	1.24	ng	0.00
8) Nitrobenzene-d5	8.96	82	28464	3.18	ng	0.00
14) 2,4,6-Tribromophenol	15.90	330	28936	9.18	ng	-0.01
15) 2-Fluorobiphenyl	13.05	172	102428	3.95	ng	0.00
29) Terphenyl-d14	19.76	244	210830	6.00	ng	0.00
Target Compounds						
24) Anthracene	17.29	178	1817	0.04	ng	Qvalue 97
33) Bis(2-ethylhexyl)phthalate	21.22	149	4590	1.15	ng	# 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE041316\
Data File : BE091654.D
Acq On : 13 Apr 2016 13:08
Operator : UM/SJ
Sample : H2090-05
Misc : Four Times Internal standard added
ALS Vial : 5 Sample Multiplier: 4

Instrument :
BNA_E
ClientSampleId :
KY038PS027-160330

Quant Time: Apr 14 04:59:32 2016
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE040816.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 08 16:48:14 2016
Response via : Initial Calibration

