

Data Path : Z:\HPCHEM1\BNA E\Data\BE041316\
 Data File : BE091657.D
 Acq On : 13 Apr 2016 14:59
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 14 01:30:54 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE040816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 08 16:48:14 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	-0.02
2	1,4-Dioxane	0.640	0.654	-2.2	84	0.00
3	n-Nitrosodimethylamine	0.767	0.679	11.5	79	0.00
4 S	2-Fluorophenol	1.193	1.058	11.3	78	0.00
5 S	Phenol-d6	1.591	1.375	13.6	79	0.00
6	bis(2-Chloroethyl)ether	1.379	1.294	6.2	83	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	85	-0.02
8 S	Nitrobenzene-d5	0.290	0.240	17.2	79	0.00
9	Nitrobenzene	0.305	0.242	20.7	77	0.00
10	Naphthalene	1.031	1.032	-0.1	86	0.00
11	Hexachlorobutadiene	0.152	0.150	1.3	84	0.00
12	2-Methylnaphthalene	0.658	0.640	2.7	85	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	83	0.00
14 S	2,4,6-Tribromophenol	0.138	0.111	19.6	83	0.00
15 S	2-Fluorobiphenyl	1.390	1.407	-1.2	85	0.00
16	Acenaphthylene	1.941	1.820	6.2	96	0.00
17	Acenaphthene	1.227	1.268	-3.3	89	-0.01
18	Fluorene	1.533	1.495	2.5	84	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	90	-0.01
20	4-Bromophenyl-phenylether	0.176	0.152	13.6	81	0.00
21	Hexachlorobenzene	0.183	0.169	7.7	84	0.00
22	Pentachlorophenol	0.068	0.040	41.2#	76	0.00
23	Phenanthrene	1.141	1.064	6.7	84	0.00
24	Anthracene	1.011	0.874	13.6	83	0.00
25	Fluoranthene	1.208	0.987	18.3	79	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	77	0.00
27	Benzydine	0.329	0.199	39.5#	68	0.00
28	Pvrene	1.000	1.002	-0.2	83	-0.02
29 S	Terphenyl-d14	0.624	0.703	-12.7	88	0.00
30	Benzo(a)anthracene	1.119	1.150	-2.8	82	0.00
31	3,3'-Dichlorobenzidine	0.599	0.396	33.9#	61	0.00
32	Chrysene	1.015	1.240	-22.2	96	-0.02
33	Bis(2-ethylhexyl)phthalate	0.494	0.177	64.2#	42#	0.00
34	Indeno(1,2,3-cd)pyrene	1.132	1.023	9.6	72	-0.02
35 I	Perylene-d12	1.000	1.000	0.0	79	-0.01
36	Benzo(b)fluoranthene	1.161	1.072	7.7	76	-0.01
37	Benzo(k)fluoranthene	1.145	1.127	1.6	81	-0.01
38 C	Benzo(a)pyrene	1.079	0.939	13.0	73	-0.01
39	Dibenzo(a,h)anthracene	1.062	0.929	12.5	72	-0.01
40	Benzo(a,h,i)perylene	1.094	0.995	9.0	74	-0.01

Data Path : Z:\HPCHEM1\BNA E\Data\BE041316\
Data File : BE091657.D
Acq On : 13 Apr 2016 14:59
Operator : UM/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 14 01:30:54 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE040816.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 08 16:48:14 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				