

Data Path : Z:\HPCHEM1\BNA_E\Data\BE041318\
 Data File : BE095997.D
 Acq On : 13 Apr 2018 15:39
 Operator : SJ/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Apr 13 16:48:18 2018
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE041118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 13 14:42:08 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	120	-0.01
2	1,4-Dioxane	0.618	0.607	1.8	127	0.00
3	n-Nitrosodimethylamine	0.757	0.712	5.9	122	0.00
4 S	2-Fluorophenol	1.237	1.176	4.9	120	0.00
5 S	Phenol-d6	1.707	1.619	5.2	119	0.00
6	bis(2-Chloroethyl)ether	1.502	1.438	4.3	120	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	118	0.00
8 S	Nitrobenzene-d5	0.469	0.433	7.7	119	-0.01
9	Naphthalene	4.217	3.989	5.4	119	0.00
10	Hexachlorobutadiene	0.187	0.108	42.2#	76	0.00
11 SURR	2-Methylnaphthalene-d10	0.634	0.605	4.6	121	0.00
12	2-Methylnaphthalene	0.698	0.661	5.3	121	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	123	0.00
14 S	2,4,6-Tribromophenol	0.193	0.182	5.7	127	0.00
15 S	2-Fluorobiphenyl	1.711	1.598	6.6	120	-0.01
16	Acenaphthylene	2.136	1.954	8.5	121	0.00
17	Acenaphthene	1.284	1.162	9.5	120	-0.01
18	Fluorene	1.588	1.458	8.2	121	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	123	0.00
20	4-Bromophenyl-phenylether	0.238	0.214	10.1	118	0.00
21	Hexachlorobenzene	0.237	0.218	8.0	117	0.00
22	Pentachlorophenol	0.078	0.072	7.7	140	0.00
23	Phenanthrene	1.124	1.044	7.1	121	0.00
24	Anthracene	1.074	0.995	7.4	124	-0.01
25 SURR	Fluoranthene-d10	5.430	4.905	9.7	120	0.00
26	Fluoranthene	1.477	1.330	10.0	120	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	114	-0.01
28	Pyrene	0.788	0.438	44.4#	71	0.00
29 S	Terphenyl-d14	0.884	0.855	3.3	121	0.00
30	Benzo(a)anthracene	1.283	1.177	8.3	115	0.00
31	Chrysene	1.239	1.177	5.0	115	-0.01
32	Bis(2-ethylhexyl)phthalate	3.311	3.015	8.9	112	0.00
33	Indeno(1,2,3-cd)pyrene	1.201	1.147	4.5	116	0.00
34 I	Perylene-d12	1.000	1.000	0.0	112	0.00
35	Benzo(b)fluoranthene	1.200	1.145	4.6	113	0.00
36	Benzo(k)fluoranthene	1.234	1.131	8.3	110	0.00
37 C	Benzo(a)pyrene	1.153	1.077	6.6	112	0.00
38	Dibenzo(a,h)anthracene	1.017	0.976	4.0	117	0.00
39	Benzo(g,h,i)perylene	1.076	1.018	5.4	112	0.00

Data Path : Z:\HPCHEM1\BNA_E\Data\BE041318\
Data File : BE095997.D
Acq On : 13 Apr 2018 15:39
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 13 16:48:18 2018
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE041118.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 13 14:42:08 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		