

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE041421\
 Data File : BE101297.D
 Acq On : 14 Apr 2021 11:44
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Apr 14 14:04:26 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE041221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 12 18:02:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.827	152	2539	0.40	ng	0.00
7) Naphthalene-d8	10.614	136	10421	0.40	ng	# 0.00
13) Acenaphthene-d10	14.442	164	6241	0.40	ng	0.00
19) Phenanthrene-d10	17.196	188	15384	0.40	ng	# 0.00
29) Chrysene-d12	21.353	240	21257	0.40	ng	# 0.00
36) Perylene-d12	23.833	264	21382	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.420	112	3013	0.40	ng	-0.01
5) Phenol-d6	6.998	99	4409	0.39	ng	0.00
8) Nitrobenzene-d5	8.977	82	3451	0.36	ng	0.00
11) 2-Methylnaphthalene-d10	12.195	152	9251	0.39	ng	0.00
14) 2,4,6-Tribromophenol	15.957	330	549	0.30	ng	0.00
15) 2-Fluorobiphenyl	13.074	172	9414	0.43	ng	0.00
27) Fluoranthene-d10	19.215	212	103770	0.39	ng	0.00
31) Terphenyl-d14	19.812	244	17895	0.35	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.359	88	1919	0.44	ng	95
3) n-Nitrosodimethylamine	3.670	42	2035	0.41	ng	# 97
6) bis(2-Chloroethyl)ether	7.251	93	3457	0.40	ng	99
9) Naphthalene	10.653	128	45732	0.41	ng	99
10) Hexachlorobutadiene	10.951	225	1977	0.41	ng	99
12) 2-Methylnaphthalene	12.267	142	7503	0.38	ng	95
16) Acenaphthylene	14.169	152	9230	0.36	ng	99
17) Acenaphthene	14.500	154	7109	0.40	ng	97
18) Fluorene	15.503	166	9202	0.39	ng	99
20) 4,6-Dinitro-2-methylph...	15.579	198	367	0.34	ng	92
21) 4-Bromophenyl-phenylether	16.395	248	2903	0.36	ng	# 81
22) Hexachlorobenzene	16.516	284	3034	0.40	ng	99
23) Atrazine	16.667	200	2211	0.31	ng	92
24) Pentachlorophenol	16.864	266	122	0.38	ng	94
25) Phenanthrene	17.242	178	17842	0.40	ng	100
26) Anthracene	17.332	178	14589	0.36	ng	99
28) Fluoranthene	19.244	202	22543	0.39	ng	100
30) Pyrene	19.605	202	22976	0.34	ng	100
32) Benzo(a)anthracene	21.341	228	25002	0.38	ng	99
33) Chrysene	21.390	228	28484	0.41	ng	99
34) Bis(2-ethylhexyl)phtha...	21.269	149	27531	0.31	ng	99
35) Indeno(1,2,3-cd)pyrene	26.449	276	28069	0.41	ng	98
37) Benzo(b)fluoranthene	23.074	252	26959	0.38	ng	99
38) Benzo(k)fluoranthene	23.124	252	29245	0.39	ng	98
39) Benzo(a)pyrene	23.723	252	23776	0.37	ng	99
40) Dibenzo(a,h)anthracene	26.477	278	24492	0.38	ng	98
41) Benzo(g,h,i)perylene	27.251	276	24772	0.38	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE041421\
Data File : BE101297.D
Acq On : 14 Apr 2021 11:44
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Quant Time: Apr 14 14:04:26 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE041221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 12 18:02:55 2021
Response via : Initial Calibration









1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.827 min Scan# 431
Delta R.T. 0.000 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	152	Resp:	2539
Ion	Ratio	Lower	Upper
152	100		
150	155.3	116.9	175.3
115	64.4	44.5	66.7



1,4-Dioxane
Concen: 0.44 ng
RT: 3.359 min Scan# 43
Delta R.T. -0.012 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44

Tgt Ion:	88	Resp:	1919
Ion	Ratio	Lower	Upper
88	100		
58	85.4	73.6	110.4
43	34.3	27.6	41.4









n-Nitrosodimethylamine
Concen: 0.41 ng
RT: 3.670 min Scan# 70
Delta R.T. -0.000 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 42 Resp: 2035
Ion Ratio Lower Upper
42 100
74 113.1 92.7 139.1
44 9.4 10.7 16.1#



2-Fluorophenol
Concen: 0.40 ng
RT: 5.420 min Scan# 222
Delta R.T. -0.012 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44

Tgt Ion: 112 Resp: 3013
Ion Ratio Lower Upper
112 100
64 76.9 62.3 93.5
63 150.9 117.3 175.9









Phenol-d6
Concen: 0.39 ng
RT: 6.998 min Scan# 359
Delta R.T. -0.000 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Tgt Ion:	99	Resp:	4409
Ion	Ratio	Lower	Upper
99	100		
42	24.2	18.5	27.7
71	33.3	26.5	39.7




bis(2-Chloroethyl)ether
Concen: 0.40 ng
RT: 7.251 min Scan# 381
Delta R.T. -0.000 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44

Tgt Ion:	93	Resp:	3457
Ion	Ratio	Lower	Upper
93	100		
63	367.5	292.5	438.7
95	32.7	25.7	38.5






Abundance Scan 650 (10.615 min): BE101280.D\data.ms (-645) (-)




Naphthalene-d8
Concen: 0.40 ng
RT: 10.614 min Scan# 650
Delta R.T. -0.001 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Tgt Ion:	136	Resp:	10421
Ion	Ratio	Lower	Upper
136	100		
137	11.4	8.6	13.0
54	32.0	34.7	52.1#
68	6.6	7.0	10.4#






Abundance Scan 524 (8.978 min): BE101280.D\data.ms (-520) (-)




Nitrobenzene-d5
Concen: 0.36 ng
RT: 8.977 min Scan# 524
Delta R.T. -0.001 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44

Tgt Ion:	82	Resp:	3451
Ion	Ratio	Lower	Upper
82	100		
128	159.5	118.7	178.1
54	203.7	164.4	246.6






Abundance Scan 654 (10.667 min): BE101280.D\data.ms (-649) (-)




Naphthalene
Concen: 0.41 ng
RT: 10.653 min Scan# 653
Delta R.T. -0.014 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Tgt Ion:	128	Resp:	45732
Ion	Ratio	Lower	Upper
128	100		
129	2.9	2.2	3.4
127	3.5	2.6	3.8







Abundance Scan 676 (10.953 min): BE101280.D\data.ms (-674) (-)



Hexachlorobutadiene
Concen: 0.41 ng
RT: 10.951 min Scan# 676
Delta R.T. -0.001 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44

Tgt Ion:	225	Resp:	1977
Ion	Ratio	Lower	Upper
225	100		
223	63.8	50.4	75.6
227	63.0	50.1	75.1









2-Methylnaphthalene-d10
Concen: 0.39 ng
RT: 12.195 min Scan# 763
Delta R.T. -0.001 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Tgt Ion:152 Resp: 9251
Ion Ratio Lower Upper
152 100
151 20.9 16.5 24.7




2-Methylnaphthalene
Concen: 0.38 ng
RT: 12.267 min Scan# 768
Delta R.T. -0.001 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44

Tgt Ion:142 Resp: 7503
Ion Ratio Lower Upper
142 100
141 86.8 73.0 109.6
115 32.5 27.8 41.8





Abundance Scan 919 (14.443 min): BE101280.D\data.ms (-913) (-)




Acenaphthene-d10
Concen: 0.40 ng
RT: 14.442 min Scan# 919
Delta R.T. -0.001 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	164	Resp:	6241
Ion	Ratio	Lower	Upper
164	100		
162	103.6	85.3	127.9
160	47.2	39.1	58.7



Abundance Scan 919 (14.442 min): BE101297.D\data.ms (-905) (-)





Abundance Scan 1016 (15.956 min): BE101280.D\data.ms (-1011) (-)




2,4,6-Tribromophenol
Concen: 0.30 ng
RT: 15.957 min Scan# 1016
Delta R.T. 0.001 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44


Tgt Ion:	330	Resp:	549
Ion	Ratio	Lower	Upper
330	100		
332	100.5	77.8	116.6
141	42.8	30.4	45.6



Abundance Scan 1016 (15.957 min): BE101297.D\data.ms (-1009) (-)






Abundance Scan 824 (13.074 min): BE101280.D\data.ms (-824) (-15)




2-Fluorobiphenyl
Concen: 0.43 ng
RT: 13.074 min Scan# 824
Delta R.T. -0.001 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Tgt Ion:	172	Resp:	9414
Ion	Ratio	Lower	Upper
172	100		
171	34.4	28.7	43.1
170	22.7	19.3	28.9







Abundance Scan 900 (14.169 min): BE101280.D\data.ms (-893) (-16)



Acenaphthylene
Concen: 0.36 ng
RT: 14.169 min Scan# 900
Delta R.T. -0.001 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44

Tgt Ion:	152	Resp:	9230
Ion	Ratio	Lower	Upper
152	100		
151	18.8	15.4	23.2
153	12.5	10.0	15.0









Acenaphthene
Concen: 0.40 ng
RT: 14.500 min Scan# 923
Delta R.T. -0.001 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	154	Resp:	7109
Ion	Ratio	Lower	Upper
154	100		
153	109.4	84.9	127.3
152	54.6	42.7	64.1



Fluorene
Concen: 0.39 ng
RT: 15.503 min Scan# 986
Delta R.T. 0.001 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44

Tgt Ion:	166	Resp:	9202
Ion	Ratio	Lower	Upper
166	100		
165	97.8	79.0	118.4
167	13.2	11.0	16.4









Phenanthrene-d10
Concen: 0.40 ng
RT: 17.196 min Scan# 1098
Delta R.T. 0.001 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Tgt Ion:	188	Resp:	15384
Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	12.7	4.6	6.8#



4,6-Dinitro-2-methylphenol
Concen: 0.34 ng
RT: 15.579 min Scan# 991
Delta R.T. 0.001 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44


Tgt Ion:	198	Resp:	367
Ion	Ratio	Lower	Upper
198	100		
51	74.7	55.9	83.9
105	49.1	33.6	50.4





Abundance Scan 1045 (16.394 min): BE101280.D\data.ms (-1022) (-)

4-Bromophenyl-phenylether
Concen: 0.36 ng
RT: 16.395 min Scan# 1045
Delta R.T. 0.001 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44


Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4



Tgt Ion:	248	Resp:	2903
Ion	Ratio	Lower	Upper
248	100		
250	98.0	84.5	126.7
141	59.3	19.8	29.8#





Abundance Scan 1045 (16.395 min): BE101297.D\data.ms (-1038) (-)




Abundance Scan 1053 (16.515 min): BE101280.D\data.ms (-1022) (-)

Hexachlorobenzene
Concen: 0.40 ng
RT: 16.516 min Scan# 1053
Delta R.T. 0.001 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44

Tgt Ion:	284	Resp:	3034
Ion	Ratio	Lower	Upper
284	100		
142	33.5	27.4	41.0
249	32.5	26.5	39.7




Abundance Scan 1053 (16.516 min): BE101297.D\data.ms (-1046) (-)






Abundance Scan 1063 (16.666 min): BE101280.D\data.ms (-1030) (-)

Atrazine
Concen: 0.31 ng
RT: 16.667 min Scan# 1063
Delta R.T. 0.001 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44


Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4



Tgt Ion:	200	Resp:	2211
Ion	Ratio	Lower	Upper
200	100		
173	22.1	21.9	32.9
215	52.4	38.6	57.8



Abundance Scan 1063 (16.667 min): BE101297.D\data.ms (-1030) (-)





Abundance Scan 1076 (16.863 min): BE101280.D\data.ms (-1072) (-)


Pentachlorophenol
Concen: 0.38 ng
RT: 16.864 min Scan# 1076
Delta R.T. 0.001 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44




Tgt Ion:	266	Resp:	122
Ion	Ratio	Lower	Upper
266	100		
264	68.0	49.1	73.7
268	65.6	50.6	76.0




Abundance Scan 1076 (16.864 min): BE101297.D\data.ms (-1069) (-)



Abundance Scan 1076 (16.864 min): BE101297.D\data.ms (-1069) (-)









Phenanthrene
Concen: 0.40 ng
RT: 17.242 min Scan# 1101
Delta R.T. 0.001 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Tgt Ion:178 Resp: 17842
Ion Ratio Lower Upper
178 100
176 19.0 15.0 22.6
179 15.1 12.2 18.4




Anthracene
Concen: 0.36 ng
RT: 17.332 min Scan# 1107
Delta R.T. 0.001 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44

Tgt Ion:178 Resp: 14589
Ion Ratio Lower Upper
178 100
176 18.0 14.4 21.6
179 15.3 11.8 17.8





Abundance Scan 1358 (19.214 min): BE101280.D\data.ms (-1329) (-)




Fluoranthene-d10
Concen: 0.39 ng
RT: 19.215 min Scan# 1358
Delta R.T. 0.001 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4


Tgt Ion:212 Resp: 103770
Ion Ratio Lower Upper
212 100
106 3.3 2.6 4.0
104 1.8 1.4 2.2



Abundance Scan 1358 (19.215 min): BE101297.D\data.ms (-1323) (-)





Abundance Scan 1363 (19.243 min): BE101280.D\data.ms (-1328) (-)




Fluoranthene
Concen: 0.39 ng
RT: 19.244 min Scan# 1363
Delta R.T. 0.001 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44

Tgt Ion:202 Resp: 22543
Ion Ratio Lower Upper
202 100
101 11.5 9.1 13.7
203 17.2 13.6 20.4




Abundance Scan 1363 (19.244 min): BE101297.D\data.ms (-1328) (-)






Abundance Scan 1604 (21.354 min): BE101280.D\data.ms (-1524) (-)

Chrysene-d12
Concen: 0.40 ng
RT: 21.353 min Scan# 1603
Delta R.T. -0.001 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4




Tgt Ion:	240	Resp:	21257
Ion	Ratio	Lower	Upper
240	100		
120	9.9	10.2	15.4#
236	27.3	21.7	32.5






Abundance Scan 1426 (19.605 min): BE101280.D\data.ms (-1430) (-)

Pyrene
Concen: 0.34 ng
RT: 19.605 min Scan# 1426
Delta R.T. 0.001 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44




Tgt Ion:	202	Resp:	22976
Ion	Ratio	Lower	Upper
202	100		
200	20.7	16.6	24.8
203	17.5	13.8	20.8





Abundance Scan 1462 (19.812 min): BE101280.D\data.ms (-1436) (-)

Terphenyl-d14
Concen: 0.35 ng
RT: 19.812 min Scan# 1462
Delta R.T. 0.001 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44


Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4



Tgt Ion:	244	Resp:	17895
Ion	Ratio	Lower	Upper
244	100		
212	27.7	21.9	32.9
122	9.8	7.3	10.9




Abundance Scan 1462 (19.812 min): BE101297.D\data.ms (-1436) (-)





Abundance Scan 1603 (21.342 min): BE101280.D\data.ms (-1586) (-)


Benzo(a)anthracene
Concen: 0.38 ng
RT: 21.341 min Scan# 1602
Delta R.T. -0.001 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44




Tgt Ion:	228	Resp:	25002
Ion	Ratio	Lower	Upper
228	100		
226	26.3	21.4	32.2
229	19.8	15.7	23.5



Abundance Scan 1602 (21.341 min): BE101297.D\data.ms (-1586) (-)









Chrysene
 Concen: 0.41 ng
 RT: 21.390 min Scan# 1606
 Delta R.T. -0.001 min
 Lab File: BE101297.D
 Acq: 14 Apr 2021 11:44




Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4

Tgt Ion:	228	Resp:	28484
Ion	Ratio	Lower	Upper
228	100		
226	30.5	24.8	37.2
229	19.1	15.2	22.8




Bis(2-ethylhexyl)phthalate
 Concen: 0.31 ng
 RT: 21.269 min Scan# 1596
 Delta R.T. -0.001 min
 Lab File: BE101297.D
 Acq: 14 Apr 2021 11:44

Tgt Ion:	149	Resp:	27531
Ion	Ratio	Lower	Upper
149	100		
167	7.2	5.5	8.3
279	1.2	1.0	1.4






Abundance Scan 2896 (26.455 min): BE101280.D\data.ms (-2896) (-)




Indeno(1,2,3-cd)pyrene
Concen: 0.41 ng
RT: 26.449 min Scan# 2893
Delta R.T. -0.007 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Tgt Ion:	276	Resp:	28069
Ion	Ratio	Lower	Upper
276	100		
138	23.2	17.4	26.0
277	25.0	19.8	29.8






Abundance Scan 2167 (23.842 min): BE101280.D\data.ms (-2167) (-)




Perylene-d12
Concen: 0.40 ng
RT: 23.833 min Scan# 2164
Delta R.T. -0.008 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44

Tgt Ion:	264	Resp:	21382
Ion	Ratio	Lower	Upper
264	100		
260	24.2	19.3	28.9
265	24.3	22.4	33.6






Abundance Scan 1953 (23.079 min): BE101280.D\data.ms (-1937) (-)




Benzo(b)fluoranthene
Concen: 0.38 ng
RT: 23.074 min Scan# 1951
Delta R.T. -0.005 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44

Tgt Ion:	252	Resp:	26959
Ion	Ratio	Lower	Upper
252	100		
253	22.8	18.1	27.1
125	10.0	7.6	11.4

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4






Abundance Scan 1966 (23.125 min): BE101280.D\data.ms (-1938) (-)



Benzo(k)fluoranthene
Concen: 0.39 ng
RT: 23.124 min Scan# 1965
Delta R.T. -0.001 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44


Tgt Ion:	252	Resp:	29245
Ion	Ratio	Lower	Upper
252	100		
253	22.9	17.8	26.6
125	10.1	7.6	11.4






Abundance Scan 2134 (23.724 min): BE101280.D\data.ms (-2377) (-)

Benzo(a)pyrene
Concen: 0.37 ng
RT: 23.723 min Scan# 2133
Delta R.T. -0.001 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4




Tgt Ion:	252	Resp:	23776
Ion	Ratio	Lower	Upper
252	100		
253	22.7	18.0	27.0
125	10.7	8.0	12.0






Abundance Scan 2903 (26.480 min): BE101280.D\data.ms (-2879) (-)


Dibenzo(a,h)anthracene
Concen: 0.38 ng
RT: 26.477 min Scan# 2901
Delta R.T. -0.003 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44



Tgt Ion:	278	Resp:	24492
Ion	Ratio	Lower	Upper
278	100		
139	16.5	12.2	18.2
279	23.8	19.3	28.9



Abundance Scan 3121 (27.258 min): BE101280.D\data.ms (-3064) (-)



Benzo(g,h,i)perylene
Concen: 0.38 ng
RT: 27.251 min Scan# 3118
Delta R.T. -0.007 min
Lab File: BE101297.D
Acq: 14 Apr 2021 11:44

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Tgt Ion:	276	Resp:	24772
Ion	Ratio	Lower	Upper
276	100		
277	23.7	19.4	29.0
138	19.8	15.2	22.8

