

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE041421\
 Data File : BE101297.D
 Acq On : 14 Apr 2021 11:44
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 14 14:04:26 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE041221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 12 18:02:55 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	89	0.00
2	1,4-Dioxane	0.400	0.440	-10.0	93	-0.01
3	n-Nitrosodimethylamine	0.400	0.408	-2.0	86	0.00
4 S	2-Fluorophenol	0.400	0.400	0.0	87	-0.01
5 S	Phenol-d6	0.400	0.386	3.5	86	0.00
6	bis(2-Chloroethyl)ether	0.400	0.399	0.3	84	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	83	0.00
8 S	Nitrobenzene-d5	0.400	0.359	10.3	77	0.00
9	Naphthalene	0.400	0.406	-1.5	83	-0.01
10	Hexachlorobutadiene	0.400	0.411	-2.7	84	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.387	3.3	80	0.00
12	2-Methylnaphthalene	0.400	0.378	5.5	78	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	72	0.00
14 S	2,4,6-Tribromophenol	0.400	0.298	25.5#	64	0.00
15 S	2-Fluorobiphenyl	0.400	0.433	-8.2	79	0.00
16	Acenaphthylene	0.400	0.362	9.5	71	0.00
17	Acenaphthene	0.400	0.400	0.0	74	0.00
18	Fluorene	0.400	0.388	3.0	71	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	72	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.340	15.0	63	0.00
21	4-Bromophenyl-phenylether	0.400	0.365	8.8	67	0.00
22	Hexachlorobenzene	0.400	0.396	1.0	74	0.00
23	Atrazine	0.400	0.314	21.5	63	0.00
24	Pentachlorophenol	0.400	0.377	5.8	50	0.00
25	Phenanthrene	0.400	0.398	0.5	72	0.00
26	Anthracene	0.400	0.364	9.0	68	0.00
27 SURR	Fluoranthene-d10	0.400	0.385	3.8	70	0.00
28	Fluoranthene	0.400	0.386	3.5	70	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	88	0.00
30	Pyrene	0.400	0.336	16.0	71	0.00
31 S	Terphenyl-d14	0.400	0.347	13.3	72	0.00
32	Benzo(a)anthracene	0.400	0.376	6.0	86	0.00
33	Chrysene	0.400	0.407	-1.7	89	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.310	22.5	79	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.407	-1.7	98	0.00
36 I	Perylene-d12	0.400	0.400	0.0	109	0.00
37	Benzo(b)fluoranthene	0.400	0.375	6.3	106	0.00
38	Benzo(k)fluoranthene	0.400	0.392	2.0	112	0.00
39 C	Benzo(a)pyrene	0.400	0.368	8.0	104	0.00
40	Dibenzo(a,h)anthracene	0.400	0.381	4.8	100	0.00
41	Benzo(g,h,i)perylene	0.400	0.378	5.5	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_E

LabSampleId :

SSTDCCC0.4