

Data Path : Z:\HPCHEM1\BNA E\DATA\BE041618\
 Data File : BE096006.D
 Acq On : 16 Apr 2018 13:00
 Operator : SJ/JU
 Sample : J2313-04
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 F1A31

Manual Integrations
 APPROVED

Sohil
 4/17/2018 4:51:31 PM

Quant Time: Apr 16 14:47:00 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE040618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 10 03:47:21 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.38	152	1877	0.40	ng/ul	0.00
2) Naphthalene-d8	11.21	136	5025	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.97	164	3061	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.68	188	6841m	0.40	ng/ul	0.00
16) Chrysene-d12	21.77	240	7112m	0.40	ng/ul	0.00
20) Perylene-d12	24.39	264	8568	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.76	152	2276	0.24	ng/ul	0.00
14) Fluoranthene-d10	19.66	212	6220m	0.22	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	11.26	128	1300	0.093	ng/ul#	86
8) Acenaphthene	15.03	153	319	0.028	ng/ul#	84
9) Fluorene	16.00	166	278	0.022	ng/ul#	90
12) Phenanthrene	17.72	178	488m	0.024	ng/ul	
15) Fluoranthene	19.69	202	1497m	0.055	ng/ul	
17) Pyrene	20.05	202	1307	0.054	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE041618\
Data File : BE096006.D
Acq On : 16 Apr 2018 13:00
Operator : SJ/JU
Sample : J2313-04
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
F1A31

Manual Integrations
APPROVED

Sohil
4/17/2018 4:51:31 PM

Quant Time: Apr 16 14:47:00 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE040618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 10 03:47:21 2018
Response via : Initial Calibration

