

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE041718\
 Data File : BE096030.D
 Acq On : 17 Apr 2018 20:50
 Operator : SJ/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.442

Quant Time: Apr 18 02:31:56 2018
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE040618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 18 01:18:08 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	108	0.00
2 I	Naphthalene-d8	0.400	0.400	0.0	109	0.00
3	Naphthalene	0.400	0.363	9.3	99	0.00
4 SURR	2-Methylnaphthalene-d10	0.400	0.377	5.8	103	0.00
5	2-Methylnaphthalene	0.400	0.370	7.5	101	0.00
6 I	Acenaphthene-d10	0.400	0.400	0.0	112	0.00
7	Acenaphthylene	0.400	0.358	10.5	99	0.00
8 C	Acenaphthene	0.400	0.363	9.3	101	0.00
9	Fluorene	0.400	0.363	9.3	100	0.00
10 I	Phenanthrene-d10	0.400	0.400	0.0	106	0.00
11	Pentachlorophenol	0.400	0.413	-3.2	122	0.00
12	Phenanthrene	0.400	0.356	11.0	96	0.00
13	Anthracene	0.400	0.362	9.5	96	0.00
14 SURR	Fluoranthene-d10	0.400	0.347	13.3	92	0.00
15 C	Fluoranthene	0.400	0.344	14.0	92	0.00
16 I	Chrysene-d12	0.400	0.400	0.0	91	0.00
17	Pyrene	0.400	0.368	8.0	89	0.00
18	Benzo(a)anthracene	0.400	0.349	12.8	86	0.00
19	Chrysene	0.400	0.346	13.5	86	0.00
20 I	Perylene-d12	0.400	0.400	0.0	89	0.00
21	Benzo(b)fluoranthene	0.400	0.366	8.5	79	0.00
22	Benzo(k)fluoranthene	0.400	0.374	6.5	88	0.00
23 C	Benzo(a)pyrene	0.400	0.366	8.5	83	0.00
24	Indeno(1,2,3-cd)pyrene	0.400	0.379	5.3	85	0.00
25	Dibenzo(a,h)anthracene	0.400	0.378	5.5	85	0.00
26	Benzo(g,h,i)perylene	0.400	0.376	6.0	85	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0