

Data Path : Z:\HPCHEM1\BNA E\DATA\BE041816\
 Data File : BE091663.D
 Acq On : 18 Apr 2016 13:50
 Operator : UM/SJ
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC1.6

Manual Integrations
 APPROVED

Sohil
 4/19/2016 6:39:54 PM

Quant Time: Apr 18 16:05:16 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE041816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 18 14:18:33 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.80	152	37085	5.00	ng	0.00
7) Naphthalene-d8	10.57	136	178865	5.00	ng	0.00
13) Acenaphthene-d10	14.42	164	108041	5.00	ng	0.00
19) Phenanthrene-d10	17.15	188	259222	5.00	ng	0.00
26) Chrysene-d12	21.31	240	299166	5.00	ng	0.00
35) Perylene-d12	23.75	264	274378	5.00	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.41	112	11779	1.33	ng	0.00
5) Phenol-d6	6.97	99	16173	1.37	ng	0.00
8) Nitrobenzene-d5	8.96	82	13657	1.31	ng	0.00
14) 2,4,6-Tribromophenol	15.90	330	3715m	1.25	ng	-0.01
15) 2-Fluorobiphenyl	13.05	172	45788	1.52	ng	0.00
29) Terphenyl-d14	19.76	244	66428	1.78	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.35	88	6525	1.38	ng	97
3) n-Nitrosodimethylamine	3.64	42	8644	1.52	ng	# 83
6) bis(2-Chloroethyl)ether	7.23	93	15311	1.50	ng	# 76
9) Nitrobenzene	8.99	77	14397	1.32	ng	96
10) Naphthalene	10.63	128	55239	1.50	ng	98
11) Hexachlorobutadiene	10.92	225	7937m	1.46	ng	
12) 2-Methylnaphthalene	12.24	142	36033	1.53	ng	97
16) Acenaphthylene	14.14	152	63388	1.51	ng	# 86
17) Acenaphthene	14.47	154	40565	1.53	ng	90
18) Fluorene	15.46	166	50427	1.52	ng	98
20) 4-Bromophenyl-phenylether	16.34	248	14012	1.53	ng	97
21) Hexachlorobenzene	16.47	284	14787	1.56	ng	97
22) Pentachlorophenol	16.79	266	4370	1.24	ng	# 89
23) Phenanthrene	17.20	178	90934	1.54	ng	99
24) Anthracene	17.28	178	82114	1.57	ng	99
25) Fluoranthene	19.20	202	89354	1.43	ng	98
27) Benzidine	19.38	184	23920m	1.21	ng	
28) Pyrene	19.55	202	97408	1.63	ng	99
30) Benzo(a)anthracene	21.29	228	111006m	1.66	ng	
31) 3,3'-Dichlorobenzidine	21.22	252	44357	1.24	ng	# 77
32) Chrysene	21.33	228	115163m	1.90	ng	
33) Bis(2-ethylhexyl)phthalate	21.22	149	19042	0.64	ng	# 85
34) Indeno(1,2,3-cd)pyrene	26.31	276	101796	1.50	ng	97
36) Benzo(b)fluoranthene	23.00	252	93952m	1.47	ng	
37) Benzo(k)fluoranthene	23.04	252	96089	1.53	ng	97
38) Benzo(a)pyrene	23.63	252	84713	1.43	ng	97
39) Dibenzo(a,h)anthracene	26.33	278	84838	1.46	ng	96
40) Benzo(g,h,i)perylene	27.09	276	86798	1.45	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA E\DATA\BE041816\
Data File : BE091663.D
Acq On : 18 Apr 2016 13:50
Operator : UM/SJ
Sample : SSTDIC1.6
Misc :
ALS Vial : 6 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6

Manual Integrations
APPROVED

Sohil
4/19/2016 6:39:54 PM

Quant Time: Apr 18 16:05:16 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE041816.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 18 14:18:33 2016
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 7.80 min Scan# 300

Delta R.T. 0.00 min

Lab File: BE091663.D

Acq: 18 Apr 2016 13:50

Instrument :

BNA_E

ClientSampleId :

SSTDICC1.6

Tot Ion:152 Resp: 37085

Ion Ratio Lower Upper

152 100


150 170.4 122.6 183.8

115 72.3 52.8 79.2




Manual Integrations
APPROVED

Sohil

4/19/2016 6:39:54 PM



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 1.38 ng

RT: 3.35 min Scan# 44

Delta R.T. -0.00 min

Lab File: BE091663.D

Acq: 18 Apr 2016 13:50


Tgt Ion: 88 Resp: 6525

Ion Ratio Lower Upper



88 100


58 81.3 65.8 98.6

43 37.0 25.3 37.9



Abundance Ion 88.00 (87.70 to 88.70): BE0
Ion 58.00 (57.70 to 58.70): BE0
Ion 43.00 (42.70 to 43.70): BE0





#3





n-Nitrosodimethylamine
 Concen: 1.52 ng
 RT: 3.64 min Scan# 61
 Delta R.T. -0.02 min
 Lab File: BE091663.D
 Acq: 18 Apr 2016 13:50

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC1.6

Tot Ion	Ratio	Lower	Upper
42	100		
74	90.8	87.9	131.9
44	6.3	6.6	9.8#

Manual Integrations
 APPROVED




Sohil
 4/19/2016 6:39:54 PM






#4

2-Fluorophenol
 Concen: 1.33 ng
 RT: 5.41 min Scan# 163
 Delta R.T. -0.00 min
 Lab File: BE091663.D
 Acq: 18 Apr 2016 13:50

Tgt Ion	Ratio	Lower	Upper
112	100		
64	67.7	30.9	46.3#
63	36.8	16.7	25.1#





#5

Phenol-d6

Concen: 1.37 ng

RT: 6.97 min Scan# 252

Delta R.T. 0.00 min

Lab File: BE091663.D

Acq: 18 Apr 2016 13:50

Instrument :

BNA_E





ClientSampleId :

SSTDICC1.6

Tot Ion:	99	Resp:	16173
Ion Ratio	Lower	Upper	
99	100		
42	26.5	16.2	24.2#
71	35.6	29.0	43.4

Manual Integrations
APPROVED

Sohil
4/19/2016 6:39:54 PM



#6

bis(2-Chloroethyl)ether

Concen: 1.50 ng


RT: 7.23 min Scan# 267


Delta R.T. 0.00 min

Lab File: BE091663.D

Acq: 18 Apr 2016 13:50

Tgt Ion:	93	Resp:	15311
Ion Ratio	Lower	Upper	
93	100		
63	85.3	49.5	74.3#
95	32.6	22.2	33.4





#7

Naphthalene-d8

Concen: 5.00 ng

RT: 10.57 min Scan# 511

Delta R.T. -0.00 min

Lab File: BE091663.D


Acq: 18 Apr 2016 13:50

Instrument :

BNA_E

ClientSampleId :

SSTDICC1.6



Tot Ion:136 Resp: 178865

Ion Ratio Lower Upper

136 100



137 9.9 8.9 13.3

54 13.4 8.2 12.4

68 8.8 6.2 9.4

Manual Integrations
APPROVED

Sohil
4/19/2016 6:39:54 PM



#8

Nitrobenzene-d5

Concen: 1.31 ng

RT: 8.96 min Scan# 392

Delta R.T. -0.00 min

Lab File: BE091663.D

Acq: 18 Apr 2016 13:50




Tgt Ion: 82 Resp: 13657


Ion Ratio Lower Upper

82 100

128 34.4 25.4 38.0

54 51.1 48.4 72.6





#9

Nitrobenzene

Concen: 1.32 ng

RT: 8.99 min Scan# 395

Delta R.T. -0.01 min

Lab File: BE091663.D

Acq: 18 Apr 2016 13:50

Instrument :

BNA_E





ClientSampleId :

SSTDICC1.6

Tot Ion:	77	Resp:	14397
Ion Ratio	Lower	Upper	
77	100		
123	35.3	26.9	40.3
65	13.9	9.4	14.2

Manual Integrations
APPROVED

Sohil
4/19/2016 6:39:54 PM



#10

Naphthalene

Concen: 1.50 ng




RT: 10.63 min Scan# 514


Delta R.T. -0.00 min

Lab File: BE091663.D

Acq: 18 Apr 2016 13:50

Tgt Ion:	128	Resp:	55239
Ion Ratio	Lower	Upper	
128	100		
129	11.4	9.3	13.9
127	12.4	10.7	16.1









#11
Hexachlorobutadiene
Concen: 1.46 ng/ml
RT: 10.92 min Scan# 529
Delta R.T. -0.00 min
Lab File: BE091663.D
Acq: 18 Apr 2016 13:50

Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6

Tot Ion	Ratio	Lower	Upper
225	100		
223	73.7	49.0	73.6
227	74.9	50.3	75.5




Manual Integrations
APPROVED


Sohil
4/19/2016 6:39:54 PM



#12
2-Methylnaphthalene
Concen: 1.53 ng/ml
RT: 12.24 min Scan# 638
Delta R.T. -0.00 min
Lab File: BE091663.D
Acq: 18 Apr 2016 13:50

Tgt Ion	Ratio	Lower	Upper
142	100		
141	87.7	71.4	107.0
115	34.7	30.3	45.5





#13

Acenaphthene-d10

Concen: 5.00 ng

RT: 14.42 min Scan# 820

Delta R.T. 0.00 min

Lab File: BE091663.D

Acq: 18 Apr 2016 13:50

Instrument :

BNA_E





ClientSampleId :

SSTDICC1.6

Tot Ion:	164	Resp:	108041
Ion Ratio	Lower	Upper	
164	100		
162	101.2	85.3	127.9
160	44.1	46.2	69.2#

Manual Integrations
APPROVED

Sohil
4/19/2016 6:39:54 PM



#14

2,4,6-Tribromophenol

Concen: 1.25 ng m




RT: 15.90 min Scan# 939


Delta R.T. -0.01 min

Lab File: BE091663.D

Acq: 18 Apr 2016 13:50

Tgt Ion:	330	Resp:	3715
Ion Ratio	Lower	Upper	
330	100		
332	123.8	79.1	118.7#
141	46.8	125.4	188.0#









#15
2-Fluorobiphenyl
Concen: 1.52 ng
RT: 13.05 min Scan# 708
Delta R.T. -0.00 min
Lab File: BE091663.D
Acq: 18 Apr 2016 13:50

Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6

Tot Ion	Ratio	Lower	Upper
172	100		
171	31.3	28.8	43.2
170	18.6	19.3	28.9




Manual Integrations
APPROVED


Sohil
4/19/2016 6:39:54 PM



#16
Acenaphthylene
Concen: 1.51 ng
RT: 14.14 min Scan# 799
Delta R.T. 0.00 min
Lab File: BE091663.D
Acq: 18 Apr 2016 13:50

Tgt Ion	Ratio	Lower	Upper
152	100		
151	18.0	3.9	5.9
153	14.3	10.3	15.5









#17
Acenaphthene
Concen: 1.53 ng
RT: 14.47 min Scan# 824
Delta R.T. 0.00 min
Lab File: BE091663.D
Acq: 18 Apr 2016 13:50

Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6

Tot Ion	Ratio	Lower	Upper
154	100		
153	111.6	80.0	120.0
152	54.9	40.0	60.0




Manual Integrations
APPROVED


Sohil
4/19/2016 6:39:54 PM



#18
Fluorene
Concen: 1.52 ng
RT: 15.46 min Scan# 901
Delta R.T. 0.00 min
Lab File: BE091663.D
Acq: 18 Apr 2016 13:50

Tgt Ion	Ratio	Lower	Upper
166	100		
165	98.9	80.8	121.2
167	13.6	10.9	16.3









#19
Phenanthrene-d10
Concen: 5.00 ng
RT: 17.15 min Scan# 1016
Delta R.T. 0.00 min
Lab File: BE091663.D
Acq: 18 Apr 2016 13:50

Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6

Tot Ion	Ratio	Resp Lower	Upper
188	100		
94	9.8	8.7	13.1
80	9.6	9.4	14.0




Manual Integrations
APPROVED


Sohil
4/19/2016 6:39:54 PM



#20
4-Bromophenyl-phenylether
Concen: 1.53 ng
RT: 16.34 min Scan# 967
Delta R.T. 0.00 min
Lab File: BE091663.D
Acq: 18 Apr 2016 13:50

Tgt Ion	Ratio	Resp Lower	Upper
248	100		
250	100.0	80.5	120.7
141	85.6	72.0	108.0





#21

Hexachlorobenzene

Concen: 1.56 ng

RT: 16.47 min Scan# 974

Delta R.T. 0.00 min

Lab File: BE091663.D

Acq: 18 Apr 2016 13:50

Instrument :

BNA_E





ClientSampleId :

SSTDICC1.6

Tot Ion:	284	Resp:	14787
Ion Ratio	Lower	Upper	
284	100		
142	41.9	32.2	48.2
249	33.1	25.4	38.2

Manual Integrations
APPROVED

Sohil
4/19/2016 6:39:54 PM



#22

Pentachlorophenol

Concen: 1.24 ng




RT: 16.79 min Scan# 993


Delta R.T. -0.02 min

Lab File: BE091663.D

Acq: 18 Apr 2016 13:50

Tgt Ion:	266	Resp:	4370
Ion Ratio	Lower	Upper	
266	100		
268	55.7	58.2	87.2#
264	70.8	56.2	84.4









#23
Phenanthrene
Concen: 1.54 ng
RT: 17.20 min Scan# 1020
Delta R.T. 0.00 min
Lab File: BE091663.D
Acq: 18 Apr 2016 13:50

Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6

Tot Ion:178 Resp: 90934
Ion Ratio Lower Upper
178 100
176 19.5 15.3 22.9
179 15.6 12.6 18.8




Manual Integrations
APPROVED


Sohil
4/19/2016 6:39:54 PM



#24
Anthracene
Concen: 1.57 ng
RT: 17.28 min Scan# 1027
Delta R.T. -0.01 min
Lab File: BE091663.D
Acq: 18 Apr 2016 13:50

Tgt Ion:178 Resp: 82114
Ion Ratio Lower Upper
178 100
176 18.7 14.6 21.8
179 15.3 11.8 17.6





#25

Fluoranthene

Concen: 1.43 ng

RT: 19.20 min Scan# 1151

Delta R.T. -0.00 min

Lab File: BE091663.D

Acq: 18 Apr 2016 13:50

Instrument :

BNA_E


ClientSampleId :

SSTDICC1.6


Tot Ion:	202	Resp:	89354
Ion Ratio	Lower	Upper	
202	100		
101	12.1	11.0	16.6
203	17.6	13.9	20.9

Manual Integrations
APPROVED



Sohil
4/19/2016 6:39:54 PM



Abundance Ion 202.00 (201.70 to 202.70): E
Ion 101.00 (100.70 to 101.70): E
Ion 203.00 (202.70 to 203.70): E



Time--> 19.00 19.20 19.40



#26

Chrysene-d12

Concen: 5.00 ng


RT: 21.31 min Scan# 1248

Delta R.T. -0.00 min


Lab File: BE091663.D

Acq: 18 Apr 2016 13:50


Tgt Ion:	240	Resp:	299166
Ion Ratio	Lower	Upper	
240	100		
120	1.2	11.0	16.4#
236	23.2	21.7	32.5




Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E



Time--> 21.20 21.40





#27

Benzidine

Concen: 1.21 ng/ml

RT: 19.38 min Scan# 1160

Delta R.T. -0.02 min

Lab File: BE091663.D

Acq: 18 Apr 2016 13:50

Instrument :

BNA_E





ClientSampleId :

SSTDICC1.6

Tot Ion:	184	Resp:	23920
Ion Ratio	Lower	Upper	
184	100		
185	13.3	22.3	33.5#
183	11.5	16.4	24.6#

Manual Integrations
APPROVED

Sohil
4/19/2016 6:39:54 PM



#28

Pyrene

Concen: 1.63 ng/ml




RT: 19.55 min Scan# 1169


Delta R.T. -0.00 min

Lab File: BE091663.D

Acq: 18 Apr 2016 13:50

Tgt Ion:	202	Resp:	97408
Ion Ratio	Lower	Upper	
202	100		
200	21.7	16.8	25.2
203	17.3	14.0	21.0









#29
 Terphenyl-d14
 Concen: 1.78 ng
 RT: 19.76 min Scan# 1180
 Delta R.T. -0.00 min
 Lab File: BE091663.D
 Acq: 18 Apr 2016 13:50

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC1.6

Tot Ion	Ratio	Lower	Upper
244	100		
212	9.0	6.9	10.3
122	5.5	11.0	16.4




Manual Integrations
 APPROVED


Sohil
 4/19/2016 6:39:54 PM



#30
 Benzo(a)anthracene
 Concen: 1.66 ng m
 RT: 21.29 min Scan# 1247
 Delta R.T. -0.00 min
 Lab File: BE091663.D
 Acq: 18 Apr 2016 13:50

Tgt Ion	Ratio	Lower	Upper
228	100		
226	24.5	21.0	31.4
229	21.1	15.7	23.5









#31
 3,3'-Dichlorobenzidine
 Concen: 1.24 ng
 RT: 21.22 min Scan# 1244
 Delta R.T. -0.02 min
 Lab File: BE091663.D
 Acq: 18 Apr 2016 13:50

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC1.6

Tot Ion	Ratio	Lower	Upper
252	100		
254	45.5	51.1	76.7#
126	23.8	12.6	18.8#




Manual Integrations
 APPROVED


Sohil
 4/19/2016 6:39:54 PM



#32
 Chrysene
 Concen: 1.90 ng m
 RT: 21.33 min Scan# 1249
 Delta R.T. -0.00 min
 Lab File: BE091663.D
 Acq: 18 Apr 2016 13:50

Tgt Ion	Ratio	Lower	Upper
228	100		
226	32.8	24.0	36.0
229	17.7	15.9	23.9





#33

Bis(2-ethylhexyl)phthalate

Concen: 0.64 ng

RT: 21.22 min Scan# 1244

Delta R.T. -0.00 min

Lab File: BE091663.D

Acq: 18 Apr 2016 13:50

Instrument :

BNA_E

ClientSampleId :

SSTDICC1.6

Tot Ion:149 Resp: 19042

Ion Ratio Lower Upper

149 100


167 29.8 19.5 29.3#

279 4.0 10.0 15.0#




Manual Integrations
APPROVED

Sohil

4/19/2016 6:39:54 PM



Abundance Ion 149.00 (148.70 to 149.70): E
Ion 167.00 (166.70 to 167.70): E
Ion 279.00 (278.70 to 279.70): E



#34

Indeno(1,2,3-cd)pyrene

Concen: 1.50 ng

RT: 26.31 min Scan# 1645

Delta R.T. -0.01 min

Lab File: BE091663.D

Acq: 18 Apr 2016 13:50


Tgt Ion:276 Resp: 101796

Ion Ratio Lower Upper



276 100


138 26.0 23.4 35.2

277 24.9 19.8 29.8



Abundance Ion 276.00 (275.70 to 276.70): E
Ion 138.00 (137.70 to 138.70): E
Ion 277.00 (276.70 to 277.70): E









#35
Pervlene-d12
Concen: 5.00 ng
RT: 23.75 min Scan# 1426
Delta R.T. -0.01 min
Lab File: BE091663.D
Acq: 18 Apr 2016 13:50

Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6

Tot Ion	Ratio	Lower	Upper
264	100		
260	25.9	20.0	30.0
265	21.3	17.5	26.3




Manual Integrations
APPROVED


Sohil
4/19/2016 6:39:54 PM



#36
Benzo(b)fluoranthene
Concen: 1.47 ng m
RT: 23.00 min Scan# 1361
Delta R.T. 0.00 min
Lab File: BE091663.D
Acq: 18 Apr 2016 13:50

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.4	26.0
125	10.1	10.2	15.2#









#37
Benzo(k)fluoranthene
Concen: 1.53 ng
RT: 23.04 min Scan# 1365
Delta R.T. -0.01 min
Lab File: BE091663.D
Acq: 18 Apr 2016 13:50

Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6

Tot Ion	Ratio	Lower	Upper
252	100		
253	21.1	17.6	26.4
125	10.5	9.9	14.9




Manual Integrations
APPROVED


Sohil
4/19/2016 6:39:54 PM



#38
Benzo(a)pyrene
Concen: 1.43 ng
RT: 23.63 min Scan# 1416
Delta R.T. -0.01 min
Lab File: BE091663.D
Acq: 18 Apr 2016 13:50

Tgt Ion	Ratio	Lower	Upper
252	100		
253	21.4	18.1	27.1
125	12.0	11.0	16.4









#39
Dibenzo(a,h)anthracene
Concen: 1.46 ng
RT: 26.33 min Scan# 1647
Delta R.T. -0.01 min
Lab File: BE091663.D
Acq: 18 Apr 2016 13:50

Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6

Tot Ion	Ratio	Lower	Upper
278	100		
139	16.8	16.0	24.0
279	23.5	19.4	29.2

Manual Integrations
APPROVED

Sohil
4/19/2016 6:39:54 PM



#40
Benzo(g,h,i)perylene
Concen: 1.45 ng
RT: 27.09 min Scan# 1713
Delta R.T. -0.01 min
Lab File: BE091663.D
Acq: 18 Apr 2016 13:50

Tgt Ion	Ratio	Lower	Upper
276	100		
277	23.0	18.6	28.0
138	22.5	20.4	30.6

