

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE041921\
 Data File : BE101305.D
 Acq On : 19 Apr 2021 16:19
 Operator : CG/JU
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.8

Quant Time: Apr 19 17:01:12 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE041921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 19 16:27:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.784	152	2627	0.40	ng	0.00
7) Naphthalene-d8	10.576	136	10654	0.40	ng	0.00
13) Acenaphthene-d10	14.400	164	6717	0.40	ng	-0.01
19) Phenanthrene-d10	17.165	188	15694	0.40	ng	0.00
29) Chrysene-d12	21.318	240	20133	0.40	ng	0.00
36) Perylene-d12	23.770	264	20163	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.377	112	5577	0.72	ng	0.00
5) Phenol-d6	6.955	99	8105	0.69	ng	0.00
8) Nitrobenzene-d5	8.926	82	7032	0.72	ng	0.00
11) 2-Methylnaphthalene-d10	12.153	152	18572	0.76	ng	0.00
14) 2,4,6-Tribromophenol	15.910	330	1192	0.60	ng	0.00
15) 2-Fluorobiphenyl	13.032	172	19403	0.83	ng	0.00
27) Fluoranthene-d10	19.174	212	200616	0.73	ng	0.00
31) Terphenyl-d14	19.772	244	34857	0.71	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.304	88	3557	0.79	ng	93
3) n-Nitrosodimethylamine	3.615	42	3767	0.73	ng	# 98
6) bis(2-Chloroethyl)ether	7.208	93	6800	0.76	ng	96
9) Naphthalene	10.615	128	90671	0.79	ng	100
10) Hexachlorobutadiene	10.913	225	3990	0.81	ng	98
12) 2-Methylnaphthalene	12.225	142	15408	0.76	ng	99
16) Acenaphthylene	14.127	152	20085	0.73	ng	100
17) Acenaphthene	14.472	154	14826	0.78	ng	97
18) Fluorene	15.472	166	19263	0.75	ng	99
20) 4,6-Dinitro-2-methylph...	15.547	198	859	0.58	ng	# 80
21) 4-Bromophenyl-phenylether	16.364	248	6446	0.79	ng	96
22) Hexachlorobenzene	16.470	284	6273	0.80	ng	98
23) Atrazine	16.621	200	4499	0.63	ng	98
24) Pentachlorophenol	16.817	266	280	0.44	ng	# 74
25) Phenanthrene	17.210	178	35613	0.78	ng	99
26) Anthracene	17.301	178	30183	0.74	ng	99
28) Fluoranthene	19.203	202	44636	0.75	ng	100
30) Pyrene	19.565	202	44476	0.69	ng	100
32) Benzo(a)anthracene	21.306	228	48108	0.76	ng	99
33) Chrysene	21.354	228	51849	0.78	ng	99
34) Bis(2-ethylhexyl)phtha...	21.233	149	50154	0.60	ng	100
35) Indeno(1,2,3-cd)pyrene	26.348	276	52019	0.80	ng	95
37) Benzo(b)fluoranthene	23.015	252	48188	0.71	ng	99
38) Benzo(k)fluoranthene	23.065	252	53637	0.76	ng	98
39) Benzo(a)pyrene	23.656	252	43530	0.71	ng	100
40) Dibenzo(a,h)anthracene	26.377	278	44927	0.74	ng	99
41) Benzo(g,h,i)perylene	27.144	276	45132	0.73	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE041921\
Data File : BE101305.D
Acq On : 19 Apr 2021 16:19
Operator : CG/JU
Sample : SSTDICC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDICC0.8

Quant Time: Apr 19 17:01:12 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE041921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 19 16:27:30 2021
Response via : Initial Calibration









1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.784 min Scan# 427
 Delta R.T. -0.008 min
 Lab File: BE101305.D
 Acq: 19 Apr 2021 16:19




Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.8


Tgt Ion:	152	Resp:	2627
Ion	Ratio	Lower	Upper
152	100		
150	161.2	124.8	187.2
115	65.4	48.7	73.1



1,4-Dioxane
 Concen: 0.79 ng
 RT: 3.304 min Scan# 38
 Delta R.T. -0.008 min
 Lab File: BE101305.D
 Acq: 19 Apr 2021 16:19

Tgt Ion:	88	Resp:	3557
Ion	Ratio	Lower	Upper
88	100		
58	88.5	65.4	98.0
43	32.6	24.4	36.6









n-Nitrosodimethylamine
Concen: 0.73 ng
RT: 3.615 min Scan# 65
Delta R.T. -0.008 min
Lab File: BE101305.D
Acq: 19 Apr 2021 16:19




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.8


Tgt Ion:	42	Resp:	3767
Ion	Ratio	Lower	Upper
42	100		
74	115.4	90.9	136.3
44	9.1	9.2	13.8#



2-Fluorophenol
Concen: 0.72 ng
RT: 5.377 min Scan# 218
Delta R.T. -0.008 min
Lab File: BE101305.D
Acq: 19 Apr 2021 16:19

Tgt Ion:	112	Resp:	5577
Ion	Ratio	Lower	Upper
112	100		
64	77.9	62.8	94.2
63	148.5	120.8	181.2









Phenol-d6
 Concen: 0.69 ng
 RT: 6.955 min Scan# 355
 Delta R.T. -0.008 min
 Lab File: BE101305.D
 Acq: 19 Apr 2021 16:19




Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.8

Tgt Ion:	99	Resp:	8105
Ion	Ratio	Lower	Upper
99	100		
42	22.9	18.8	28.2
71	34.3	26.6	40.0




bis(2-Chloroethyl)ether
 Concen: 0.76 ng
 RT: 7.208 min Scan# 377
 Delta R.T. -0.008 min
 Lab File: BE101305.D
 Acq: 19 Apr 2021 16:19

Tgt Ion:	93	Resp:	6800
Ion	Ratio	Lower	Upper
93	100		
63	360.9	296.7	445.1
95	32.5	25.9	38.9





Abundance Scan 647 (10.576 min): BE101304.D\data.ms (-647) (-)




Naphthalene-d8
Concen: 0.40 ng
RT: 10.576 min Scan# 647
Delta R.T. -0.001 min
Lab File: BE101305.D
Acq: 19 Apr 2021 16:19

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.8


Tgt Ion:	136	Resp:	10654
Ion	Ratio	Lower	Upper
136	100		
137	11.5	9.0	13.6
54	26.1	21.3	31.9
68	5.4	4.2	6.4



Abundance Scan 647 (10.576 min): BE101305.D\data.ms (-632) (-)





Abundance Scan 520 (8.926 min): BE101304.D\data.ms (-517) (-)




Nitrobenzene-d5
Concen: 0.72 ng
RT: 8.926 min Scan# 520
Delta R.T. -0.001 min
Lab File: BE101305.D
Acq: 19 Apr 2021 16:19


Tgt Ion:	82	Resp:	7032
Ion	Ratio	Lower	Upper
82	100		
128	108.0	91.2	136.8
54	253.6	204.0	306.0



Abundance Scan 520 (8.926 min): BE101305.D\data.ms (-507) (-)






Abundance Scan 650 (10.615 min): BE101304.D\data.ms (-642) (-)




Naphthalene
Concen: 0.79 ng
RT: 10.615 min Scan# 650
Delta R.T. -0.001 min
Lab File: BE101305.D
Acq: 19 Apr 2021 16:19

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.8

Tgt Ion:	128	Resp:	90671
Ion	Ratio	Lower	Upper
128	100		
129	2.8	2.3	3.5
127	3.4	2.6	4.0







Abundance Scan 673 (10.913 min): BE101304.D\data.ms (-667) (-)



Hexachlorobutadiene
Concen: 0.81 ng
RT: 10.913 min Scan# 673
Delta R.T. -0.001 min
Lab File: BE101305.D
Acq: 19 Apr 2021 16:19

Tgt Ion:	225	Resp:	3990
Ion	Ratio	Lower	Upper
225	100		
223	63.1	49.4	74.0
227	64.2	49.7	74.5









2-Methylnaphthalene-d10
Concen: 0.76 ng
RT: 12.153 min Scan# 760
Delta R.T. 0.000 min
Lab File: BE101305.D
Acq: 19 Apr 2021 16:19




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.8


Tgt Ion:152 Resp: 18572
Ion Ratio Lower Upper
152 100
151 20.6 16.2 24.4



2-Methylnaphthalene
Concen: 0.76 ng
RT: 12.225 min Scan# 765
Delta R.T. 0.000 min
Lab File: BE101305.D
Acq: 19 Apr 2021 16:19

Tgt Ion:142 Resp: 15408
Ion Ratio Lower Upper
142 100
141 87.5 71.4 107.0
115 33.4 27.0 40.6









Acenaphthene-d10
 Concen: 0.40 ng
 RT: 14.400 min Scan# 916
 Delta R.T. -0.014 min
 Lab File: BE101305.D
 Acq: 19 Apr 2021 16:19




Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.8


Tgt Ion:	164	Resp:	6717
Ion	Ratio	Lower	Upper
164	100		
162	107.2	78.4	117.6
160	48.9	33.3	49.9



2,4,6-Tribromophenol
 Concen: 0.60 ng
 RT: 15.910 min Scan# 1013
 Delta R.T. 0.000 min
 Lab File: BE101305.D
 Acq: 19 Apr 2021 16:19

Tgt Ion:	330	Resp:	1192
Ion	Ratio	Lower	Upper
330	100		
332	100.8	79.7	119.5
141	40.6	33.2	49.8









2-Fluorobiphenyl
 Concen: 0.83 ng
 RT: 13.032 min Scan# 821
 Delta R.T. 0.000 min
 Lab File: BE101305.D
 Acq: 19 Apr 2021 16:19




Instrument :
 BNA_E
 ClientSampleId :
 SSTDIC0.8


Tgt Ion:	172	Resp:	19403
Ion	Ratio	Lower	Upper
172	100		
171	35.5	29.0	43.6
170	23.5	19.1	28.7



Acenaphthylene
 Concen: 0.73 ng
 RT: 14.127 min Scan# 897
 Delta R.T. 0.000 min
 Lab File: BE101305.D
 Acq: 19 Apr 2021 16:19

Tgt Ion:	152	Resp:	20085
Ion	Ratio	Lower	Upper
152	100		
151	19.3	15.6	23.4
153	13.1	10.7	16.1









Acenaphthene
Concen: 0.78 ng
RT: 14.472 min Scan# 921
Delta R.T. 0.000 min
Lab File: BE101305.D
Acq: 19 Apr 2021 16:19




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.8


Tgt Ion	Ratio	Lower	Upper
154	100		
153	109.0	90.1	135.1
152	54.1	44.2	66.2



Fluorene
Concen: 0.75 ng
RT: 15.472 min Scan# 984
Delta R.T. -0.000 min
Lab File: BE101305.D
Acq: 19 Apr 2021 16:19

Tgt Ion	Ratio	Lower	Upper
166	100		
165	97.4	78.9	118.3
167	13.2	11.0	16.4









Phenanthrene-d10
 Concen: 0.40 ng
 RT: 17.165 min Scan# 1096
 Delta R.T. -0.000 min
 Lab File: BE101305.D
 Acq: 19 Apr 2021 16:19




Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.8

Tgt Ion:	188	Resp:	15694
Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	7.2	5.8	8.8



4,6-Dinitro-2-methylphenol
 Concen: 0.58 ng
 RT: 15.547 min Scan# 989
 Delta R.T. -0.000 min
 Lab File: BE101305.D
 Acq: 19 Apr 2021 16:19


Tgt Ion:	198	Resp:	859
Ion	Ratio	Lower	Upper
198	100		
51	51.4	56.0	84.0#
105	34.8	36.7	55.1#






Abundance Scan 1043 (16.364 min): BE101304.D\data.ms (-1041) (-)

4-Bromophenyl-phenylether
Concen: 0.79 ng
RT: 16.364 min Scan# 1043
Delta R.T. -0.000 min
Lab File: BE101305.D
Acq: 19 Apr 2021 16:19

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.8




Tgt Ion:	248	Resp:	6446
Ion	Ratio	Lower	Upper
248	100		
250	99.6	79.2	118.8
141	32.1	31.1	46.7







Abundance Scan 1050 (16.470 min): BE101304.D\data.ms (-1042) (-)

Hexachlorobenzene
Concen: 0.80 ng
RT: 16.470 min Scan# 1050
Delta R.T. -0.000 min
Lab File: BE101305.D
Acq: 19 Apr 2021 16:19



Tgt Ion:	284	Resp:	6273
Ion	Ratio	Lower	Upper
284	100		
142	35.1	27.2	40.8
249	34.1	26.6	40.0









Atrazine
Concen: 0.63 ng
RT: 16.621 min Scan# 1060
Delta R.T. -0.000 min
Lab File: BE101305.D
Acq: 19 Apr 2021 16:19




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.8


Tgt Ion:	200	Resp:	4499
Ion	Ratio	Lower	Upper
200	100		
173	34.0	27.8	41.8
215	47.5	39.8	59.6



Pentachlorophenol
Concen: 0.44 ng
RT: 16.817 min Scan# 1073
Delta R.T. -0.000 min
Lab File: BE101305.D
Acq: 19 Apr 2021 16:19

Tgt Ion:	266	Resp:	280
Ion	Ratio	Lower	Upper
266	100		
264	50.0	67.0	100.6#
268	58.6	54.1	81.1









Scan 1098 (17.196 min): BE101304.D\data.ms (-1025) (-)
 Phenanthrene
 Concen: 0.78 ng
 RT: 17.210 min Scan# 1099
 Delta R.T. 0.015 min
 Lab File: BE101305.D
 Acq: 19 Apr 2021 16:19




Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.8

Tgt Ion:	178	Resp:	35613
Ion	Ratio	Lower	Upper
178	100		
176	18.7	15.0	22.4
179	15.3	12.6	19.0




Scan 1105 (17.301 min): BE101304.D\data.ms (-1026) (-)
 Anthracene
 Concen: 0.74 ng
 RT: 17.301 min Scan# 1105
 Delta R.T. -0.000 min
 Lab File: BE101305.D
 Acq: 19 Apr 2021 16:19

Tgt Ion:	178	Resp:	30183
Ion	Ratio	Lower	Upper
178	100		
176	18.4	14.5	21.7
179	15.3	12.1	18.1





Abundance Scan 1351 (19.175 min): BE101304.D\data.ms (-1323) (-)




Fluoranthene-d10
Concen: 0.73 ng
RT: 19.174 min Scan# 1351
Delta R.T. -0.000 min
Lab File: BE101305.D
Acq: 19 Apr 2021 16:19

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.8


Tgt Ion:	212	Resp:	200616
Ion	Ratio	Lower	Upper
212	100		
106	3.3	2.6	4.0
104	1.8	1.4	2.2



Abundance Scan 1351 (19.174 min): BE101305.D\data.ms (-1316) (-)





Abundance Scan 1357 (19.209 min): BE101304.D\data.ms (-1323) (-)





Fluoranthene
Concen: 0.75 ng
RT: 19.203 min Scan# 1356
Delta R.T. -0.006 min
Lab File: BE101305.D
Acq: 19 Apr 2021 16:19

Tgt Ion:	202	Resp:	44636
Ion	Ratio	Lower	Upper
202	100		
101	11.6	9.2	13.8
203	17.4	13.7	20.5



Abundance Scan 1356 (19.203 min): BE101305.D\data.ms (-1322) (-)









Chrysene-d12
 Concen: 0.40 ng
 RT: 21.318 min Scan# 1600
 Delta R.T. 0.000 min
 Lab File: BE101305.D
 Acq: 19 Apr 2021 16:19




Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.8

Tgt Ion:	240	Resp:	20133
Ion	Ratio	Lower	Upper
240	100		
120	6.0	5.3	7.9
236	25.9	21.0	31.4



Pyrene
 Concen: 0.69 ng
 RT: 19.565 min Scan# 1419
 Delta R.T. -0.000 min
 Lab File: BE101305.D
 Acq: 19 Apr 2021 16:19


Tgt Ion:	202	Resp:	44476
Ion	Ratio	Lower	Upper
202	100		
200	20.7	16.5	24.7
203	17.5	13.8	20.8






Abundance Scan 1456 (19.778 min): BE101304.D\data.ms (-1430) (-)

Terphenyl-d14
Concen: 0.71 ng
RT: 19.772 min Scan# 1455
Delta R.T. -0.006 min
Lab File: BE101305.D
Acq: 19 Apr 2021 16:19

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.8







Tgt Ion	Ratio	Lower	Upper
244	100		
212	28.5	21.7	32.5
122	10.7	6.7	10.1#




Abundance Scan 1599 (21.306 min): BE101304.D\data.ms (-1582) (-)

Benzo(a)anthracene
Concen: 0.76 ng
RT: 21.306 min Scan# 1599
Delta R.T. 0.000 min
Lab File: BE101305.D
Acq: 19 Apr 2021 16:19


Tgt Ion	Ratio	Lower	Upper
228	100		
226	25.8	20.4	30.6
229	20.4	16.1	24.1








Scan 1603 (21.354 min): BE101304.D\data.ms (-1595) (-)
 Chrysene
 Concen: 0.78 ng
 RT: 21.354 min Scan# 1603
 Delta R.T. 0.000 min
 Lab File: BE101305.D
 Acq: 19 Apr 2021 16:19


Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.8





Tgt Ion:	228	Resp:	51849
Ion	Ratio	Lower	Upper
228	100		
226	29.1	23.7	35.5
229	19.7	15.8	23.8




Scan 1593 (21.233 min): BE101304.D\data.ms (-1595) (-)
 Bis(2-ethylhexyl)phthalate
 Concen: 0.60 ng
 RT: 21.233 min Scan# 1593
 Delta R.T. 0.000 min
 Lab File: BE101305.D
 Acq: 19 Apr 2021 16:19



Tgt Ion:	149	Resp:	50154
Ion	Ratio	Lower	Upper
149	100		
167	7.1	5.7	8.5
279	1.1	1.0	1.6





Abundance Scan 2866 (26.352 min): BE101304.D\data.ms (-2838) (-)




Indeno(1,2,3-cd)pyrene
Concen: 0.80 ng
RT: 26.348 min Scan# 2865
Delta R.T. -0.003 min
Lab File: BE101305.D
Acq: 19 Apr 2021 16:19

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.8


Tgt Ion	Ratio	Lower	Upper
276	100		
138	22.6	16.6	24.8
277	24.9	17.6	26.4



Abundance Scan 2865 (26.348 min): BE101305.D\data.ms (-2838) (-)





Abundance Scan 2147 (23.773 min): BE101304.D\data.ms (-2126) (-)




Perylene-d12
Concen: 0.40 ng
RT: 23.770 min Scan# 2146
Delta R.T. -0.002 min
Lab File: BE101305.D
Acq: 19 Apr 2021 16:19


Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.3	20.1	30.1
265	24.0	19.4	29.0



Abundance Scan 2146 (23.770 min): BE101305.D\data.ms (-2105) (-)




Abundance Scan 1935 (23.017 min): BE101304.D\data.ms (-1935) (-)





Benzo(b)fluoranthene
Concen: 0.71 ng
RT: 23.015 min Scan# 1934
Delta R.T. -0.002 min
Lab File: BE101305.D
Acq: 19 Apr 2021 16:19


Tgt Ion:	252	Resp:	48188
Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.7	26.5
125	9.0	7.5	11.3



Abundance Scan 1934 (23.015 min): BE101305.D\data.ms (-1878) (-)




Abundance Scan 1949 (23.067 min): BE101304.D\data.ms (-1949) (-)





Benzo(k)fluoranthene
Concen: 0.76 ng
RT: 23.065 min Scan# 1948
Delta R.T. -0.002 min
Lab File: BE101305.D
Acq: 19 Apr 2021 16:19

Tgt Ion:	252	Resp:	53637
Ion	Ratio	Lower	Upper
252	100		
253	22.1	18.1	27.1
125	9.2	8.8	13.2




Abundance Scan 1948 (23.065 min): BE101305.D\data.ms (-1864) (-)





Abundance Scan 2115 (23.659 min): BE101304.D\data.ms (-2058) (-)

Benzo(a)pyrene
Concen: 0.71 ng
RT: 23.656 min Scan# 2114
Delta R.T. -0.002 min
Lab File: BE101305.D
Acq: 19 Apr 2021 16:19


Instrument :
BNA_E
ClientSampleId :
SSTDIC0.8



Tgt Ion:	252	Resp:	43530
Ion Ratio	Lower	Upper	
252	100		
253	22.3	17.9	26.9
125	10.5	8.5	12.7




Abundance Scan 2114 (23.656 min): BE101305.D\data.ms (-2058) (-)





Abundance Scan 2873 (26.377 min): BE101304.D\data.ms (-2851) (-)


Dibenzo(a,h)anthracene
Concen: 0.74 ng
RT: 26.377 min Scan# 2873
Delta R.T. 0.000 min
Lab File: BE101305.D
Acq: 19 Apr 2021 16:19




Tgt Ion:	278	Resp:	44927
Ion Ratio	Lower	Upper	
278	100		
139	15.5	12.9	19.3
279	24.0	19.3	28.9



Abundance Scan 2873 (26.377 min): BE101305.D\data.ms (-2851) (-)





Abundance Scan 3089 (27.147 min): BE101304.D\data.ms (-3033) (-)



Benzo(g,h,i)perylene
Concen: 0.73 ng
RT: 27.144 min Scan# 3088
Delta R.T. -0.003 min
Lab File: BE101305.D
Acq: 19 Apr 2021 16:19

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.8

Tgt Ion:	276	Resp:	45132
Ion	Ratio	Lower	Upper
276	100		
277	23.8	19.3	28.9
138	19.1	15.8	23.8



Abundance Scan 3088 (27.144 min): BE101305.D\data.ms (-3033) (-)

