

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE041921\
 Data File : BE101306.D
 Acq On : 19 Apr 2021 16:55
 Operator : CG/JU
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC1.6

Quant Time: Apr 19 17:40:18 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE041921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 19 16:27:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.782	152	2823	0.40	ng	-0.01
7) Naphthalene-d8	10.575	136	11520	0.40	ng	0.00
13) Acenaphthene-d10	14.414	164	7532	0.40	ng	0.00
19) Phenanthrene-d10	17.166	188	16852	0.40	ng	0.00
29) Chrysene-d12	21.317	240	22578	0.40	ng	0.00
36) Perylene-d12	23.770	264	21437	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.375	112	12495	1.49	ng	-0.01
5) Phenol-d6	6.964	99	18435	1.45	ng	0.00
8) Nitrobenzene-d5	8.925	82	16089	1.51	ng	0.00
11) 2-Methylnaphthalene-d10	12.152	152	41602	1.57	ng	0.00
14) 2,4,6-Tribromophenol	15.911	330	3302	1.48	ng	0.00
15) 2-Fluorobiphenyl	13.030	172	43387	1.65	ng	0.00
27) Fluoranthene-d10	19.175	212	471546	1.60	ng	0.00
31) Terphenyl-d14	19.773	244	82071	1.50	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.301	88	7504	1.55	ng	92
3) n-Nitrosodimethylamine	3.612	42	8368	1.51	ng	# 96
6) bis(2-Chloroethyl)ether	7.206	93	14814	1.54	ng	98
9) Naphthalene	10.614	128	200239	1.61	ng	100
10) Hexachlorobutadiene	10.913	225	8739	1.64	ng	97
12) 2-Methylnaphthalene	12.224	142	34759	1.58	ng	100
16) Acenaphthylene	14.125	152	49220	1.60	ng	99
17) Acenaphthene	14.471	154	33960	1.59	ng	97
18) Fluorene	15.473	166	44772	1.56	ng	99
20) 4,6-Dinitro-2-methylph...	15.548	198	3089	1.94	ng	# 57
21) 4-Bromophenyl-phenylether	16.365	248	14884	1.71	ng	# 94
22) Hexachlorobenzene	16.471	284	14022	1.67	ng	99
23) Atrazine	16.622	200	11492	1.49	ng	98
24) Pentachlorophenol	16.818	266	1218	1.78	ng	# 87
25) Phenanthrene	17.196	178	78394	1.60	ng	99
26) Anthracene	17.287	178	71653	1.63	ng	99
28) Fluoranthene	19.204	202	104280	1.63	ng	99
30) Pyrene	19.566	202	105993	1.46	ng	100
32) Benzo(a)anthracene	21.305	228	114300	1.62	ng	99
33) Chrysene	21.353	228	119000	1.60	ng	100
34) Bis(2-ethylhexyl)phtha...	21.244	149	138539	1.47	ng	100
35) Indeno(1,2,3-cd)pyrene	26.356	276	117199	1.60	ng	95
37) Benzo(b)fluoranthene	23.017	252	111740	1.55	ng	99
38) Benzo(k)fluoranthene	23.067	252	119262	1.59	ng	98
39) Benzo(a)pyrene	23.659	252	100884	1.56	ng	99
40) Dibenzo(a,h)anthracene	26.381	278	102377	1.59	ng	98
41) Benzo(g,h,i)perylene	27.144	276	102936	1.57	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE041921\
Data File : BE101306.D
Acq On : 19 Apr 2021 16:55
Operator : CG/JU
Sample : SSTDICC1.6
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6

Quant Time: Apr 19 17:40:18 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE041921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 19 16:27:30 2021
Response via : Initial Calibration









1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.782 min Scan# 427
 Delta R.T. -0.010 min
 Lab File: BE101306.D
 Acq: 19 Apr 2021 16:55




Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:	152	Resp:	2823
Ion	Ratio	Lower	Upper
152	100		
150	160.3	124.8	187.2
115	69.5	48.7	73.1



1,4-Dioxane
 Concen: 1.55 ng
 RT: 3.301 min Scan# 38
 Delta R.T. -0.011 min
 Lab File: BE101306.D
 Acq: 19 Apr 2021 16:55

Tgt Ion:	88	Resp:	7504
Ion	Ratio	Lower	Upper
88	100		
58	89.8	65.4	98.0
43	34.1	24.4	36.6









#3
n-Nitrosodimethylamine
Concen: 1.51 ng
RT: 3.612 min Scan# 65
Delta R.T. -0.011 min
Lab File: BE101306.D
Acq: 19 Apr 2021 16:55




Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6


Tgt Ion:	42	Resp:	8368
Ion	Ratio	Lower	Upper
42	100		
74	117.2	90.9	136.3
44	7.9	9.2	13.8#



#4
2-Fluorophenol
Concen: 1.49 ng
RT: 5.375 min Scan# 218
Delta R.T. -0.011 min
Lab File: BE101306.D
Acq: 19 Apr 2021 16:55

Tgt Ion:	112	Resp:	12495
Ion	Ratio	Lower	Upper
112	100		
64	75.9	62.8	94.2
63	150.2	120.8	181.2









Phenol-d6
Concen: 1.45 ng
RT: 6.964 min Scan# 356
Delta R.T. 0.001 min
Lab File: BE101306.D
Acq: 19 Apr 2021 16:55




Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6

Tgt Ion:	99	Resp:	18435
Ion	Ratio	Lower	Upper
99	100		
42	23.2	18.8	28.2
71	33.9	26.6	40.0




bis(2-Chloroethyl)ether
Concen: 1.54 ng
RT: 7.206 min Scan# 377
Delta R.T. -0.011 min
Lab File: BE101306.D
Acq: 19 Apr 2021 16:55

Tgt Ion:	93	Resp:	14814
Ion	Ratio	Lower	Upper
93	100		
63	366.5	296.7	445.1
95	32.7	25.9	38.9






Abundance Scan 647 (10.576 min): BE101304.D\data.ms (-647) (-)




Naphthalene-d8
Concen: 0.40 ng
RT: 10.575 min Scan# 647
Delta R.T. -0.001 min
Lab File: BE101306.D
Acq: 19 Apr 2021 16:55

Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6

Tgt Ion:	136	Resp:	11520
Ion	Ratio	Lower	Upper
136	100		
137	11.3	9.0	13.6
54	26.7	21.3	31.9
68	5.5	4.2	6.4





Abundance Scan 520 (8.926 min): BE101304.D\data.ms (-517) (-)




Nitrobenzene-d5
Concen: 1.51 ng
RT: 8.925 min Scan# 520
Delta R.T. -0.001 min
Lab File: BE101306.D
Acq: 19 Apr 2021 16:55


Tgt Ion:	82	Resp:	16089
Ion	Ratio	Lower	Upper
82	100		
128	106.8	91.2	136.8
54	250.8	204.0	306.0



Abundance Scan 520 (8.925 min): BE101306.D\data.ms (-507) (-)





Abundance Scan 650 (10.615 min): BE101304.D\data.ms (-642) (-)




Naphthalene
Concen: 1.61 ng
RT: 10.614 min Scan# 650
Delta R.T. -0.001 min
Lab File: BE101306.D
Acq: 19 Apr 2021 16:55

Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6


Tgt Ion:	128	Resp:	200239
Ion	Ratio	Lower	Upper
128	100		
129	2.7	2.3	3.5
127	3.2	2.6	4.0



Abundance Scan 650 (10.614 min): BE101306.D\data.ms (-642) (-)





Abundance Scan 673 (10.914 min): BE101304.D\data.ms (-667) (-)





Hexachlorobutadiene
Concen: 1.64 ng
RT: 10.913 min Scan# 673
Delta R.T. -0.001 min
Lab File: BE101306.D
Acq: 19 Apr 2021 16:55

Tgt Ion:	225	Resp:	8739
Ion	Ratio	Lower	Upper
225	100		
223	63.3	49.4	74.0
227	64.5	49.7	74.5



Abundance Scan 673 (10.913 min): BE101306.D\data.ms (-665) (-)









2-Methylnaphthalene-d10
Concen: 1.57 ng
RT: 12.152 min Scan# 760
Delta R.T. -0.001 min
Lab File: BE101306.D
Acq: 19 Apr 2021 16:55




Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6


Tgt Ion:152 Resp: 41602
Ion Ratio Lower Upper
152 100
151 20.4 16.2 24.4



2-Methylnaphthalene
Concen: 1.58 ng
RT: 12.224 min Scan# 765
Delta R.T. -0.001 min
Lab File: BE101306.D
Acq: 19 Apr 2021 16:55

Tgt Ion:142 Resp: 34759
Ion Ratio Lower Upper
142 100
141 89.2 71.4 107.0
115 33.6 27.0 40.6









Acenaphthene-d10
 Concen: 0.40 ng
 RT: 14.414 min Scan# 917
 Delta R.T. -0.001 min
 Lab File: BE101306.D
 Acq: 19 Apr 2021 16:55




Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:	164	Resp:	7532
Ion	Ratio	Lower	Upper
164	100		
162	98.4	78.4	117.6
160	41.9	33.3	49.9



2,4,6-Tribromophenol
 Concen: 1.48 ng
 RT: 15.911 min Scan# 1013
 Delta R.T. 0.001 min
 Lab File: BE101306.D
 Acq: 19 Apr 2021 16:55

Tgt Ion:	330	Resp:	3302
Ion	Ratio	Lower	Upper
330	100		
332	95.7	79.7	119.5
141	39.0	33.2	49.8









2-Fluorobiphenyl
Concen: 1.65 ng
RT: 13.030 min Scan# 821
Delta R.T. -0.001 min
Lab File: BE101306.D
Acq: 19 Apr 2021 16:55




Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6


Tgt Ion:	172	Resp:	43387
Ion	Ratio	Lower	Upper
172	100		
171	36.0	29.0	43.6
170	24.6	19.1	28.7



Acenaphthylene
Concen: 1.60 ng
RT: 14.125 min Scan# 897
Delta R.T. -0.001 min
Lab File: BE101306.D
Acq: 19 Apr 2021 16:55

Tgt Ion:	152	Resp:	49220
Ion	Ratio	Lower	Upper
152	100		
151	19.2	15.6	23.4
153	13.0	10.7	16.1









Acenaphthene
Concen: 1.59 ng
RT: 14.471 min Scan# 921
Delta R.T. -0.001 min
Lab File: BE101306.D
Acq: 19 Apr 2021 16:55




Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6


Tgt Ion:154 Resp: 33960
Ion Ratio Lower Upper
154 100
153 108.7 90.1 135.1
152 54.0 44.2 66.2



Fluorene
Concen: 1.56 ng
RT: 15.473 min Scan# 984
Delta R.T. 0.001 min
Lab File: BE101306.D
Acq: 19 Apr 2021 16:55

Tgt Ion:166 Resp: 44772
Ion Ratio Lower Upper
166 100
165 98.0 78.9 118.3
167 13.3 11.0 16.4









Scan 1096 (17.165 min): BE101304.D\data.ms (-1096) (-)
 Phenanthrene-d10
 Concen: 0.40 ng
 RT: 17.166 min Scan# 1096
 Delta R.T. 0.001 min
 Lab File: BE101306.D
 Acq: 19 Apr 2021 16:55




Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC1.6

Tgt Ion:	188	Resp:	16852
Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	7.3	5.8	8.8



Scan 989 (15.547 min): BE101304.D\data.ms (-989) (-)
 4,6-Dinitro-2-methylphenol
 Concen: 1.94 ng
 RT: 15.548 min Scan# 989
 Delta R.T. 0.001 min
 Lab File: BE101306.D
 Acq: 19 Apr 2021 16:55


Tgt Ion:	198	Resp:	3089
Ion	Ratio	Lower	Upper
198	100		
51	29.7	56.0	84.0#
105	24.0	36.7	55.1#






Abundance Scan 1043 (16.364 min): BE101304.D\data.ms (-1042) (-)

4-Bromophenyl-phenylether
Concen: 1.71 ng
RT: 16.365 min Scan# 1043
Delta R.T. 0.001 min
Lab File: BE101306.D
Acq: 19 Apr 2021 16:55

Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6




Tgt Ion:	248	Resp:	14884
Ion	Ratio	Lower	Upper
248	100		
250	101.2	79.2	118.8
141	29.5	31.1	46.7#






Abundance Scan 1050 (16.470 min): BE101304.D\data.ms (-1042) (-)


Hexachlorobenzene
Concen: 1.67 ng
RT: 16.471 min Scan# 1050
Delta R.T. 0.001 min
Lab File: BE101306.D
Acq: 19 Apr 2021 16:55



Tgt Ion:	284	Resp:	14022
Ion	Ratio	Lower	Upper
284	100		
142	34.2	27.2	40.8
249	33.8	26.6	40.0






Abundance Scan 1060 (16.621 min): BE101304.D\data.ms (-1023) (-)




Atrazine
Concen: 1.49 ng
RT: 16.622 min Scan# 1060
Delta R.T. 0.001 min
Lab File: BE101306.D
Acq: 19 Apr 2021 16:55

Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6

Tgt Ion:	200	Resp:	11492
Ion Ratio	Lower	Upper	
200	100		
173	34.4	27.8	41.8
215	48.0	39.8	59.6







Abundance Scan 1073 (16.818 min): BE101304.D\data.ms (-1024) (-)



Pentachlorophenol
Concen: 1.78 ng
RT: 16.818 min Scan# 1073
Delta R.T. 0.001 min
Lab File: BE101306.D
Acq: 19 Apr 2021 16:55

Tgt Ion:	266	Resp:	1218
Ion Ratio	Lower	Upper	
266	100		
264	64.0	67.0	100.6#
268	66.1	54.1	81.1









Scan 1098 (17.196 min): BE101304.D\data.ms (-1025) (-)
 Phenanthrene
 Concen: 1.60 ng
 RT: 17.196 min Scan# 1098
 Delta R.T. 0.001 min
 Lab File: BE101306.D
 Acq: 19 Apr 2021 16:55




Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC1.6

Tgt Ion:	178	Resp:	78394
Ion	Ratio	Lower	Upper
178	100		
176	19.1	15.0	22.4
179	15.2	12.6	19.0




Scan 1105 (17.301 min): BE101304.D\data.ms (-1025) (-)
 Anthracene
 Concen: 1.63 ng
 RT: 17.287 min Scan# 1104
 Delta R.T. -0.014 min
 Lab File: BE101306.D
 Acq: 19 Apr 2021 16:55

Tgt Ion:	178	Resp:	71653
Ion	Ratio	Lower	Upper
178	100		
176	18.3	14.5	21.7
179	15.4	12.1	18.1






Abundance Scan 1351 (19.175 min): BE101304.D\data.ms (-1323) (-)




Fluoranthene-d10
Concen: 1.60 ng
RT: 19.175 min Scan# 1351
Delta R.T. 0.000 min
Lab File: BE101306.D
Acq: 19 Apr 2021 16:55

Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6

Tgt Ion:	212	Resp:	471546
Ion	Ratio	Lower	Upper
212	100		
106	3.4	2.6	4.0
104	1.8	1.4	2.2







Abundance Scan 1357 (19.209 min): BE101304.D\data.ms (-1323) (-)



Fluoranthene
Concen: 1.63 ng
RT: 19.204 min Scan# 1356
Delta R.T. -0.005 min
Lab File: BE101306.D
Acq: 19 Apr 2021 16:55

Tgt Ion:	202	Resp:	104280
Ion	Ratio	Lower	Upper
202	100		
101	11.7	9.2	13.8
203	17.6	13.7	20.5









Scan 1600 (21.318 min): BE101304.D\data.ms (-1529) (-)
 Chrysene-d12
 Concen: 0.40 ng
 RT: 21.317 min Scan# 1600
 Delta R.T. -0.001 min
 Lab File: BE101306.D
 Acq: 19 Apr 2021 16:55




Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC1.6

Tgt Ion:	240	Resp:	22578
Ion	Ratio	Lower	Upper
240	100		
120	7.3	5.3	7.9
236	26.3	21.0	31.4



Scan 1419 (19.565 min): BE101304.D\data.ms (-1408) (-)
 Pyrene
 Concen: 1.46 ng
 RT: 19.566 min Scan# 1419
 Delta R.T. 0.000 min
 Lab File: BE101306.D
 Acq: 19 Apr 2021 16:55


Tgt Ion:	202	Resp:	105993
Ion	Ratio	Lower	Upper
202	100		
200	20.7	16.5	24.7
203	17.5	13.8	20.8






Abundance Scan 1456 (19.778 min): BE101304.D\data.ms (-1430) (-)

Terphenyl-d14
Concen: 1.50 ng
RT: 19.773 min Scan# 1455
Delta R.T. -0.005 min
Lab File: BE101306.D
Acq: 19 Apr 2021 16:55

Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6




Tgt Ion:	244	Resp:	82071
Ion Ratio	Lower	Upper	
244	100		
212	28.2	21.7	32.5
122	10.9	6.7	10.1#







Abundance Scan 1599 (21.306 min): BE101304.D\data.ms (-1583) (-)

Benzo(a)anthracene
Concen: 1.62 ng
RT: 21.305 min Scan# 1599
Delta R.T. -0.001 min
Lab File: BE101306.D
Acq: 19 Apr 2021 16:55



Tgt Ion:	228	Resp:	114300
Ion Ratio	Lower	Upper	
228	100		
226	25.9	20.4	30.6
229	20.2	16.1	24.1









Chrysene
Concen: 1.60 ng
RT: 21.353 min Scan# 1603
Delta R.T. -0.001 min
Lab File: BE101306.D
Acq: 19 Apr 2021 16:55




Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6

Tgt Ion:	228	Resp:	119000
Ion	Ratio	Lower	Upper
228	100		
226	29.9	23.7	35.5
229	19.6	15.8	23.8



Bis(2-ethylhexyl)phthalate
Concen: 1.47 ng
RT: 21.244 min Scan# 1594
Delta R.T. 0.011 min
Lab File: BE101306.D
Acq: 19 Apr 2021 16:55


Tgt Ion:	149	Resp:	138539
Ion	Ratio	Lower	Upper
149	100		
167	7.1	5.7	8.5
279	1.1	1.0	1.6






Abundance Scan 2866 (26.352 min): BE101304.D\data.ms (-2835) (-)

Indeno(1,2,3-cd)pyrene
Concen: 1.60 ng
RT: 26.356 min Scan# 2867
Delta R.T. 0.004 min
Lab File: BE101306.D
Acq: 19 Apr 2021 16:55

Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6




Tgt Ion:	276	Resp:	117199
Ion	Ratio	Lower	Upper
276	100		
138	22.3	16.6	24.8
277	25.0	17.6	26.4






Abundance Scan 2147 (23.773 min): BE101304.D\data.ms (-2129) (-)

Perylene-d12
Concen: 0.40 ng
RT: 23.770 min Scan# 2146
Delta R.T. -0.003 min
Lab File: BE101306.D
Acq: 19 Apr 2021 16:55




Tgt Ion:	264	Resp:	21437
Ion	Ratio	Lower	Upper
264	100		
260	25.2	20.1	30.1
265	23.6	19.4	29.0






Abundance Scan 1935 (23.017 min): BE101304.D\data.ms (-1935) (-)

Benzo(b)fluoranthene
Concen: 1.55 ng
RT: 23.017 min Scan# 1935
Delta R.T. 0.000 min
Lab File: BE101306.D
Acq: 19 Apr 2021 16:55

Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6







Tgt Ion:	252	Resp:	111740
Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.7	26.5
125	8.8	7.5	11.3



Abundance Scan 1949 (23.067 min): BE101304.D\data.ms (-1949) (-)

Benzo(k)fluoranthene
Concen: 1.59 ng
RT: 23.067 min Scan# 1949
Delta R.T. 0.000 min
Lab File: BE101306.D
Acq: 19 Apr 2021 16:55


Tgt Ion:	252	Resp:	119262
Ion	Ratio	Lower	Upper
252	100		
253	22.3	18.1	27.1
125	8.9	8.8	13.2





Abundance Scan 2115 (23.659 min): BE101304.D\data.ms (-2059) (-)

Benzo(a)pyrene
Concen: 1.56 ng
RT: 23.659 min Scan# 2115
Delta R.T. 0.000 min
Lab File: BE101306.D
Acq: 19 Apr 2021 16:55


Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6



Tgt Ion:	252	Resp:	100884
Ion Ratio	Lower	Upper	
252	100		
253	21.9	17.9	26.9
125	10.2	8.5	12.7






Abundance Scan 2115 (23.659 min): BE101306.D\data.ms (-2059) (-)




Abundance Scan 2873 (26.377 min): BE101304.D\data.ms (-2848) (-)

Dibenzo(a,h)anthracene
Concen: 1.59 ng
RT: 26.381 min Scan# 2874
Delta R.T. 0.004 min
Lab File: BE101306.D
Acq: 19 Apr 2021 16:55


Tgt Ion:	278	Resp:	102377
Ion Ratio	Lower	Upper	
278	100		
139	14.6	12.9	19.3
279	23.9	19.3	28.9



Abundance Scan 2874 (26.381 min): BE101306.D\data.ms (-2850) (-)




Abundance Scan 3089 (27.147 min): BE101304.D\data.ms (-3033) (-)




Benzo(g,h,i)perylene
Concen: 1.57 ng
RT: 27.144 min Scan# 3088
Delta R.T. -0.003 min
Lab File: BE101306.D
Acq: 19 Apr 2021 16:55

Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6

Abundance Scan 3088 (27.144 min): BE101306.D\data.ms



Tgt Ion:	276	Resp:	102936
Ion	Ratio	Lower	Upper
276	100		
277	23.9	19.3	28.9
138	19.4	15.8	23.8



Abundance Scan 3088 (27.144 min): BE101306.D\data.ms (-3033) (-)

