

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE041921\
 Data File : BE101307.D
 Acq On : 19 Apr 2021 17:33
 Operator : CG/JU
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC3.2

Quant Time: Apr 19 18:17:45 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE041921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 19 16:27:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.784	152	2858	0.40	ng	0.00
7) Naphthalene-d8	10.575	136	11287	0.40	ng	0.00
13) Acenaphthene-d10	14.414	164	7275	0.40	ng	0.00
19) Phenanthrene-d10	17.166	188	16600	0.40	ng	0.00
29) Chrysene-d12	21.317	240	23178	0.40	ng	0.00
36) Perylene-d12	23.770	264	20993	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.377	112	25108	2.96	ng	0.00
5) Phenol-d6	6.955	99	38446	2.99	ng	0.00
8) Nitrobenzene-d5	8.925	82	33608	3.23	ng	0.00
11) 2-Methylnaphthalene-d10	12.152	152	83503	3.22	ng	0.00
14) 2,4,6-Tribromophenol	15.911	330	7874	3.66	ng	0.00
15) 2-Fluorobiphenyl	13.031	172	85087	3.36	ng	0.00
27) Fluoranthene-d10	19.175	212	994233	3.42	ng	0.00
31) Terphenyl-d14	19.773	244	173748	3.09	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.304	88	14455	2.95	ng	91
3) n-Nitrosodimethylamine	3.604	42	16909	3.01	ng	# 96
6) bis(2-Chloroethyl)ether	7.209	93	29446	3.02	ng	98
9) Naphthalene	10.614	128	394222	3.23	ng	100
10) Hexachlorobutadiene	10.913	225	17247	3.31	ng	98
12) 2-Methylnaphthalene	12.224	142	69797	3.25	ng	99
16) Acenaphthylene	14.126	152	103436	3.48	ng	100
17) Acenaphthene	14.471	154	68121	3.29	ng	96
18) Fluorene	15.473	166	90381	3.27	ng	99
20) 4,6-Dinitro-2-methylph...	15.548	198	8885	5.66	ng	# 53
21) 4-Bromophenyl-phenylether	16.365	248	29555	3.44	ng	# 93
22) Hexachlorobenzene	16.470	284	28373	3.44	ng	98
23) Atrazine	16.622	200	26047	3.43	ng	96
24) Pentachlorophenol	16.818	266	4340	6.43	ng	# 84
25) Phenanthrene	17.196	178	160240	3.31	ng	99
26) Anthracene	17.287	178	150787	3.49	ng	99
28) Fluoranthene	19.204	202	219555	3.48	ng	99
30) Pyrene	19.566	202	222207	2.98	ng	99
32) Benzo(a)anthracene	21.305	228	238306	3.28	ng	99
33) Chrysene	21.354	228	241947	3.17	ng	100
34) Bis(2-ethylhexyl)phtha...	21.233	149	332145	3.43	ng	100
35) Indeno(1,2,3-cd)pyrene	26.352	276	237714	3.16	ng	95
37) Benzo(b)fluoranthene	23.018	252	225358	3.20	ng	99
38) Benzo(k)fluoranthene	23.068	252	237785	3.24	ng	98
39) Benzo(a)pyrene	23.659	252	204374	3.22	ng	99
40) Dibenzo(a,h)anthracene	26.377	278	207804	3.29	ng	98
41) Benzo(g,h,i)perylene	27.148	276	206293	3.20	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE041921\
Data File : BE101307.D
Acq On : 19 Apr 2021 17:33
Operator : CG/JU
Sample : SSTDIC3.2
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDICC3.2

Quant Time: Apr 19 18:17:45 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE041921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 19 16:27:30 2021
Response via : Initial Calibration









1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.784 min Scan# 427
 Delta R.T. -0.008 min
 Lab File: BE101307.D
 Acq: 19 Apr 2021 17:33




Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC3.2


Tgt Ion:	152	Resp:	2858
Ion	Ratio	Lower	Upper
152	100		
150	154.6	124.8	187.2
115	64.6	48.7	73.1



1,4-Dioxane
 Concen: 2.95 ng
 RT: 3.304 min Scan# 38
 Delta R.T. -0.008 min
 Lab File: BE101307.D
 Acq: 19 Apr 2021 17:33

Tgt Ion:	88	Resp:	14455
Ion	Ratio	Lower	Upper
88	100		
58	90.5	65.4	98.0
43	34.0	24.4	36.6









n-Nitrosodimethylamine
Concen: 3.01 ng
RT: 3.604 min Scan# 64
Delta R.T. -0.019 min
Lab File: BE101307.D
Acq: 19 Apr 2021 17:33




Instrument :
BNA_E
ClientSampleId :
SSTDICC3.2


Tgt Ion:	42	Resp:	16909
Ion	Ratio	Lower	Upper
42	100		
74	117.0	90.9	136.3
44	7.4	9.2	13.8#



2-Fluorophenol
Concen: 2.96 ng
RT: 5.377 min Scan# 218
Delta R.T. -0.008 min
Lab File: BE101307.D
Acq: 19 Apr 2021 17:33

Tgt Ion:	112	Resp:	25108
Ion	Ratio	Lower	Upper
112	100		
64	76.3	62.8	94.2
63	148.0	120.8	181.2









Phenol-d6
Concen: 2.99 ng
RT: 6.955 min Scan# 355
Delta R.T. -0.008 min
Lab File: BE101307.D
Acq: 19 Apr 2021 17:33




Instrument :
BNA_E
ClientSampleId :
SSTDICC3.2

Tgt Ion:	99	Resp:	38446
Ion	Ratio	Lower	Upper
99	100		
42	22.6	18.8	28.2
71	33.7	26.6	40.0




bis(2-Chloroethyl)ether
Concen: 3.02 ng
RT: 7.209 min Scan# 377
Delta R.T. -0.008 min
Lab File: BE101307.D
Acq: 19 Apr 2021 17:33

Tgt Ion:	93	Resp:	29446
Ion	Ratio	Lower	Upper
93	100		
63	367.1	296.7	445.1
95	32.5	25.9	38.9





Abundance Scan 647 (10.576 min): BE101304.D\data.ms (-647) (-)




Naphthalene-d8
Concen: 0.40 ng
RT: 10.575 min Scan# 647
Delta R.T. -0.001 min
Lab File: BE101307.D
Acq: 19 Apr 2021 17:33

Instrument :
BNA_E
ClientSampleId :
SSTDICC3.2


Tgt Ion:	136	Resp:	11287
Ion	Ratio	Lower	Upper
136	100		
137	11.6	9.0	13.6
54	25.1	21.3	31.9
68	5.8	4.2	6.4



Abundance Scan 647 (10.575 min): BE101307.D\data.ms (-632) (-)





Abundance Scan 520 (8.926 min): BE101304.D\data.ms (-517) (-)




Nitrobenzene-d5
Concen: 3.23 ng
RT: 8.925 min Scan# 520
Delta R.T. -0.001 min
Lab File: BE101307.D
Acq: 19 Apr 2021 17:33


Tgt Ion:	82	Resp:	33608
Ion	Ratio	Lower	Upper
82	100		
128	108.8	91.2	136.8
54	250.1	204.0	306.0



Abundance Scan 520 (8.925 min): BE101307.D\data.ms (-507) (-)






Abundance Scan 650 (10.615 min): BE101304.D\data.ms (-642) (-)




Naphthalene
Concen: 3.23 ng
RT: 10.614 min Scan# 650
Delta R.T. -0.001 min
Lab File: BE101307.D
Acq: 19 Apr 2021 17:33

Instrument :
BNA_E
ClientSampleId :
SSTDICC3.2

Tgt Ion:	128	Resp:	394222
Ion	Ratio	Lower	Upper
128	100		
129	2.8	2.3	3.5
127	3.3	2.6	4.0







Abundance Scan 673 (10.914 min): BE101304.D\data.ms (-667) (-)



Hexachlorobutadiene
Concen: 3.31 ng
RT: 10.913 min Scan# 673
Delta R.T. -0.001 min
Lab File: BE101307.D
Acq: 19 Apr 2021 17:33

Tgt Ion:	225	Resp:	17247
Ion	Ratio	Lower	Upper
225	100		
223	62.4	49.4	74.0
227	64.1	49.7	74.5









2-Methylnaphthalene-d10
Concen: 3.22 ng
RT: 12.152 min Scan# 760
Delta R.T. -0.001 min
Lab File: BE101307.D
Acq: 19 Apr 2021 17:33




Instrument :
BNA_E
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 83503
Ion Ratio Lower Upper
152 100
151 20.5 16.2 24.4



2-Methylnaphthalene
Concen: 3.25 ng
RT: 12.224 min Scan# 765
Delta R.T. -0.001 min
Lab File: BE101307.D
Acq: 19 Apr 2021 17:33

Tgt Ion:142 Resp: 69797
Ion Ratio Lower Upper
142 100
141 87.7 71.4 107.0
115 33.1 27.0 40.6









Acenaphthene-d10
 Concen: 0.40 ng
 RT: 14.414 min Scan# 917
 Delta R.T. -0.001 min
 Lab File: BE101307.D
 Acq: 19 Apr 2021 17:33




Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC3.2


Tgt Ion:	164	Resp:	7275
Ion	Ratio	Lower	Upper
164	100		
162	98.1	78.4	117.6
160	41.6	33.3	49.9



2,4,6-Tribromophenol
 Concen: 3.66 ng
 RT: 15.911 min Scan# 1013
 Delta R.T. 0.001 min
 Lab File: BE101307.D
 Acq: 19 Apr 2021 17:33

Tgt Ion:	330	Resp:	7874
Ion	Ratio	Lower	Upper
330	100		
332	98.4	79.7	119.5
141	39.1	33.2	49.8









2-Fluorobiphenyl
Concen: 3.36 ng
RT: 13.031 min Scan# 821
Delta R.T. -0.001 min
Lab File: BE101307.D
Acq: 19 Apr 2021 17:33




Instrument :
BNA_E
ClientSampleId :
SSTDICC3.2


Tgt Ion:	172	Resp:	85087
Ion	Ratio	Lower	Upper
172	100		
171	35.0	29.0	43.6
170	23.6	19.1	28.7



Acenaphthylene
Concen: 3.48 ng
RT: 14.126 min Scan# 897
Delta R.T. -0.001 min
Lab File: BE101307.D
Acq: 19 Apr 2021 17:33

Tgt Ion:	152	Resp:	103436
Ion	Ratio	Lower	Upper
152	100		
151	19.4	15.6	23.4
153	13.1	10.7	16.1









Acenaphthene
Concen: 3.29 ng
RT: 14.471 min Scan# 921
Delta R.T. -0.001 min
Lab File: BE101307.D
Acq: 19 Apr 2021 17:33




Instrument :
BNA_E
ClientSampleId :
SSTDICC3.2


Tgt Ion:	154	Resp:	68121
Ion	Ratio	Lower	Upper
154	100		
153	108.5	90.1	135.1
152	52.7	44.2	66.2



Fluorene
Concen: 3.27 ng
RT: 15.473 min Scan# 984
Delta R.T. 0.001 min
Lab File: BE101307.D
Acq: 19 Apr 2021 17:33

Tgt Ion:	166	Resp:	90381
Ion	Ratio	Lower	Upper
166	100		
165	97.5	78.9	118.3
167	13.5	11.0	16.4









Phenanthrene-d10
 Concen: 0.40 ng
 RT: 17.166 min Scan# 1096
 Delta R.T. 0.001 min
 Lab File: BE101307.D
 Acq: 19 Apr 2021 17:33




Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC3.2

Tgt Ion:	188	Resp:	16600
Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	7.3	5.8	8.8



4,6-Dinitro-2-methylphenol
 Concen: 5.66 ng
 RT: 15.548 min Scan# 989
 Delta R.T. 0.001 min
 Lab File: BE101307.D
 Acq: 19 Apr 2021 17:33


Tgt Ion:	198	Resp:	8885
Ion	Ratio	Lower	Upper
198	100		
51	25.7	56.0	84.0#
105	21.6	36.7	55.1#





Abundance Scan 1043 (16.364 min): BE101304.D\data.ms (-1042) (-)

4-Bromophenyl-phenylether
Concen: 3.44 ng
RT: 16.365 min Scan# 1043
Delta R.T. 0.001 min
Lab File: BE101307.D
Acq: 19 Apr 2021 17:33


Instrument :
BNA_E
ClientSampleId :
SSTDICC3.2



Tgt Ion:	248	Resp:	29555
Ion Ratio	Lower	Upper	
248	100		
250	103.1	79.2	118.8
141	29.7	31.1	46.7#




Abundance Scan 1043 (16.365 min): BE101307.D\data.ms (-1036) (-)





Abundance Scan 1050 (16.470 min): BE101304.D\data.ms (-1042) (-)


Hexachlorobenzene
Concen: 3.44 ng
RT: 16.470 min Scan# 1050
Delta R.T. 0.001 min
Lab File: BE101307.D
Acq: 19 Apr 2021 17:33




Tgt Ion:	284	Resp:	28373
Ion Ratio	Lower	Upper	
284	100		
142	32.6	27.2	40.8
249	33.9	26.6	40.0



Abundance Scan 1050 (16.470 min): BE101307.D\data.ms (-1043) (-)





Abundance Scan 1060 (16.622 min): BE101304.D\data.ms (-1023) (-)




Atrazine
Concen: 3.43 ng
RT: 16.622 min Scan# 1060
Delta R.T. 0.001 min
Lab File: BE101307.D
Acq: 19 Apr 2021 17:33

Instrument :
BNA_E
ClientSampleId :
SSTDICC3.2


Tgt Ion:	200	Resp:	26047
Ion Ratio	Lower	Upper	
200	100		
173	32.9	27.8	41.8
215	46.5	39.8	59.6



Abundance Scan 1060 (16.622 min): BE101307.D\data.ms (-1027) (-)





Abundance Scan 1073 (16.818 min): BE101304.D\data.ms (-1024) (-)





Pentachlorophenol
Concen: 6.43 ng
RT: 16.818 min Scan# 1073
Delta R.T. 0.001 min
Lab File: BE101307.D
Acq: 19 Apr 2021 17:33

Tgt Ion:	266	Resp:	4340
Ion Ratio	Lower	Upper	
266	100		
264	63.5	67.0	100.6#
268	61.8	54.1	81.1



Abundance Scan 1073 (16.818 min): BE101307.D\data.ms (-1066) (-)









Scan 1098 (17.196 min): BE101304.D\data.ms (-1025) (-)
 Phenanthrene
 Concen: 3.31 ng
 RT: 17.196 min Scan# 1098
 Delta R.T. 0.001 min
 Lab File: BE101307.D
 Acq: 19 Apr 2021 17:33




Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC3.2

Tgt Ion:	178	Resp:	160240
Ion	Ratio	Lower	Upper
178	100		
176	19.0	15.0	22.4
179	15.4	12.6	19.0




Scan 1105 (17.301 min): BE101304.D\data.ms (-1125) (-)
 Anthracene
 Concen: 3.49 ng
 RT: 17.287 min Scan# 1104
 Delta R.T. -0.015 min
 Lab File: BE101307.D
 Acq: 19 Apr 2021 17:33

Tgt Ion:	178	Resp:	150787
Ion	Ratio	Lower	Upper
178	100		
176	18.3	14.5	21.7
179	15.4	12.1	18.1






Abundance Scan 1351 (19.175 min): BE101304.D\data.ms (-1323) (-)




Fluoranthene-d10
Concen: 3.42 ng
RT: 19.175 min Scan# 1351
Delta R.T. 0.001 min
Lab File: BE101307.D
Acq: 19 Apr 2021 17:33

Instrument :
BNA_E
ClientSampleId :
SSTDICC3.2

Tgt Ion:	212	Resp:	994233
Ion	Ratio	Lower	Upper
212	100		
106	3.4	2.6	4.0
104	1.9	1.4	2.2







Abundance Scan 1357 (19.209 min): BE101304.D\data.ms (-1323) (-)



Fluoranthene
Concen: 3.48 ng
RT: 19.204 min Scan# 1356
Delta R.T. -0.005 min
Lab File: BE101307.D
Acq: 19 Apr 2021 17:33

Tgt Ion:	202	Resp:	219555
Ion	Ratio	Lower	Upper
202	100		
101	11.8	9.2	13.8
203	17.6	13.7	20.5









Chrysene-d12
 Concen: 0.40 ng
 RT: 21.317 min Scan# 1600
 Delta R.T. -0.000 min
 Lab File: BE101307.D
 Acq: 19 Apr 2021 17:33




Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC3.2

Tgt Ion:	240	Resp:	23178
Ion	Ratio	Lower	Upper
240	100		
120	7.0	5.3	7.9
236	26.0	21.0	31.4




Pyrene
 Concen: 2.98 ng
 RT: 19.566 min Scan# 1419
 Delta R.T. 0.001 min
 Lab File: BE101307.D
 Acq: 19 Apr 2021 17:33

Tgt Ion:	202	Resp:	222207
Ion	Ratio	Lower	Upper
202	100		
200	21.0	16.5	24.7
203	17.8	13.8	20.8





Abundance Scan 1456 (19.778 min): BE101304.D\data.ms (-1430) (-)




Terphenyl-d14
Concen: 3.09 ng
RT: 19.773 min Scan# 1455
Delta R.T. -0.005 min
Lab File: BE101307.D
Acq: 19 Apr 2021 17:33

Instrument :
BNA_E
ClientSampleId :
SSTDICC3.2


Tgt Ion:	244	Resp:	173748
Ion	Ratio	Lower	Upper
244	100		
212	28.5	21.7	32.5
122	11.5	6.7	10.1#



Abundance Scan 1455 (19.773 min): BE101307.D\data.ms (-1430) (-)





Abundance Scan 1599 (21.306 min): BE101304.D\data.ms (-1582) (-)





Benzo(a)anthracene
Concen: 3.28 ng
RT: 21.305 min Scan# 1599
Delta R.T. -0.000 min
Lab File: BE101307.D
Acq: 19 Apr 2021 17:33

Tgt Ion:	228	Resp:	238306
Ion	Ratio	Lower	Upper
228	100		
226	26.0	20.4	30.6
229	20.5	16.1	24.1




Abundance Scan 1599 (21.305 min): BE101307.D\data.ms (-1583) (-)








Scan 1603 (21.354 min): BE101304.D\data.ms (-1603) (-)
 Chrysene
 Concen: 3.17 ng
 RT: 21.354 min Scan# 1603
 Delta R.T. -0.000 min
 Lab File: BE101307.D
 Acq: 19 Apr 2021 17:33

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC3.2






Tgt Ion:	228	Resp:	241947
Ion	Ratio	Lower	Upper
228	100		
226	29.7	23.7	35.5
229	20.1	15.8	23.8



Scan 1593 (21.233 min): BE101304.D\data.ms (-1593) (-)
 Bis(2-ethylhexyl)phthalate
 Concen: 3.43 ng
 RT: 21.233 min Scan# 1593
 Delta R.T. -0.000 min
 Lab File: BE101307.D
 Acq: 19 Apr 2021 17:33


Tgt Ion:	149	Resp:	332145
Ion	Ratio	Lower	Upper
149	100		
167	7.0	5.7	8.5
279	1.2	1.0	1.6






Abundance Scan 2866 (26.352 min): BE101304.D\data.ms (-2835) (-)

Indeno(1,2,3-cd)pyrene
Concen: 3.16 ng
RT: 26.352 min Scan# 2866
Delta R.T. 0.001 min
Lab File: BE101307.D
Acq: 19 Apr 2021 17:33

Instrument :
BNA_E
ClientSampleId :
SSTDICC3.2




Tgt Ion	Ratio	Lower	Upper
276	100		
138	22.4	16.6	24.8
277	25.2	17.6	26.4






Abundance Scan 2147 (23.773 min): BE101304.D\data.ms (-2126) (-)


Perylene-d12
Concen: 0.40 ng
RT: 23.770 min Scan# 2146
Delta R.T. -0.003 min
Lab File: BE101307.D
Acq: 19 Apr 2021 17:33



Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.4	20.1	30.1
265	24.6	19.4	29.0





Abundance Scan 1935 (23.017 min): BE101304.D\data.ms (-1879) (-)




Benzo(b)fluoranthene
Concen: 3.20 ng
RT: 23.018 min Scan# 1935
Delta R.T. 0.001 min
Lab File: BE101307.D
Acq: 19 Apr 2021 17:33


Tgt Ion:	252	Resp:	225358
Ion	Ratio	Lower	Upper
252	100		
253	22.5	17.7	26.5
125	9.0	7.5	11.3



Abundance Scan 1935 (23.018 min): BE101307.D\data.ms (-1879) (-)





Abundance Scan 1949 (23.067 min): BE101304.D\data.ms (-1879) (-)




Benzo(k)fluoranthene
Concen: 3.24 ng
RT: 23.068 min Scan# 1949
Delta R.T. 0.001 min
Lab File: BE101307.D
Acq: 19 Apr 2021 17:33

Tgt Ion:	252	Resp:	237785
Ion	Ratio	Lower	Upper
252	100		
253	22.4	18.1	27.1
125	9.0	8.8	13.2



Abundance Scan 1949 (23.068 min): BE101307.D\data.ms (-1865) (-)




Instrument :
BNA_E
ClientSampleId :
SSTDICC3.2




Abundance Scan 2115 (23.659 min): BE101304.D\data.ms (-2059) (-)

Benzo(a)pyrene
Concen: 3.22 ng
RT: 23.659 min Scan# 2115
Delta R.T. 0.001 min
Lab File: BE101307.D
Acq: 19 Apr 2021 17:33

Instrument :
BNA_E
ClientSampleId :
SSTDICC3.2




Tgt Ion:	252	Resp:	204374
Ion Ratio	Lower	Upper	
252	100		
253	22.2	17.9	26.9
125	9.8	8.5	12.7






Abundance Scan 2873 (26.377 min): BE101304.D\data.ms (-2850) (-)


Dibenzo(a,h)anthracene
Concen: 3.29 ng
RT: 26.377 min Scan# 2873
Delta R.T. 0.001 min
Lab File: BE101307.D
Acq: 19 Apr 2021 17:33



Tgt Ion:	278	Resp:	207804
Ion Ratio	Lower	Upper	
278	100		
139	14.4	12.9	19.3
279	23.7	19.3	28.9




Abundance Scan 3089 (27.147 min): BE101304.D\data.ms (-3033) (-)




Benzo(g,h,i)perylene
Concen: 3.20 ng
RT: 27.148 min Scan# 3089
Delta R.T. 0.001 min
Lab File: BE101307.D
Acq: 19 Apr 2021 17:33

Instrument :
BNA_E
ClientSampleId :
SSTDICC3.2

Abundance Scan 3089 (27.148 min): BE101307.D\data.ms



Tgt Ion:	276	Resp:	206293
Ion	Ratio	Lower	Upper
276	100		
277	23.6	19.3	28.9
138	19.1	15.8	23.8



Abundance Scan 3089 (27.148 min): BE101307.D\data.ms (-3033) (-)

