

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE041921\
 Data File : BE101308.D
 Acq On : 19 Apr 2021 18:09
 Operator : CG/JU
 Sample : SSTDICC5.0
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC5.0

Quant Time: Apr 19 18:53:04 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE041921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 19 16:27:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.782	152	3271	0.40	ng	-0.01
7) Naphthalene-d8	10.575	136	12376	0.40	ng	0.00
13) Acenaphthene-d10	14.414	164	8179	0.40	ng	0.00
19) Phenanthrene-d10	17.166	188	17949	0.40	ng	0.00
29) Chrysene-d12	21.318	240	24131	0.40	ng	0.00
36) Perylene-d12	23.769	264	22222	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.375	112	40666	4.19	ng	-0.01
5) Phenol-d6	6.964	99	62643	4.25	ng	0.00
8) Nitrobenzene-d5	8.925	82	56248	4.93	ng	0.00
11) 2-Methylnaphthalene-d10	12.152	152	133411	4.69	ng	0.00
14) 2,4,6-Tribromophenol	15.911	330	13979	5.79	ng	0.00
15) 2-Fluorobiphenyl	13.031	172	136457	4.79	ng	0.00
27) Fluoranthene-d10	19.179	212	1507632	4.80	ng	0.00
31) Terphenyl-d14	19.777	244	259843	4.44	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.302	88	22923	4.08	ng	92
3) n-Nitrosodimethylamine	3.613	42	27308	4.25	ng	# 96
6) bis(2-Chloroethyl)ether	7.206	93	47034	4.21	ng	98
9) Naphthalene	10.614	128	627607	4.69	ng	100
10) Hexachlorobutadiene	10.913	225	27838	4.87	ng	98
12) 2-Methylnaphthalene	12.224	142	112035	4.75	ng	99
16) Acenaphthylene	14.126	152	171980	5.14	ng	100
17) Acenaphthene	14.471	154	110739	4.76	ng	96
18) Fluorene	15.473	166	147858	4.76	ng	99
20) 4,6-Dinitro-2-methylph...	15.548	198	18043	10.62	ng	# 51
21) 4-Bromophenyl-phenylether	16.365	248	48792	5.26	ng	# 93
22) Hexachlorobenzene	16.471	284	45952	5.15	ng	99
23) Atrazine	16.622	200	42707	5.20	ng	97
24) Pentachlorophenol	16.819	266	9666	13.25	ng	# 86
25) Phenanthrene	17.197	178	249162	4.76	ng	99
26) Anthracene	17.287	178	237894	5.09	ng	99
28) Fluoranthene	19.208	202	330024	4.84	ng	98
30) Pyrene	19.570	202	336240	4.33	ng	99
32) Benzo(a)anthracene	21.306	228	366607	4.85	ng	99
33) Chrysene	21.354	228	370330	4.66	ng	99
34) Bis(2-ethylhexyl)phtha...	21.233	149	565073	5.61	ng	100
35) Indeno(1,2,3-cd)pyrene	26.353	276	375160	4.79	ng	95
37) Benzo(b)fluoranthene	23.017	252	349422	4.68	ng	99
38) Benzo(k)fluoranthene	23.067	252	378448	4.88	ng	98
39) Benzo(a)pyrene	23.659	252	323918	4.82	ng	99
40) Dibenzo(a,h)anthracene	26.378	278	328001	4.91	ng	99
41) Benzo(g,h,i)perylene	27.144	276	325139	4.77	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE041921\
Data File : BE101308.D
Acq On : 19 Apr 2021 18:09
Operator : CG/JU
Sample : SSTDICC5.0
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Quant Time: Apr 19 18:53:04 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE041921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 19 16:27:30 2021
Response via : Initial Calibration









1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.782 min Scan# 427
 Delta R.T. -0.010 min
 Lab File: BE101308.D
 Acq: 19 Apr 2021 18:09




Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC5.0


Tgt Ion:	152	Resp:	3271
Ion	Ratio	Lower	Upper
152	100		
150	156.4	124.8	187.2
115	67.3	48.7	73.1



1,4-Dioxane
 Concen: 4.08 ng
 RT: 3.302 min Scan# 38
 Delta R.T. -0.011 min
 Lab File: BE101308.D
 Acq: 19 Apr 2021 18:09

Tgt Ion:	88	Resp:	22923
Ion	Ratio	Lower	Upper
88	100		
58	89.2	65.4	98.0
43	33.6	24.4	36.6









n-Nitrosodimethylamine
Concen: 4.25 ng
RT: 3.613 min Scan# 65
Delta R.T. -0.011 min
Lab File: BE101308.D
Acq: 19 Apr 2021 18:09




Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0


Tgt Ion: 42 Resp: 27308
Ion Ratio Lower Upper
42 100
74 117.0 90.9 136.3
44 7.4 9.2 13.8#



2-Fluorophenol
Concen: 4.19 ng
RT: 5.375 min Scan# 218
Delta R.T. -0.011 min
Lab File: BE101308.D
Acq: 19 Apr 2021 18:09

Tgt Ion: 112 Resp: 40666
Ion Ratio Lower Upper
112 100
64 76.6 62.8 94.2
63 148.8 120.8 181.2









Phenol-d6
Concen: 4.25 ng
RT: 6.964 min Scan# 356
Delta R.T. 0.001 min
Lab File: BE101308.D
Acq: 19 Apr 2021 18:09




Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tgt Ion:	99	Resp:	62643
Ion	Ratio	Lower	Upper
99	100		
42	23.2	18.8	28.2
71	33.7	26.6	40.0




bis(2-Chloroethyl)ether
Concen: 4.21 ng
RT: 7.206 min Scan# 377
Delta R.T. -0.011 min
Lab File: BE101308.D
Acq: 19 Apr 2021 18:09

Tgt Ion:	93	Resp:	47034
Ion	Ratio	Lower	Upper
93	100		
63	365.4	296.7	445.1
95	32.7	25.9	38.9





Abundance Scan 647 (10.576 min): BE101304.D\data.ms (-647) (-)




Naphthalene-d8
Concen: 0.40 ng
RT: 10.575 min Scan# 647
Delta R.T. -0.001 min
Lab File: BE101308.D
Acq: 19 Apr 2021 18:09

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0


Tgt Ion:	136	Resp:	12376
Ion	Ratio	Lower	Upper
136	100		
137	11.4	9.0	13.6
54	25.8	21.3	31.9
68	5.5	4.2	6.4



Abundance Scan 647 (10.575 min): BE101308.D\data.ms (-632) (-)





Abundance Scan 520 (8.926 min): BE101304.D\data.ms (-517) (-)




Nitrobenzene-d5
Concen: 4.93 ng
RT: 8.925 min Scan# 520
Delta R.T. -0.001 min
Lab File: BE101308.D
Acq: 19 Apr 2021 18:09


Tgt Ion:	82	Resp:	56248
Ion	Ratio	Lower	Upper
82	100		
128	108.5	91.2	136.8
54	247.1	204.0	306.0



Abundance Scan 520 (8.925 min): BE101308.D\data.ms (-507) (-)






Abundance Scan 650 (10.615 min): BE101304.D\data.ms (-642) (-)




Naphthalene
Concen: 4.69 ng
RT: 10.614 min Scan# 650
Delta R.T. -0.001 min
Lab File: BE101308.D
Acq: 19 Apr 2021 18:09

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tgt Ion:	128	Resp:	627607
Ion	Ratio	Lower	Upper
128	100		
129	2.7	2.3	3.5
127	3.4	2.6	4.0







Abundance Scan 673 (10.914 min): BE101304.D\data.ms (-667) (-)



Hexachlorobutadiene
Concen: 4.87 ng
RT: 10.913 min Scan# 673
Delta R.T. -0.001 min
Lab File: BE101308.D
Acq: 19 Apr 2021 18:09

Tgt Ion:	225	Resp:	27838
Ion	Ratio	Lower	Upper
225	100		
223	62.5	49.4	74.0
227	63.8	49.7	74.5









2-Methylnaphthalene-d10
Concen: 4.69 ng
RT: 12.152 min Scan# 760
Delta R.T. -0.001 min
Lab File: BE101308.D
Acq: 19 Apr 2021 18:09




Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0


Tgt Ion:152 Resp: 133411
Ion Ratio Lower Upper
152 100
151 20.5 16.2 24.4



2-Methylnaphthalene
Concen: 4.75 ng
RT: 12.224 min Scan# 765
Delta R.T. -0.001 min
Lab File: BE101308.D
Acq: 19 Apr 2021 18:09

Tgt Ion:142 Resp: 112035
Ion Ratio Lower Upper
142 100
141 88.1 71.4 107.0
115 33.3 27.0 40.6









Acenaphthene-d10
 Concen: 0.40 ng
 RT: 14.414 min Scan# 917
 Delta R.T. -0.001 min
 Lab File: BE101308.D
 Acq: 19 Apr 2021 18:09




Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC5.0


Tgt Ion:	164	Resp:	8179
Ion	Ratio	Lower	Upper
164	100		
162	98.4	78.4	117.6
160	41.7	33.3	49.9



2,4,6-Tribromophenol
 Concen: 5.79 ng
 RT: 15.911 min Scan# 1013
 Delta R.T. 0.001 min
 Lab File: BE101308.D
 Acq: 19 Apr 2021 18:09

Tgt Ion:	330	Resp:	13979
Ion	Ratio	Lower	Upper
330	100		
332	98.6	79.7	119.5
141	39.5	33.2	49.8









2-Fluorobiphenyl
Concen: 4.79 ng
RT: 13.031 min Scan# 821
Delta R.T. -0.001 min
Lab File: BE101308.D
Acq: 19 Apr 2021 18:09




Instrument :
BNA_E
ClientSampled :
SSTDICC5.0


Tgt Ion:	172	Resp:	136457
Ion	Ratio	Lower	Upper
172	100		
171	35.4	29.0	43.6
170	24.0	19.1	28.7



Acenaphthylene
Concen: 5.14 ng
RT: 14.126 min Scan# 897
Delta R.T. -0.001 min
Lab File: BE101308.D
Acq: 19 Apr 2021 18:09

Tgt Ion:	152	Resp:	171980
Ion	Ratio	Lower	Upper
152	100		
151	19.3	15.6	23.4
153	13.3	10.7	16.1









Acenaphthene
Concen: 4.76 ng
RT: 14.471 min Scan# 921
Delta R.T. -0.001 min
Lab File: BE101308.D
Acq: 19 Apr 2021 18:09




Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0


Tgt Ion:	154	Resp:	110739
Ion	Ratio	Lower	Upper
154	100		
153	108.8	90.1	135.1
152	52.7	44.2	66.2



Fluorene
Concen: 4.76 ng
RT: 15.473 min Scan# 984
Delta R.T. 0.001 min
Lab File: BE101308.D
Acq: 19 Apr 2021 18:09

Tgt Ion:	166	Resp:	147858
Ion	Ratio	Lower	Upper
166	100		
165	97.2	78.9	118.3
167	13.5	11.0	16.4









Phenanthrene-d10
 Concen: 0.40 ng
 RT: 17.166 min Scan# 1096
 Delta R.T. 0.001 min
 Lab File: BE101308.D
 Acq: 19 Apr 2021 18:09




Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC5.0

Tgt Ion:	188	Resp:	17949
Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	7.1	5.8	8.8



4,6-Dinitro-2-methylphenol
 Concen: 10.62 ng
 RT: 15.548 min Scan# 989
 Delta R.T. 0.001 min
 Lab File: BE101308.D
 Acq: 19 Apr 2021 18:09


Tgt Ion:	198	Resp:	18043
Ion	Ratio	Lower	Upper
198	100		
51	23.0	56.0	84.0#
105	21.1	36.7	55.1#






Abundance Scan 1043 (16.364 min): BE101304.D\data.ms (-1041) (-)

4-Bromophenyl-phenylether
Concen: 5.26 ng
RT: 16.365 min Scan# 1043
Delta R.T. 0.001 min
Lab File: BE101308.D
Acq: 19 Apr 2021 18:09

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0




Tgt Ion:	248	Resp:	48792
Ion Ratio	Lower	Upper	
248	100		
250	100.3	79.2	118.8
141	25.9	31.1	46.7#







Abundance Scan 1050 (16.470 min): BE101304.D\data.ms (-1042) (-)

Hexachlorobenzene
Concen: 5.15 ng
RT: 16.471 min Scan# 1050
Delta R.T. 0.001 min
Lab File: BE101308.D
Acq: 19 Apr 2021 18:09



Tgt Ion:	284	Resp:	45952
Ion Ratio	Lower	Upper	
284	100		
142	32.6	27.2	40.8
249	33.1	26.6	40.0









Scan 1060 (16.621 min): BE101304.D\data.ms (-1023) (-)
 Atrazine
 Concen: 5.20 ng
 RT: 16.622 min Scan# 1060
 Delta R.T. 0.001 min
 Lab File: BE101308.D
 Acq: 19 Apr 2021 18:09




Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC5.0


Tgt Ion:	200	Resp:	42707
Ion Ratio	Lower	Upper	
200	100		
173	34.7	27.8	41.8
215	46.5	39.8	59.6



Scan 1073 (16.818 min): BE101304.D\data.ms (-1024) (-)
 Pentachlorophenol
 Concen: 13.25 ng
 RT: 16.819 min Scan# 1073
 Delta R.T. 0.001 min
 Lab File: BE101308.D
 Acq: 19 Apr 2021 18:09

Tgt Ion:	266	Resp:	9666
Ion Ratio	Lower	Upper	
266	100		
264	64.2	67.0	100.6#
268	64.7	54.1	81.1









Scan 1098 (17.196 min): BE101304.D\data.ms (-1025) (-)
 Phenanthrene
 Concen: 4.76 ng
 RT: 17.197 min Scan# 1098
 Delta R.T. 0.001 min
 Lab File: BE101308.D
 Acq: 19 Apr 2021 18:09




Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC5.0

Tgt Ion:	178	Resp:	249162
Ion	Ratio	Lower	Upper
178	100		
176	19.1	15.0	22.4
179	15.5	12.6	19.0




Scan 1105 (17.301 min): BE101304.D\data.ms (-1025) (-)
 Anthracene
 Concen: 5.09 ng
 RT: 17.287 min Scan# 1104
 Delta R.T. -0.014 min
 Lab File: BE101308.D
 Acq: 19 Apr 2021 18:09

Tgt Ion:	178	Resp:	237894
Ion	Ratio	Lower	Upper
178	100		
176	18.5	14.5	21.7
179	15.6	12.1	18.1






Abundance Scan 1351 (19.175 min): BE101304.D\data.ms (-1323) (-)




Fluoranthene-d10
Concen: 4.80 ng
RT: 19.179 min Scan# 1352
Delta R.T. 0.004 min
Lab File: BE101308.D
Acq: 19 Apr 2021 18:09

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tgt Ion:212 Resp: 1507632
Ion Ratio Lower Upper
212 100
106 3.5 2.6 4.0
104 1.9 1.4 2.2







Abundance Scan 1357 (19.209 min): BE101304.D\data.ms (-1323) (-)



Fluoranthene
Concen: 4.84 ng
RT: 19.208 min Scan# 1357
Delta R.T. -0.001 min
Lab File: BE101308.D
Acq: 19 Apr 2021 18:09

Tgt Ion:202 Resp: 330024
Ion Ratio Lower Upper
202 100
101 12.0 9.2 13.8
203 17.9 13.7 20.5









Chrysene-d12
 Concen: 0.40 ng
 RT: 21.318 min Scan# 1601
 Delta R.T. 0.000 min
 Lab File: BE101308.D
 Acq: 19 Apr 2021 18:09




Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC5.0

Tgt Ion:	240	Resp:	24131
Ion	Ratio	Lower	Upper
240	100		
120	6.4	5.3	7.9
236	25.7	21.0	31.4




Pyrene
 Concen: 4.33 ng
 RT: 19.570 min Scan# 1420
 Delta R.T. 0.004 min
 Lab File: BE101308.D
 Acq: 19 Apr 2021 18:09

Tgt Ion:	202	Resp:	336240
Ion	Ratio	Lower	Upper
202	100		
200	21.1	16.5	24.7
203	17.9	13.8	20.8





Abundance Scan 1456 (19.778 min): BE101304.D\data.ms (-1430) (-)




Terphenyl-d14
Concen: 4.44 ng
RT: 19.777 min Scan# 1456
Delta R.T. -0.001 min
Lab File: BE101308.D
Acq: 19 Apr 2021 18:09

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0


Tgt Ion:	244	Resp:	259843
Ion	Ratio	Lower	Upper
244	100		
212	27.4	21.7	32.5
122	9.3	6.7	10.1



Abundance Scan 1456 (19.777 min): BE101308.D\data.ms (-1430) (-)





Abundance Scan 1599 (21.306 min): BE101304.D\data.ms (-1582) (-)





Benzo(a)anthracene
Concen: 4.85 ng
RT: 21.306 min Scan# 1600
Delta R.T. 0.000 min
Lab File: BE101308.D
Acq: 19 Apr 2021 18:09

Tgt Ion:	228	Resp:	366607
Ion	Ratio	Lower	Upper
228	100		
226	26.0	20.4	30.6
229	20.7	16.1	24.1



Abundance Scan 1600 (21.306 min): BE101308.D\data.ms (-1583) (-)









Scan 1603 (21.354 min): BE101304.D\data.ms (-1596) (-)
 Chrysene
 Concen: 4.66 ng
 RT: 21.354 min Scan# 1604
 Delta R.T. 0.000 min
 Lab File: BE101308.D
 Acq: 19 Apr 2021 18:09




Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC5.0

Tgt Ion:	228	Resp:	370330
Ion	Ratio	Lower	Upper
228	100		
226	29.6	23.7	35.5
229	20.5	15.8	23.8



Scan 1593 (21.233 min): BE101304.D\data.ms (-1596) (-)
 Bis(2-ethylhexyl)phthalate
 Concen: 5.61 ng
 RT: 21.233 min Scan# 1594
 Delta R.T. 0.000 min
 Lab File: BE101308.D
 Acq: 19 Apr 2021 18:09


Tgt Ion:	149	Resp:	565073
Ion	Ratio	Lower	Upper
149	100		
167	7.2	5.7	8.5
279	1.2	1.0	1.6






Abundance Scan 2866 (26.352 min): BE101304.D\data.ms (-2839) (-)

Indeno(1,2,3-cd)pyrene
Concen: 4.79 ng
RT: 26.353 min Scan# 2867
Delta R.T. 0.001 min
Lab File: BE101308.D
Acq: 19 Apr 2021 18:09

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0




Tgt Ion	Ratio	Lower	Upper
276	100		
138	22.6	16.6	24.8
277	25.2	17.6	26.4






Abundance Scan 2147 (23.773 min): BE101304.D\data.ms (-2126) (-)


Perylene-d12
Concen: 0.40 ng
RT: 23.769 min Scan# 2147
Delta R.T. -0.003 min
Lab File: BE101308.D
Acq: 19 Apr 2021 18:09



Tgt Ion	Ratio	Lower	Upper
264	100		
260	25.1	20.1	30.1
265	23.7	19.4	29.0





Abundance Scan 1935 (23.017 min): BE101304.D\data.ms (-1880) (-)




Benzo(b)fluoranthene
Concen: 4.68 ng
RT: 23.017 min Scan# 1936
Delta R.T. 0.000 min
Lab File: BE101308.D
Acq: 19 Apr 2021 18:09


Tgt Ion:	252	Resp:	349422
Ion	Ratio	Lower	Upper
252	100		
253	22.5	17.7	26.5
125	9.1	7.5	11.3



Abundance Scan 1936 (23.017 min): BE101308.D\data.ms (-1880) (-)





Abundance Scan 1949 (23.067 min): BE101304.D\data.ms (-1880) (-)




Benzo(k)fluoranthene
Concen: 4.88 ng
RT: 23.067 min Scan# 1950
Delta R.T. 0.000 min
Lab File: BE101308.D
Acq: 19 Apr 2021 18:09

Tgt Ion:	252	Resp:	378448
Ion	Ratio	Lower	Upper
252	100		
253	22.4	18.1	27.1
125	9.0	8.8	13.2




Abundance Scan 1950 (23.067 min): BE101308.D\data.ms (-1866) (-)






Abundance Scan 2115 (23.659 min): BE101304.D\data.ms (-2060) (-)

Benzo(a)pyrene
Concen: 4.82 ng
RT: 23.659 min Scan# 2116
Delta R.T. 0.000 min
Lab File: BE101308.D
Acq: 19 Apr 2021 18:09

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0




Tgt Ion:	252	Resp:	323918
Ion Ratio	Lower	Upper	
252	100		
253	22.2	17.9	26.9
125	9.9	8.5	12.7






Abundance Scan 2873 (26.377 min): BE101304.D\data.ms (-2848) (-)


Dibenzo(a,h)anthracene
Concen: 4.91 ng
RT: 26.378 min Scan# 2874
Delta R.T. 0.001 min
Lab File: BE101308.D
Acq: 19 Apr 2021 18:09



Tgt Ion:	278	Resp:	328001
Ion Ratio	Lower	Upper	
278	100		
139	14.8	12.9	19.3
279	24.1	19.3	28.9





Abundance Scan 3089 (27.147 min): BE101304.D\data.ms (-3034) (-)



Benzo(g,h,i)perylene
Concen: 4.77 ng
RT: 27.144 min Scan# 3089
Delta R.T. -0.003 min
Lab File: BE101308.D
Acq: 19 Apr 2021 18:09

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tgt Ion:276 Resp: 325139
Ion Ratio Lower Upper
276 100
277 24.0 19.3 28.9
138 19.1 15.8 23.8



Abundance Scan 3089 (27.144 min): BE101308.D\data.ms (-3034) (-)

