

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE041921\
 Data File : BE101302.D
 Acq On : 19 Apr 2021 14:25
 Operator : CG/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.1

Manual Integrations
 APPROVED

mohammad
 4/21/2021 8:37:35 AM

Quant Time: Apr 19 16:33:31 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE041921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 19 16:27:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.792	152	2559	0.40	ng	0.00
7) Naphthalene-d8	10.576	136	10932	0.40	ng	0.00
13) Acenaphthene-d10	14.414	164	7036	0.40	ng	0.00
19) Phenanthrene-d10	17.166	188	17851	0.40	ng	0.00
29) Chrysene-d12	21.319	240	19733	0.40	ng	0.00
36) Perylene-d12	23.770	264	19595	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.374	112	771	0.10	ng	-0.01
5) Phenol-d6	6.963	99	1135	0.10	ng	0.00
8) Nitrobenzene-d5	8.939	82	886	0.09	ng	0.01
11) 2-Methylnaphthalene-d10	12.152	152	2478	0.10	ng	0.00
14) 2,4,6-Tribromophenol	15.912	330	162	0.08	ng	0.00
15) 2-Fluorobiphenyl	13.031	172	2643	0.11	ng	0.00
27) Fluoranthene-d10	19.179	212	26188	0.08	ng	0.00
31) Terphenyl-d14	19.777	244	4381	0.09	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.301	88	554	0.13	ng	97
6) bis(2-Chloroethyl)ether	7.216	93	899	0.10	ng	93
9) Naphthalene	10.615	128	12033	0.10	ng	99
10) Hexachlorobutadiene	10.913	225	536	0.11	ng	99
12) 2-Methylnaphthalene	12.224	142	1928	0.09	ng	97
16) Acenaphthylene	14.126	152	2367	0.08	ng	100
17) Acenaphthene	14.472	154	1955	0.10	ng	98
18) Fluorene	15.473	166	2664	0.10	ng	96
20) 4,6-Dinitro-2-methylph...	15.549	198	65	0.04	ng	# 16
21) 4-Bromophenyl-phenylether	16.365	248	891	0.10	ng	96
22) Hexachlorobenzene	16.471	284	905	0.10	ng	99
23) Atrazine	16.637	200	603	0.07	ng	# 89
25) Phenanthrene	17.197	178	5209	0.10	ng	100
26) Anthracene	17.287	178	3978	0.09	ng	99
28) Fluoranthene	19.208	202	5528	0.08	ng	100
30) Pyrene	19.570	202	5752	0.09	ng	99
32) Benzo(a)anthracene	21.307	228	5892	0.10	ng	98
33) Chrysene	21.355	228	6453	0.10	ng	99
34) Bis(2-ethylhexyl)phtha...	21.234	149	7186	0.09	ng	100
35) Indeno(1,2,3-cd)pyrene	26.355	276	6155	0.10	ng	94
37) Benzo(b)fluoranthene	23.018	252	5935	0.09	ng	# 94
38) Benzo(k)fluoranthene	23.068	252	6807m	0.10	ng	
39) Benzo(a)pyrene	23.660	252	5078	0.09	ng	# 94
40) Dibenzo(a,h)anthracene	26.380	278	5334	0.09	ng	95
41) Benzo(g,h,i)perylene	27.147	276	5500	0.09	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE041921\
Data File : BE101302.D
Acq On : 19 Apr 2021 14:25
Operator : CG/JU
Sample : SSTDICC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
SSTDICC0.1

Manual Integrations
APPROVED

mohammad
4/21/2021 8:37:35 AM


Quant Time: Apr 19 16:33:31 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE041921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 19 16:27:30 2021
Response via : Initial Calibration





#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.792 min Scan# 428
 Delta R.T. 0.000 min
 Lab File: BE101302.D
 Acq: 19 Apr 2021 14:25



Instrument :
 BNA_E
 ClientSampled :
 SSTDICC0.1




Tgt Ion: 152 Resp: 2559
 Ion Ratio Lower Upper
 152 100
 150 150.9 124.8 187.2
 115 56.1 48.7 73.1

Manual Integrations
 APPROVED



mohammad
 4/21/2021 8:37:35 AM




#2
 1,4-Dioxane
 Concen: 0.13 ng
 RT: 3.301 min Scan# 38
 Delta R.T. -0.012 min
 Lab File: BE101302.D
 Acq: 19 Apr 2021 14:25



Tgt Ion: 88 Resp: 554
 Ion Ratio Lower Upper
 88 100
 58 78.9 65.4 98.0
 43 29.4 24.4 36.6









#4
2-Fluorophenol
Concen: 0.10 ng
RT: 5.374 min Scan# 218
Delta R.T. -0.012 min
Lab File: BE101302.D
Acq: 19 Apr 2021 14:25

Instrument :
BNA_E
ClientSampled :
SSTDIC0.1

Tgt Ion: 112 Resp: 771
Ion Ratio Lower Upper
112 100
64 78.2 62.8 94.2
63 166.0 120.8 181.2




Manual Integrations
APPROVED


mohammad
4/21/2021 8:37:35 AM



#5
Phenol-d6
Concen: 0.10 ng
RT: 6.963 min Scan# 356
Delta R.T. -0.000 min
Lab File: BE101302.D
Acq: 19 Apr 2021 14:25

Tgt Ion: 99 Resp: 1135
Ion Ratio Lower Upper
99 100
42 26.5 18.8 28.2
71 33.1 26.6 40.0









#6
bis(2-Chloroethyl)ether
Concen: 0.10 ng
RT: 7.216 min Scan# 378
Delta R.T. -0.000 min
Lab File: BE101302.D
Acq: 19 Apr 2021 14:25

Tgt Ion:	93	Resp:	899
Ion	Ratio	Lower	Upper
93	100		
63	356.2	296.7	445.1
95	36.0	25.9	38.9

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
4/21/2021 8:37:35 AM



#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.576 min Scan# 647
Delta R.T. -0.001 min
Lab File: BE101302.D
Acq: 19 Apr 2021 14:25


Tgt Ion:	136	Resp:	10932
Ion	Ratio	Lower	Upper
136	100		
137	11.2	9.0	13.6
54	28.5	21.3	31.9
68	6.2	4.2	6.4





#8
Nitrobenzene-d5
Concen: 0.09 ng
RT: 8.939 min Scan# 521
Delta R.T. 0.012 min
Lab File: BE101302.D
Acq: 19 Apr 2021 14:25




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1



Tgt Ion: 82 Resp: 886
Ion Ratio Lower Upper
82 100
128 178.4 91.2 136.8#
54 208.7 204.0 306.0




Manual Integrations
APPROVED


mohammad
4/21/2021 8:37:35 AM



#9
Naphthalene
Concen: 0.10 ng
RT: 10.615 min Scan# 650
Delta R.T. -0.001 min
Lab File: BE101302.D
Acq: 19 Apr 2021 14:25

Tgt Ion: 128 Resp: 12033
Ion Ratio Lower Upper
128 100
129 3.2 2.3 3.5
127 3.7 2.6 4.0









#10
Hexachlorobutadiene
Concen: 0.11 ng
RT: 10.913 min Scan# 673
Delta R.T. -0.001 min
Lab File: BE101302.D
Acq: 19 Apr 2021 14:25

Tgt Ion:	225	Resp:	536
Ion	Ratio	Lower	Upper
225	100		
223	62.3	49.4	74.0
227	63.2	49.7	74.5

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
4/21/2021 8:37:35 AM



#11
2-Methylnaphthalene-d10
Concen: 0.10 ng
RT: 12.152 min Scan# 760
Delta R.T. -0.001 min
Lab File: BE101302.D
Acq: 19 Apr 2021 14:25


Tgt Ion:	152	Resp:	2478
Ion	Ratio	Lower	Upper
152	100		
151	20.7	16.2	24.4





#12
2-Methylnaphthalene
Concen: 0.09 ng
RT: 12.224 min Scan# 765
Delta R.T. -0.001 min
Lab File: BE101302.D
Acq: 19 Apr 2021 14:25




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1



Tgt Ion:	142	Resp:	1928
Ion	Ratio	Lower	Upper
142	100		
141	90.9	71.4	107.0
115	37.3	27.0	40.6




Manual Integrations
APPROVED


mohammad
4/21/2021 8:37:35 AM



#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.414 min Scan# 917
Delta R.T. -0.001 min
Lab File: BE101302.D
Acq: 19 Apr 2021 14:25


Tgt Ion:	164	Resp:	7036
Ion	Ratio	Lower	Upper
164	100		
162	99.5	78.4	117.6
160	43.4	33.3	49.9





#14
2,4,6-Tribromophenol
Concen: 0.08 ng
RT: 15.912 min Scan# 1013
Delta R.T. 0.002 min
Lab File: BE101302.D
Acq: 19 Apr 2021 14:25




Instrument :
BNA_E
ClientSampled :
SSTDIC0.1




Tgt Ion: 330 Resp: 162
Ion Ratio Lower Upper
330 100
332 81.5 79.7 119.5
141 40.7 33.2 49.8

Manual Integrations
APPROVED



mohammad
4/21/2021 8:37:35 AM




#15
2-Fluorobiphenyl
Concen: 0.11 ng
RT: 13.031 min Scan# 821
Delta R.T. -0.001 min
Lab File: BE101302.D
Acq: 19 Apr 2021 14:25



Tgt Ion: 172 Resp: 2643
Ion Ratio Lower Upper
172 100
171 36.7 29.0 43.6
170 26.1 19.1 28.7









#16
Acenaphthylene
Concen: 0.08 ng
RT: 14.126 min Scan# 897
Delta R.T. -0.001 min
Lab File: BE101302.D
Acq: 19 Apr 2021 14:25

Tgt Ion:	152	Resp:	2367
Ion Ratio	Lower	Upper	
152	100		
151	19.6	15.6	23.4
153	13.2	10.7	16.1

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
4/21/2021 8:37:35 AM



#17
Acenaphthene
Concen: 0.10 ng
RT: 14.472 min Scan# 921
Delta R.T. -0.001 min
Lab File: BE101302.D
Acq: 19 Apr 2021 14:25


Tgt Ion:	154	Resp:	1955
Ion Ratio	Lower	Upper	
154	100		
153	109.5	90.1	135.1
152	55.6	44.2	66.2





#18
Fluorene
Concen: 0.10 ng
RT: 15.473 min Scan# 984
Delta R.T. 0.001 min
Lab File: BE101302.D
Acq: 19 Apr 2021 14:25




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1




Tgt Ion:166 Resp: 2664
Ion Ratio Lower Upper
166 100
165 94.0 78.9 118.3
167 13.8 11.0 16.4

Manual Integrations
APPROVED



mohammad
4/21/2021 8:37:35 AM




#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.166 min Scan# 1096
Delta R.T. 0.001 min
Lab File: BE101302.D
Acq: 19 Apr 2021 14:25



Tgt Ion:188 Resp: 17851
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 8.2 5.8 8.8









#20
4,6-Dinitro-2-methylphenol
Concen: 0.04 ng
RT: 15.549 min Scan# 989
Delta R.T. 0.001 min
Lab File: BE101302.D
Acq: 19 Apr 2021 14:25

Tgt Ion:198 Resp: 65
Ion Ratio Lower Upper
198 100
51 146.4 56.0 84.0#
105 91.7 36.7 55.1#

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
4/21/2021 8:37:35 AM



#21
4-Bromophenyl-phenylether
Concen: 0.10 ng
RT: 16.365 min Scan# 1043
Delta R.T. 0.001 min
Lab File: BE101302.D
Acq: 19 Apr 2021 14:25


Tgt Ion:248 Resp: 891
Ion Ratio Lower Upper
248 100
250 95.3 79.2 118.8
141 42.0 31.1 46.7





#22
Hexachlorobenzene
Concen: 0.10 ng
RT: 16.471 min Scan# 1050
Delta R.T. 0.001 min
Lab File: BE101302.D
Acq: 19 Apr 2021 14:25




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1




Tgt Ion:284 Resp: 905
Ion Ratio Lower Upper
284 100
142 34.3 27.2 40.8
249 33.7 26.6 40.0

Manual Integrations
APPROVED



mohammad
4/21/2021 8:37:35 AM




#23
Atrazine
Concen: 0.07 ng
RT: 16.637 min Scan# 1061
Delta R.T. 0.016 min
Lab File: BE101302.D
Acq: 19 Apr 2021 14:25




Tgt Ion:200 Resp: 603
Ion Ratio Lower Upper
200 100
173 27.1 27.8 41.8#
215 56.3 39.8 59.6





#25
Phenanthrene
Concen: 0.10 ng
RT: 17.197 min Scan# 1098
Delta R.T. 0.001 min
Lab File: BE101302.D
Acq: 19 Apr 2021 14:25




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1




Tgt Ion:178 Resp: 5209
Ion Ratio Lower Upper
178 100
176 18.9 15.0 22.4
179 15.7 12.6 19.0

Manual Integrations
APPROVED



mohammad
4/21/2021 8:37:35 AM




#26
Anthracene
Concen: 0.09 ng
RT: 17.287 min Scan# 1104
Delta R.T. -0.014 min
Lab File: BE101302.D
Acq: 19 Apr 2021 14:25




Tgt Ion:178 Resp: 3978
Ion Ratio Lower Upper
178 100
176 18.0 14.5 21.7
179 16.2 12.1 18.1





#27
Fluoranthene-d10
Concen: 0.08 ng
RT: 19.179 min Scan# 1352
Delta R.T. 0.005 min
Lab File: BE101302.D
Acq: 19 Apr 2021 14:25



Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1




Tgt Ion:212 Resp: 26188
Ion Ratio Lower Upper
212 100
106 3.4 2.6 4.0
104 1.8 1.4 2.2

Manual Integrations
APPROVED



mohammad
4/21/2021 8:37:35 AM




#28
Fluoranthene
Concen: 0.08 ng
RT: 19.208 min Scan# 1357
Delta R.T. -0.001 min
Lab File: BE101302.D
Acq: 19 Apr 2021 14:25



Tgt Ion:202 Resp: 5528
Ion Ratio Lower Upper
202 100
101 11.4 9.2 13.8
203 17.2 13.7 20.5









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.319 min Scan# 1601
Delta R.T. 0.001 min
Lab File: BE101302.D
Acq: 19 Apr 2021 14:25

Tgt Ion:240 Resp: 19733
Ion Ratio Lower Upper
240 100
120 6.4 5.3 7.9
236 25.9 21.0 31.4

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
4/21/2021 8:37:35 AM




#30
Pyrene
Concen: 0.09 ng
RT: 19.570 min Scan# 1420
Delta R.T. 0.005 min
Lab File: BE101302.D
Acq: 19 Apr 2021 14:25

Tgt Ion:202 Resp: 5752
Ion Ratio Lower Upper
202 100
200 20.6 16.5 24.7
203 18.2 13.8 20.8






#31
Terphenyl-d14
Concen: 0.09 ng
RT: 19.777 min Scan# 1456
Delta R.T. -0.001 min
Lab File: BE101302.D
Acq: 19 Apr 2021 14:25




Tgt Ion:244 Resp: 4381
Ion Ratio Lower Upper
244 100
212 30.7 21.7 32.5
122 11.1 6.7 10.1#




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1

Manual Integrations
APPROVED



mohammad
4/21/2021 8:37:35 AM




#32
Benzo(a)anthracene
Concen: 0.10 ng
RT: 21.307 min Scan# 1600
Delta R.T. 0.001 min
Lab File: BE101302.D
Acq: 19 Apr 2021 14:25



Tgt Ion:228 Resp: 5892
Ion Ratio Lower Upper
228 100
226 27.1 20.4 30.6
229 20.6 16.1 24.1









#33
Chrysene
Concen: 0.10 ng
RT: 21.355 min Scan# 1604
Delta R.T. 0.001 min
Lab File: BE101302.D
Acq: 19 Apr 2021 14:25

Tgt Ion:	228	Resp:	6453
Ion Ratio	Lower	Upper	
228	100		
226	30.0	23.7	35.5
229	20.6	15.8	23.8

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
4/21/2021 8:37:35 AM




#34
Bis(2-ethylhexyl)phthalate
Concen: 0.09 ng
RT: 21.234 min Scan# 1594
Delta R.T. 0.001 min
Lab File: BE101302.D
Acq: 19 Apr 2021 14:25

Tgt Ion:	149	Resp:	7186
Ion Ratio	Lower	Upper	
149	100		
167	7.2	5.7	8.5
279	1.3	1.0	1.6









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.10 ng
RT: 26.355 min Scan# 2868
Delta R.T. 0.003 min
Lab File: BE101302.D
Acq: 19 Apr 2021 14:25





Tgt Ion: 276 Resp: 6155
Ion Ratio Lower Upper
276 100
138 23.4 16.6 24.8
277 25.2 17.6 26.4



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.770 min Scan# 2147
Delta R.T. -0.002 min
Lab File: BE101302.D
Acq: 19 Apr 2021 14:25




Tgt Ion: 264 Resp: 19595
Ion Ratio Lower Upper
264 100
260 24.7 20.1 30.1
265 24.2 19.4 29.0



Instrument :
BNA_E
ClientSampled :
SSTDIC0.1

Manual Integrations
APPROVED

mohammad
4/21/2021 8:37:35 AM







#37
Benzo(b)fluoranthene
Concen: 0.09 ng
RT: 23.018 min Scan# 1936
Delta R.T. 0.001 min
Lab File: BE101302.D
Acq: 19 Apr 2021 14:25

Tgt Ion:252 Resp: 5935
Ion Ratio Lower Upper
252 100
253 24.9 17.7 26.5
125 11.5 7.5 11.3#

Instrument :
BNA_E
ClientSampleId :
SSTDICC0.1




Manual Integrations
APPROVED


mohammad
4/21/2021 8:37:35 AM



#38
Benzo(k)fluoranthene
Concen: 0.10 ng m
RT: 23.068 min Scan# 1950
Delta R.T. 0.001 min
Lab File: BE101302.D
Acq: 19 Apr 2021 14:25

Tgt Ion:252 Resp: 6807
Ion Ratio Lower Upper
252 100
253 25.6 18.1 27.1
125 16.7 8.8 13.2#









#39
Benzo(a)pyrene
Concen: 0.09 ng
RT: 23.660 min Scan# 2116
Delta R.T. 0.001 min
Lab File: BE101302.D
Acq: 19 Apr 2021 14:25

Tgt Ion:	252	Resp:	5078
Ion Ratio	Lower	Upper	
252	100		
253	25.2	17.9	26.9
125	13.2	8.5	12.7#

Instrument :
BNA_E
ClientSampleId :
SSTDICC0.1




Manual Integrations
APPROVED


mohammad
4/21/2021 8:37:35 AM



#40
Dibenzo(a,h)anthracene
Concen: 0.09 ng
RT: 26.380 min Scan# 2875
Delta R.T. 0.003 min
Lab File: BE101302.D
Acq: 19 Apr 2021 14:25

Tgt Ion:	278	Resp:	5334
Ion Ratio	Lower	Upper	
278	100		
139	18.8	12.9	19.3
279	26.4	19.3	28.9





#41
Benzo(g,h,i)perylene
Concen: 0.09 ng
RT: 27.147 min Scan# 3090
Delta R.T. -0.000 min
Lab File: BE101302.D
Acq: 19 Apr 2021 14:25

Tgt Ion:	276	Resp:	5500
Ion Ratio	Lower	Upper	
276	100		
277	25.5	19.3	28.9
138	20.4	15.8	23.8

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1

Manual Integrations
APPROVED

mohammad
4/21/2021 8:37:35 AM

