

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE041921\
 Data File : BE101304.D
 Acq On : 19 Apr 2021 15:40
 Operator : CG/JU
 Sample : SSTDICCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICCC0.4

Quant Time: Apr 19 16:36:15 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE041921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 19 16:27:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.792	152	2618	0.40	ng	0.00
7) Naphthalene-d8	10.576	136	10954	0.40	ng	0.00
13) Acenaphthene-d10	14.415	164	6752	0.40	ng	0.00
19) Phenanthrene-d10	17.165	188	16444	0.40	ng	0.00
29) Chrysene-d12	21.318	240	21390	0.40	ng	0.00
36) Perylene-d12	23.773	264	21755	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.385	112	2927	0.38	ng	0.00
5) Phenol-d6	6.963	99	4262	0.36	ng	0.00
8) Nitrobenzene-d5	8.926	82	3530	0.35	ng	0.00
11) 2-Methylnaphthalene-d10	12.153	152	9775	0.39	ng	0.00
14) 2,4,6-Tribromophenol	15.910	330	521	0.26	ng	0.00
15) 2-Fluorobiphenyl	13.032	172	9971	0.42	ng	0.00
27) Fluoranthene-d10	19.175	212	109744	0.38	ng	0.00
31) Terphenyl-d14	19.778	244	19175	0.37	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.312	88	1958	0.44	ng	96
3) n-Nitrosodimethylamine	3.623	42	1934	0.38	ng	99
6) bis(2-Chloroethyl)ether	7.217	93	3521	0.39	ng	100
9) Naphthalene	10.615	128	47431	0.40	ng	100
10) Hexachlorobutadiene	10.914	225	2195	0.43	ng	100
12) 2-Methylnaphthalene	12.225	142	7929	0.38	ng	100
16) Acenaphthylene	14.127	152	9624	0.35	ng	100
17) Acenaphthene	14.472	154	7361	0.38	ng	100
18) Fluorene	15.472	166	9754	0.38	ng	100
20) 4,6-Dinitro-2-methylph...	15.547	198	342	0.22	ng	100
21) 4-Bromophenyl-phenylether	16.364	248	3228	0.38	ng	100
22) Hexachlorobenzene	16.470	284	3287	0.40	ng	100
23) Atrazine	16.621	200	2164	0.29	ng	100
24) Pentachlorophenol	16.818	266	76	0.11	ng	# 87
25) Phenanthrene	17.196	178	18689	0.39	ng	100
26) Anthracene	17.301	178	14980	0.35	ng	100
28) Fluoranthene	19.209	202	24332	0.39	ng	100
30) Pyrene	19.565	202	24246	0.35	ng	100
32) Benzo(a)anthracene	21.306	228	25171	0.38	ng	100
33) Chrysene	21.354	228	28168	0.40	ng	100
34) Bis(2-ethylhexyl)phtha...	21.233	149	27331	0.31	ng	100
35) Indeno(1,2,3-cd)pyrene	26.352	276	27604	0.40	ng	94
37) Benzo(b)fluoranthene	23.017	252	26003	0.36	ng	100
38) Benzo(k)fluoranthene	23.067	252	29223	0.38	ng	100
39) Benzo(a)pyrene	23.659	252	22652	0.34	ng	100
40) Dibenzo(a,h)anthracene	26.377	278	24233	0.37	ng	100
41) Benzo(g,h,i)perylene	27.147	276	24489	0.37	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE041921\
Data File : BE101304.D
Acq On : 19 Apr 2021 15:40
Operator : CG/JU
Sample : SSTDICCC0.4
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4

Quant Time: Apr 19 16:36:15 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE041921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 19 16:27:30 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.792 min Scan# 428
 Delta R.T. 0.000 min
 Lab File: BE101304.D
 Acq: 19 Apr 2021 15:40

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.4

Tgt Ion: 152 Resp: 2618




Ion	Ratio	Lower	Upper
152	100		
150	156.0	124.8	187.2
115	60.9	48.7	73.1






#2
 1,4-Dioxane
 Concen: 0.44 ng
 RT: 3.312 min Scan# 39
 Delta R.T. 0.000 min
 Lab File: BE101304.D
 Acq: 19 Apr 2021 15:40

Tgt Ion: 88 Resp: 1958

Ion	Ratio	Lower	Upper
88	100		
58	85.6	65.4	98.0
43	32.0	24.4	36.6







Abundance Scan 66 (3.623 min): BE101304.D\data.ms (-62)



#3

n-Nitrosodimethylamine
Concen: 0.38 ng
RT: 3.623 min Scan# 66
Delta R.T. 0.000 min
Lab File: BE101304.D
Acq: 19 Apr 2021 15:40


Tgt Ion:	42	Resp:	1934
Ion	Ratio	Lower	Upper
42	100		
74	112.9	90.9	136.3
44	13.0	9.2	13.8




#4

2-Fluorophenol
Concen: 0.38 ng
RT: 5.385 min Scan# 219
Delta R.T. 0.000 min
Lab File: BE101304.D
Acq: 19 Apr 2021 15:40

Tgt Ion:	112	Resp:	2927
Ion	Ratio	Lower	Upper
112	100		
64	78.5	62.8	94.2
63	151.0	120.8	181.2







Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4



#5
Phenol-d6
Concen: 0.36 ng
RT: 6.963 min Scan# 356
Delta R.T. 0.000 min
Lab File: BE101304.D
Acq: 19 Apr 2021 15:40




Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	99	Resp:	4262
Ion	Ratio	Lower	Upper
99	100		
42	23.5	18.8	28.2
71	34.1	26.6	40.0



#6
bis(2-Chloroethyl)ether
Concen: 0.39 ng
RT: 7.217 min Scan# 378
Delta R.T. 0.000 min
Lab File: BE101304.D
Acq: 19 Apr 2021 15:40

Tgt Ion:	93	Resp:	3521
Ion	Ratio	Lower	Upper
93	100		
63	370.9	296.7	445.1
95	32.4	25.9	38.9









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.576 min Scan# 647
Delta R.T. 0.000 min
Lab File: BE101304.D
Acq: 19 Apr 2021 15:40




Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	136	Resp:	10954
Ion	Ratio	Lower	Upper
136	100		
137	11.3	9.0	13.6
54	26.6	21.3	31.9
68	5.3	4.2	6.4



#8
Nitrobenzene-d5
Concen: 0.35 ng
RT: 8.926 min Scan# 520
Delta R.T. 0.000 min
Lab File: BE101304.D
Acq: 19 Apr 2021 15:40

Tgt Ion:	82	Resp:	3530
Ion	Ratio	Lower	Upper
82	100		
128	114.0	91.2	136.8
54	255.0	204.0	306.0









#9
Naphthalene
Concen: 0.40 ng
RT: 10.615 min Scan# 650
Delta R.T. 0.000 min
Lab File: BE101304.D
Acq: 19 Apr 2021 15:40




Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	128	Resp:	47431
Ion	Ratio	Lower	Upper
128	100		
129	2.9	2.3	3.5
127	3.3	2.6	4.0



#10
Hexachlorobutadiene
Concen: 0.43 ng
RT: 10.914 min Scan# 673
Delta R.T. 0.000 min
Lab File: BE101304.D
Acq: 19 Apr 2021 15:40

Tgt Ion:	225	Resp:	2195
Ion	Ratio	Lower	Upper
225	100		
223	61.7	49.4	74.0
227	62.1	49.7	74.5









#11
2-Methylnaphthalene-d10
Concen: 0.39 ng
RT: 12.153 min Scan# 760
Delta R.T. 0.000 min
Lab File: BE101304.D
Acq: 19 Apr 2021 15:40




Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4


Tgt Ion:152 Resp: 9775
Ion Ratio Lower Upper
152 100
151 20.3 16.2 24.4



#12
2-Methylnaphthalene
Concen: 0.38 ng
RT: 12.225 min Scan# 765
Delta R.T. 0.000 min
Lab File: BE101304.D
Acq: 19 Apr 2021 15:40

Tgt Ion:142 Resp: 7929
Ion Ratio Lower Upper
142 100
141 89.2 71.4 107.0
115 33.8 27.0 40.6









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.415 min Scan# 917
Delta R.T. 0.000 min
Lab File: BE101304.D
Acq: 19 Apr 2021 15:40




Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4


Tgt Ion:164 Resp: 6752
Ion Ratio Lower Upper
164 100
162 98.0 78.4 117.6
160 41.6 33.3 49.9



#14
2,4,6-Tribromophenol
Concen: 0.26 ng
RT: 15.910 min Scan# 1013
Delta R.T. 0.000 min
Lab File: BE101304.D
Acq: 19 Apr 2021 15:40

Tgt Ion:330 Resp: 521
Ion Ratio Lower Upper
330 100
332 101.2 79.7 119.5
141 41.1 33.2 49.8









#15
2-Fluorobiphenyl
Concen: 0.42 ng
RT: 13.032 min Scan# 821
Delta R.T. 0.000 min
Lab File: BE101304.D
Acq: 19 Apr 2021 15:40

Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4

Tgt Ion:172 Resp: 9971




Ion	Ratio	Lower	Upper
172	100		
171	36.3	29.0	43.6
170	23.9	19.1	28.7




#16
Acenaphthylene
Concen: 0.35 ng
RT: 14.127 min Scan# 897
Delta R.T. 0.000 min
Lab File: BE101304.D
Acq: 19 Apr 2021 15:40

Tgt Ion:152 Resp: 9624

Ion	Ratio	Lower	Upper
152	100		
151	19.6	15.6	23.4
153	13.4	10.7	16.1









#17
Acenaphthene
Concen: 0.38 ng
RT: 14.472 min Scan# 921
Delta R.T. 0.000 min
Lab File: BE101304.D
Acq: 19 Apr 2021 15:40

Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4

Tgt Ion:154 Resp: 7361




Ion	Ratio	Lower	Upper
154	100		
153	112.6	90.1	135.1
152	55.2	44.2	66.2




#18
Fluorene
Concen: 0.38 ng
RT: 15.472 min Scan# 984
Delta R.T. 0.000 min
Lab File: BE101304.D
Acq: 19 Apr 2021 15:40

Tgt Ion:166 Resp: 9754

Ion	Ratio	Lower	Upper
166	100		
165	98.6	78.9	118.3
167	13.7	11.0	16.4









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.165 min Scan# 1096
Delta R.T. 0.000 min
Lab File: BE101304.D
Acq: 19 Apr 2021 15:40




Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4


Tgt Ion:188 Resp: 16444
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 7.3 5.8 8.8



#20
4,6-Dinitro-2-methylphenol
Concen: 0.22 ng
RT: 15.547 min Scan# 989
Delta R.T. 0.000 min
Lab File: BE101304.D
Acq: 19 Apr 2021 15:40

Tgt Ion:198 Resp: 342
Ion Ratio Lower Upper
198 100
51 70.0 56.0 84.0
105 45.9 36.7 55.1









#21
4-Bromophenyl-phenylether
Concen: 0.38 ng
RT: 16.364 min Scan# 1043
Delta R.T. 0.000 min
Lab File: BE101304.D
Acq: 19 Apr 2021 15:40

Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4

Tgt Ion:248 Resp: 3228




Ion	Ratio	Lower	Upper
248	100		
250	99.0	79.2	118.8
141	38.9	31.1	46.7




#22
Hexachlorobenzene
Concen: 0.40 ng
RT: 16.470 min Scan# 1050
Delta R.T. 0.000 min
Lab File: BE101304.D
Acq: 19 Apr 2021 15:40

Tgt Ion:284 Resp: 3287

Ion	Ratio	Lower	Upper
284	100		
142	34.1	27.2	40.8
249	33.6	26.6	40.0









#23
Atrazine
Concen: 0.29 ng
RT: 16.621 min Scan# 1060
Delta R.T. 0.000 min
Lab File: BE101304.D
Acq: 19 Apr 2021 15:40




Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	200	Resp:	2164
Ion	Ratio	Lower	Upper
200	100		
173	34.8	27.8	41.8
215	49.7	39.8	59.6



#24
Pentachlorophenol
Concen: 0.11 ng
RT: 16.818 min Scan# 1073
Delta R.T. 0.000 min
Lab File: BE101304.D
Acq: 19 Apr 2021 15:40

Tgt Ion:	266	Resp:	76
Ion	Ratio	Lower	Upper
266	100		
264	92.1	67.0	100.6
268	81.6	54.1	81.1#









#25
Phenanthrene
Concen: 0.39 ng
RT: 17.196 min Scan# 1098
Delta R.T. 0.000 min
Lab File: BE101304.D
Acq: 19 Apr 2021 15:40




Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	178	Resp:	18689
Ion	Ratio	Lower	Upper
178	100		
176	18.7	15.0	22.4
179	15.8	12.6	19.0



#26
Anthracene
Concen: 0.35 ng
RT: 17.301 min Scan# 1105
Delta R.T. 0.000 min
Lab File: BE101304.D
Acq: 19 Apr 2021 15:40

Tgt Ion:	178	Resp:	14980
Ion	Ratio	Lower	Upper
178	100		
176	18.1	14.5	21.7
179	15.1	12.1	18.1









#27
Fluoranthene-d10
Concen: 0.38 ng
RT: 19.175 min Scan# 1351
Delta R.T. 0.000 min
Lab File: BE101304.D
Acq: 19 Apr 2021 15:40




Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4


Tgt Ion:212 Resp: 109744
Ion Ratio Lower Upper
212 100
106 3.3 2.6 4.0
104 1.8 1.4 2.2



#28
Fluoranthene
Concen: 0.39 ng
RT: 19.209 min Scan# 1357
Delta R.T. 0.000 min
Lab File: BE101304.D
Acq: 19 Apr 2021 15:40

Tgt Ion:202 Resp: 24332
Ion Ratio Lower Upper
202 100
101 11.5 9.2 13.8
203 17.1 13.7 20.5









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.318 min Scan# 1600
Delta R.T. 0.000 min
Lab File: BE101304.D
Acq: 19 Apr 2021 15:40




Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4


Tgt Ion:240 Resp: 21390
Ion Ratio Lower Upper
240 100
120 6.6 5.3 7.9
236 26.2 21.0 31.4



#30
Pyrene
Concen: 0.35 ng
RT: 19.565 min Scan# 1419
Delta R.T. 0.000 min
Lab File: BE101304.D
Acq: 19 Apr 2021 15:40

Tgt Ion:202 Resp: 24246
Ion Ratio Lower Upper
202 100
200 20.6 16.5 24.7
203 17.3 13.8 20.8









#31
 Terphenyl-d14
 Concen: 0.37 ng
 RT: 19.778 min Scan# 1456
 Delta R.T. 0.000 min
 Lab File: BE101304.D
 Acq: 19 Apr 2021 15:40




Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.4


Tgt Ion:244 Resp: 19175
 Ion Ratio Lower Upper
 244 100
 212 27.1 21.7 32.5
 122 8.4 6.7 10.1



#32
 Benzo(a)anthracene
 Concen: 0.38 ng
 RT: 21.306 min Scan# 1599
 Delta R.T. 0.000 min
 Lab File: BE101304.D
 Acq: 19 Apr 2021 15:40

Tgt Ion:228 Resp: 25171
 Ion Ratio Lower Upper
 228 100
 226 25.5 20.4 30.6
 229 20.1 16.1 24.1









#33
Chrysene
Concen: 0.40 ng
RT: 21.354 min Scan# 1603
Delta R.T. 0.000 min
Lab File: BE101304.D
Acq: 19 Apr 2021 15:40




Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4


Tgt Ion:228 Resp: 28168
Ion Ratio Lower Upper
228 100
226 29.6 23.7 35.5
229 19.8 15.8 23.8



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.31 ng
RT: 21.233 min Scan# 1593
Delta R.T. 0.000 min
Lab File: BE101304.D
Acq: 19 Apr 2021 15:40

Tgt Ion:149 Resp: 27331
Ion Ratio Lower Upper
149 100
167 7.1 5.7 8.5
279 1.3 1.0 1.6









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.40 ng
RT: 26.352 min Scan# 2866
Delta R.T. 0.000 min
Lab File: BE101304.D
Acq: 19 Apr 2021 15:40




Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4


Tgt Ion:276 Resp: 27604
Ion Ratio Lower Upper
276 100
138 23.5 16.6 24.8
277 24.9 17.6 26.4



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.773 min Scan# 2147
Delta R.T. 0.000 min
Lab File: BE101304.D
Acq: 19 Apr 2021 15:40





Tgt Ion:264 Resp: 21755
Ion Ratio Lower Upper
264 100
260 25.1 20.1 30.1
265 24.2 19.4 29.0








#37
Benzo(b)fluoranthene
Concen: 0.36 ng
RT: 23.017 min Scan# 1935
Delta R.T. 0.000 min
Lab File: BE101304.D
Acq: 19 Apr 2021 15:40

Tgt Ion:252 Resp: 26003
Ion Ratio Lower Upper
252 100
253 22.1 17.7 26.5
125 9.4 7.5 11.3




#38
Benzo(k)fluoranthene
Concen: 0.38 ng
RT: 23.067 min Scan# 1949
Delta R.T. 0.000 min
Lab File: BE101304.D
Acq: 19 Apr 2021 15:40

Tgt Ion:252 Resp: 29223
Ion Ratio Lower Upper
252 100
253 22.6 18.1 27.1
125 11.0 8.8 13.2







Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4



#39
Benzo(a)pyrene
Concen: 0.34 ng
RT: 23.659 min Scan# 2115
Delta R.T. 0.000 min
Lab File: BE101304.D
Acq: 19 Apr 2021 15:40




Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	252	Resp:	22652
Ion Ratio	Lower	Upper	
252	100		
253	22.4	17.9	26.9
125	10.6	8.5	12.7



#40
Dibenzo(a,h)anthracene
Concen: 0.37 ng
RT: 26.377 min Scan# 2873
Delta R.T. 0.000 min
Lab File: BE101304.D
Acq: 19 Apr 2021 15:40


Tgt Ion:	278	Resp:	24233
Ion Ratio	Lower	Upper	
278	100		
139	16.1	12.9	19.3
279	24.1	19.3	28.9





#41
Benzo(g,h,i)perylene
Concen: 0.37 ng
RT: 27.147 min Scan# 3089
Delta R.T. 0.000 min
Lab File: BE101304.D
Acq: 19 Apr 2021 15:40

Tgt Ion:	276	Resp:	24489
Ion	Ratio	Lower	Upper
276	100		
277	24.1	19.3	28.9
138	19.8	15.8	23.8



Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4