

Data Path : Z:\HPCHEM1\BNA E\DATA\BE042016\
 Data File : BE091690.D
 Acq On : 20 Apr 2016 18:12
 Operator : UM/SJ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE042016

Quant Time: Apr 20 19:02:46 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE042016.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 20 18:55:38 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.615	0.631	-2.6	93	0.00
3	n-Nitrosodimethylamine	0.718	0.726	-1.1	107	0.00
4 S	2-Fluorophenol	1.106	1.081	2.3	102	0.00
5 S	Phenol-d6	1.496	1.394	6.8	100	0.00
6	bis(2-Chloroethyl)ether	1.344	1.299	3.3	100	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	99	0.00
8 S	Nitrobenzene-d5	0.267	0.258	3.4	107	0.00
9	Nitrobenzene	0.267	0.256	4.1	105	-0.01
10	Naphthalene	1.019	1.022	-0.3	98	0.00
11	Hexachlorobutadiene	0.146	0.148	-1.4	99	0.00
12	2-Methylnaphthalene	0.654	0.639	2.3	98	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	97	0.00
14 S	2,4,6-Tribromophenol	0.117	0.096	17.9	100	-0.01
15 S	2-Fluorobiphenyl	1.383	1.388	-0.4	97	0.00
16	Acenaphthylene	1.999	1.866	6.7	100	0.00
17	Acenaphthene	1.231	1.213	1.5	98	0.00
18	Fluorene	1.526	1.501	1.6	97	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
20	4-Bromophenyl-phenylether	0.171	0.168	1.8	98	0.00
21	Hexachlorobenzene	0.183	0.182	0.5	96	0.00
22	Pentachlorophenol	0.060	0.043	28.3#	104	-0.02
23	Phenanthrene	1.123	1.149	-2.3	97	-0.01
24	Anthracene	1.002	0.962	4.0	99	-0.01
25	Fluoranthene	1.110	1.095	1.4	101	-0.02
26 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
27	Benzydine	0.292	0.225	22.9	125	-0.02
28	Pvrene	1.155	1.161	-0.5	106	0.00
29 S	Terphenyl-d14	0.793	0.778	1.9	103	0.00
30	Benzo(a)anthracene	1.215	1.240	-2.1	109	0.00
31	3,3'-Dichlorobenzidine	0.510	0.456	10.6	121	-0.02
32	Chrysene	1.324	1.381	-4.3	105	0.00
33	Bis(2-ethylhexyl)phthalate	0.233	0.225	3.4	130	-0.02
34	Indeno(1,2,3-cd)pyrene	1.185	1.154	2.6	105	-0.01
35 I	Perylene-d12	1.000	1.000	0.0	103	-0.01
36	Benzo(b)fluoranthene	1.199	1.166	2.8	106	0.00
37	Benzo(k)fluoranthene	1.180	1.134	3.9	104	-0.01
38 C	Benzo(a)pyrene	1.037	0.977	5.8	106	-0.01
39	Dibenzo(a,h)anthracene	1.063	1.003	5.6	106	-0.01
40	Benzo(a,h,i)perylene	1.090	1.033	5.2	103	-0.01

Data Path : Z:\HPCHEM1\BNA E\DATA\BE042016\
Data File : BE091690.D
Acq On : 20 Apr 2016 18:12
Operator : UM/SJ
Sample : SSTDICV0.4
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE042016

Quant Time: Apr 20 19:02:46 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE042016.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 20 18:55:38 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0