

Data Path : Z:\HPCHEM1\BNA E\DATA\BE042016\
 Data File : BE091687.D
 Acq On : 20 Apr 2016 16:18
 Operator : UM/SJ
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC3.2

Manual Integrations
 APPROVED

Sohil
 4/21/2016 10:35:23 AM

Quant Time: Apr 20 17:19:06 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE042016.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 20 15:30:28 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.80	152	31388	5.00	ng	0.00
7) Naphthalene-d8	10.57	136	155147	5.00	ng	0.00
13) Acenaphthene-d10	14.42	164	94725	5.00	ng	0.00
19) Phenanthrene-d10	17.15	188	223179	5.00	ng	0.00
26) Chrysene-d12	21.31	240	248818	5.00	ng	0.00
35) Perylene-d12	23.75	264	230068	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.41	112	22606	3.31	ng	0.00
5) Phenol-d6	6.95	99	31730	3.45	ng	-0.02
8) Nitrobenzene-d5	8.96	82	28489	3.58	ng	0.00
14) 2,4,6-Tribromophenol	15.90	330	8966m	4.48	ng	0.00
15) 2-Fluorobiphenyl	13.03	172	84803	3.21	ng	-0.01
29) Terphenyl-d14	19.76	244	133408	3.93	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.35	88	11184	2.76	ng	# 90
3) n-Nitrosodimethylamine	3.64	42	15314	3.40	ng	93
6) bis(2-Chloroethyl)ether	7.23	93	28105	3.32	ng	# 76
9) Nitrobenzene	8.99	77	30890	3.86	ng	95
10) Naphthalene	10.63	128	100843	3.19	ng	97
11) Hexachlorobutadiene	10.92	225	14604m	3.19	ng	
12) 2-Methylnaphthalene	12.24	142	67291	3.34	ng	94
16) Acenaphthylene	14.12	152	127649	3.66	ng	# 86
17) Acenaphthene	14.47	154	76041	3.26	ng	89
18) Fluorene	15.46	166	96392	3.38	ng	98
20) 4-Bromophenyl-phenylether	16.33	248	26597	3.44	ng	97
21) Hexachlorobenzene	16.45	284	27392	3.34	ng	96
22) Pentachlorophenol	16.79	266	11257	4.14	ng	88
23) Phenanthrene	17.19	178	166843	3.29	ng	99
24) Anthracene	17.27	178	157705	3.50	ng	98
25) Fluoranthene	19.19	202	172009	3.41	ng	98
27) Benzidine	19.38	184	54697	3.03	ng	# 76
28) Pyrene	19.55	202	197768	3.98	ng	99
30) Benzo(a)anthracene	21.29	228	205485m	3.56	ng	
31) 3,3'-Dichlorobenzidine	21.22	252	97247	4.30	ng	# 80
32) Chrysene	21.33	228	211799m	3.62	ng	
33) Bis(2-ethylhexyl)phthalate	21.20	149	45708	4.32	ng	# 57
34) Indeno(1,2,3-cd)pyrene	26.31	276	196984	3.70	ng	97
36) Benzo(b)fluoranthene	23.00	252	181448m	3.58	ng	
37) Benzo(k)fluoranthene	23.04	252	179418	3.34	ng	97
38) Benzo(a)pyrene	23.63	252	163422	3.46	ng	97
39) Dibenzo(a,h)anthracene	26.33	278	163666	3.55	ng	96
40) Benzo(g,h,i)perylene	27.09	276	163977	3.43	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA E\DATA\BE042016\
Data File : BE091687.D
Acq On : 20 Apr 2016 16:18
Operator : UM/SJ
Sample : SSTDICC3.2
Misc :
ALS Vial : 8 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
SSTDICC3.2

Manual Integrations
APPROVED

Sohil
4/21/2016 10:35:23 AM

Quant Time: Apr 20 17:19:06 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE042016.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 20 15:30:28 2016
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 7.80 min Scan# 300

Delta R.T. 0.00 min

Lab File: BE091687.D

Acq: 20 Apr 2016 16:18

Instrument :

BNA_E





ClientSampleId :

SSTDICC3.2

Tot Ion	Ratio	Lower	Upper
152	100		
150	162.5	122.6	183.8
115	65.7	52.8	79.2

Manual Integrations
APPROVED

Sohil
4/21/2016 10:35:23 AM



#2

1,4-Dioxane

Concen: 2.76 ng




RT: 3.35 min Scan# 44


Delta R.T. -0.00 min

Lab File: BE091687.D

Acq: 20 Apr 2016 16:18

Tgt Ion	Ratio	Lower	Upper
88	100		
58	88.9	65.8	98.6
43	40.5	25.3	37.9





#3





n-Nitrosodimethylamine
Concen: 3.40 ng
RT: 3.64 min Scan# 61
Delta R.T. -0.02 min
Lab File: BE091687.D
Acq: 20 Apr 2016 16:18

Instrument :
BNA_E
ClientSampleId :
SSTDICC3.2

Tot Ion	Ratio	Lower	Upper
42	100		
74	102.4	87.9	131.9
44	6.6	6.6	9.8

Manual Integrations
APPROVED




Sohil
4/21/2016 10:35:23 AM




#4

2-Fluorophenol
Concen: 3.31 ng
RT: 5.41 min Scan# 163
Delta R.T. -0.00 min
Lab File: BE091687.D
Acq: 20 Apr 2016 16:18

Tgt Ion	Ratio	Lower	Upper
112	100		
64	67.4	30.9	46.3#
63	36.8	16.7	25.1#





#5

Phenol-d6

Concen: 3.45 ng

RT: 6.95 min Scan# 251

Delta R.T. -0.02 min

Lab File: BE091687.D

Acq: 20 Apr 2016 16:18

Instrument :

BNA_E


ClientSampleId :

SSTDICC3.2




Tot Ion:	99	Resp:	31730
Ion Ratio	Lower	Upper	
99	100		
42	27.2	16.2	24.2#
71	36.2	29.0	43.4

Manual Integrations
APPROVED

Sohil
4/21/2016 10:35:23 AM



Abundance Ion 99.00 (98.70 to 99.70): BE0
Ion 42.00 (41.70 to 42.70): BE0
Ion 71.00 (70.70 to 71.70): BE0



#6

bis(2-Chloroethyl)ether

Concen: 3.32 ng


RT: 7.23 min Scan# 267

Delta R.T. 0.00 min



Lab File: BE091687.D


Acq: 20 Apr 2016 16:18

Tgt Ion:	93	Resp:	28105
Ion Ratio	Lower	Upper	
93	100		
63	86.0	49.5	74.3#
95	32.3	22.2	33.4



Abundance Ion 93.00 (92.70 to 93.70): BE0
Ion 63.00 (62.70 to 63.70): BE0
Ion 95.00 (94.70 to 95.70): BE0





#7

Naphthalene-d8

Concen: 5.00 ng

RT: 10.57 min Scan# 511

Delta R.T. -0.00 min

Lab File: BE091687.D

Acq: 20 Apr 2016 16:18

Instrument :

BNA_E


ClientSampleId :

SSTDICC3.2

Tot Ion	Ratio	Resp	Lower	Upper
136	100	155147		
137	10.3	8.9	13.3	
54	9.7	8.2	12.4	
68	6.7	6.2	9.4	

Manual Integrations
APPROVED

Sohil
4/21/2016 10:35:23 AM






Abundance Ion 136.00 (135.70 to 136.70): BE091687.D (-507) (-)

Ion 137.00 (136.70 to 137.70): BE091687.D (-507) (-)

Ion 54.00 (53.70 to 54.70): BE091687.D (-507) (-)

Ion 68.00 (67.70 to 68.70): BE091687.D (-507) (-)



#8

Nitrobenzene-d5

Concen: 3.58 ng


RT: 8.96 min Scan# 392

Delta R.T. -0.00 min

Lab File: BE091687.D

Acq: 20 Apr 2016 16:18



Tgt Ion	Ratio	Resp	Lower	Upper
82	100	28489		
128	39.4	25.4	38.0#	
54	44.3	48.4	72.6#	




Abundance Ion 82.00 (81.70 to 82.70): BE091687.D (-375) (-)

Ion 128.00 (127.70 to 128.70): BE091687.D (-375) (-)

Ion 54.00 (53.70 to 54.70): BE091687.D (-375) (-)





#9

Nitrobenzene

Concen: 3.86 ng

RT: 8.99 min Scan# 395

Delta R.T. -0.01 min

Lab File: BE091687.D

Acq: 20 Apr 2016 16:18

Instrument :

BNA_E





ClientSampleId :

SSTDICC3.2

Tot Ion:	77	Resp:	30890
Ion	Ratio	Lower	Upper
77	100		
123	36.6	26.9	40.3
65	13.8	9.4	14.2

Manual Integrations
APPROVED

Sohil
4/21/2016 10:35:23 AM



#10

Naphthalene

Concen: 3.19 ng




RT: 10.63 min Scan# 514


Delta R.T. -0.00 min

Lab File: BE091687.D

Acq: 20 Apr 2016 16:18

Tgt Ion:	128	Resp:	100843
Ion	Ratio	Lower	Upper
128	100		
129	11.8	9.3	13.9
127	11.7	10.7	16.1









#11
Hexachlorobutadiene
Concen: 3.19 ng m
RT: 10.92 min Scan# 529
Delta R.T. -0.00 min
Lab File: BE091687.D
Acq: 20 Apr 2016 16:18

Instrument :
BNA_E
ClientSampleId :
SSTDICC3.2

Tot Ion	Ratio	Lower	Upper
225	100		
223	70.5	49.0	73.6
227	71.3	50.3	75.5




Manual Integrations
APPROVED


Sohil
4/21/2016 10:35:23 AM



#12
2-Methylnaphthalene
Concen: 3.34 ng
RT: 12.24 min Scan# 638
Delta R.T. -0.00 min
Lab File: BE091687.D
Acq: 20 Apr 2016 16:18

Tgt Ion	Ratio	Lower	Upper
142	100		
141	85.0	71.4	107.0
115	32.6	30.3	45.5









#13
Acenaphthene-d10
Concen: 5.00 ng
RT: 14.42 min Scan# 820
Delta R.T. -0.00 min
Lab File: BE091687.D
Acq: 20 Apr 2016 16:18

Instrument :
BNA_E
ClientSampleId :
SSTDICC3.2

Tot Ion	Ratio	Lower	Upper
164	100		
162	94.9	85.3	127.9
160	38.5	46.2	69.2#




Manual Integrations
APPROVED


Sohil
4/21/2016 10:35:23 AM



#14
2,4,6-Tribromophenol
Concen: 4.48 ng m
RT: 15.90 min Scan# 939
Delta R.T. 0.00 min
Lab File: BE091687.D
Acq: 20 Apr 2016 16:18

Tgt Ion	Ratio	Lower	Upper
330	100		
332	114.8	79.1	118.7
141	46.6	125.4	188.0#









#15
2-Fluorobiphenyl
Concen: 3.21 ng
RT: 13.03 min Scan# 707
Delta R.T. -0.01 min
Lab File: BE091687.D
Acq: 20 Apr 2016 16:18

Instrument :
BNA_E
ClientSampleId :
SSTDICC3.2

Tot Ion	Ratio	Lower	Upper
172	100		
171	37.7	28.8	43.2
170	27.3	19.3	28.9




Manual Integrations
APPROVED


Sohil
4/21/2016 10:35:23 AM



#16
Acenaphthylene
Concen: 3.66 ng
RT: 14.12 min Scan# 798
Delta R.T. -0.01 min
Lab File: BE091687.D
Acq: 20 Apr 2016 16:18

Tgt Ion	Ratio	Lower	Upper
152	100		
151	20.3	3.9	5.9#
153	13.8	10.3	15.5









#17
Acenaphthene
Concen: 3.26 ng
RT: 14.47 min Scan# 824
Delta R.T. -0.00 min
Lab File: BE091687.D
Acq: 20 Apr 2016 16:18

Instrument :
BNA_E
ClientSampleId :
SSTDICC3.2

Tot Ion	Ratio	Lower	Upper
154	100		
153	112.7	80.0	120.0
152	56.2	40.0	60.0




Manual Integrations
APPROVED


Sohil
4/21/2016 10:35:23 AM



#18
Fluorene
Concen: 3.38 ng
RT: 15.46 min Scan# 901
Delta R.T. 0.00 min
Lab File: BE091687.D
Acq: 20 Apr 2016 16:18

Tgt Ion	Ratio	Lower	Upper
166	100		
165	98.7	80.8	121.2
167	13.6	10.9	16.3









#19
Phenanthrene-d10
Concen: 5.00 ng
RT: 17.15 min Scan# 1015
Delta R.T. -0.00 min
Lab File: BE091687.D
Acq: 20 Apr 2016 16:18

Instrument :
BNA_E
ClientSampleId :
SSTDICC3.2

Tot Ion	Ratio	Resp Lower	Resp Upper
188	100		
94	9.6	8.7	13.1
80	9.4	9.4	14.0




Manual Integrations
APPROVED


Sohil
4/21/2016 10:35:23 AM



#20
4-Bromophenyl-phenylether
Concen: 3.44 ng
RT: 16.33 min Scan# 966
Delta R.T. -0.02 min
Lab File: BE091687.D
Acq: 20 Apr 2016 16:18

Tgt Ion	Ratio	Resp Lower	Resp Upper
248	100		
250	98.2	80.5	120.7
141	87.0	72.0	108.0









#21
Hexachlorobenzene
Concen: 3.34 ng
RT: 16.45 min Scan# 973
Delta R.T. -0.02 min
Lab File: BE091687.D
Acq: 20 Apr 2016 16:18

Instrument :
BNA_E
ClientSampleId :
SSTDICC3.2

Tot Ion	Ratio	Lower	Upper
284	100		
142	42.6	32.2	48.2
249	33.5	25.4	38.2




Manual Integrations
APPROVED


Sohil
4/21/2016 10:35:23 AM



#22
Pentachlorophenol
Concen: 4.14 ng
RT: 16.79 min Scan# 993
Delta R.T. -0.02 min
Lab File: BE091687.D
Acq: 20 Apr 2016 16:18

Tgt Ion	Ratio	Lower	Upper
266	100		
268	58.6	58.2	87.2
264	65.3	56.2	84.4









#23
Phenanthrene
Concen: 3.29 ng
RT: 17.19 min Scan# 1019
Delta R.T. -0.00 min
Lab File: BE091687.D
Acq: 20 Apr 2016 16:18

Instrument :
BNA_E
ClientSampleId :
SSTDICC3.2

Tot Ion	Ratio	Lower	Upper
178	100		
176	19.6	15.3	22.9
179	15.7	12.6	18.8




Manual Integrations
APPROVED


Sohil
4/21/2016 10:35:23 AM



#24
Anthracene
Concen: 3.50 ng
RT: 17.27 min Scan# 1026
Delta R.T. -0.01 min
Lab File: BE091687.D
Acq: 20 Apr 2016 16:18

Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.1	14.6	21.8
179	15.4	11.8	17.6





#25

Fluoranthene

Concen: 3.41 ng

RT: 19.19 min Scan# 1149

Delta R.T. -0.02 min

Lab File: BE091687.D

Acq: 20 Apr 2016 16:18

Instrument :

BNA_E


ClientSampleId :

SSTDICC3.2

Tot Ion:	202	Resp:	172009
Ion Ratio	Lower	Upper	
202	100		
101	12.1	11.0	16.6
203	17.5	13.9	20.9

Manual Integrations
APPROVED




Sohil
4/21/2016 10:35:23 AM



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#26

Chrysene-d12

Concen: 5.00 ng


RT: 21.31 min Scan# 1247

Delta R.T. -0.00 min

Lab File: BE091687.D

Acq: 20 Apr 2016 16:18



Tgt Ion:	240	Resp:	248818
Ion Ratio	Lower	Upper	
240	100		
120	1.2	11.0	16.4#
236	23.1	21.7	32.5




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E





#27

Benzidine

Concen: 3.03 ng

RT: 19.38 min Scan# 1159

Delta R.T. -0.02 min

Lab File: BE091687.D

Acq: 20 Apr 2016 16:18

Instrument :

BNA_E





ClientSampleId :

SSTDICC3.2

Tot Ion:	184	Resp:	54697
Ion Ratio	Lower	Upper	
184	100		
185	14.1	22.3	33.5#
183	10.6	16.4	24.6#

Manual Integrations
APPROVED

Sohil
4/21/2016 10:35:23 AM



#28

Pyrene

Concen: 3.98 ng




RT: 19.55 min Scan# 1168


Delta R.T. -0.00 min

Lab File: BE091687.D

Acq: 20 Apr 2016 16:18

Tgt Ion:	202	Resp:	197768
Ion Ratio	Lower	Upper	
202	100		
200	21.7	16.8	25.2
203	17.3	14.0	21.0









#29
 Terphenyl-d14
 Concen: 3.93 ng
 RT: 19.76 min Scan# 1179
 Delta R.T. -0.00 min
 Lab File: BE091687.D
 Acq: 20 Apr 2016 16:18

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC3.2

Tot Ion	Ratio	Lower	Upper
244	100		
212	8.1	6.9	10.3
122	2.6	11.0	16.4




Manual Integrations
 APPROVED


Sohil
 4/21/2016 10:35:23 AM



#30
 Benzo(a)anthracene
 Concen: 3.56 ng m
 RT: 21.29 min Scan# 1246
 Delta R.T. -0.00 min
 Lab File: BE091687.D
 Acq: 20 Apr 2016 16:18

Tgt Ion	Ratio	Lower	Upper
228	100		
226	23.7	21.0	31.4
229	21.3	15.7	23.5






#31
 3,3'-Dichlorobenzidine
 Concen: 4.30 ng
 RT: 21.22 min Scan# 1243
 Delta R.T. -0.02 min
 Lab File: BE091687.D
 Acq: 20 Apr 2016 16:18

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC3.2




Tot Ion	Ratio	Lower	Upper
252	100		
254	45.0	51.1	76.7#
126	17.4	12.6	18.8

Manual Integrations
 APPROVED

Sohil
 4/21/2016 10:35:23 AM






Abundance Ion 252.00 (251.70 to 252.70): E
 Ion 254.00 (253.70 to 254.70): E
 Ion 126.00 (125.70 to 126.70): E




#32
 Chrysene
 Concen: 3.62 ng m
 RT: 21.33 min Scan# 1248
 Delta R.T. -0.00 min
 Lab File: BE091687.D
 Acq: 20 Apr 2016 16:18

Tgt Ion	Ratio	Lower	Upper
228	100		
226	32.8	24.0	36.0
229	17.2	15.9	23.9

Abundance Ion 228.00 (227.70 to 228.70): E
 Ion 226.00 (225.70 to 226.70): E
 Ion 229.00 (228.70 to 229.70): E









#33
 Bis(2-ethylhexyl)phthalate
 Concen: 4.32 ng
 RT: 21.20 min Scan# 1242
 Delta R.T. -0.00 min
 Lab File: BE091687.D
 Acq: 20 Apr 2016 16:18

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC3.2

Tot Ion	Ratio	Lower	Upper
149	100		
167	0.0	19.5	29.3#
279	0.0	10.0	15.0#




Manual Integrations
 APPROVED


Sohil
 4/21/2016 10:35:23 AM



#34
 Indeno(1,2,3-cd)pyrene
 Concen: 3.70 ng
 RT: 26.31 min Scan# 1644
 Delta R.T. -0.00 min
 Lab File: BE091687.D
 Acq: 20 Apr 2016 16:18

Tgt Ion	Ratio	Lower	Upper
276	100		
138	26.3	23.4	35.2
277	24.6	19.8	29.8









#35
 Pervlene-d12
 Concen: 5.00 ng
 RT: 23.75 min Scan# 1425
 Delta R.T. -0.00 min
 Lab File: BE091687.D
 Acq: 20 Apr 2016 16:18

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC3.2

Tot Ion	Ratio	Lower	Upper
264	100		
260	25.0	20.0	30.0
265	21.6	17.5	26.3




Manual Integrations
 APPROVED


Sohil
 4/21/2016 10:35:23 AM



#36
 Benzo(b)fluoranthene
 Concen: 3.58 ng m
 RT: 23.00 min Scan# 1360
 Delta R.T. -0.00 min
 Lab File: BE091687.D
 Acq: 20 Apr 2016 16:18

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.4	26.0
125	10.0	10.2	15.2#





#37

Benzo(k)fluoranthene

Concen: 3.34 ng

RT: 23.04 min Scan# 1364

Delta R.T. -0.00 min

Lab File: BE091687.D

Acq: 20 Apr 2016 16:18

Instrument :

BNA_E





ClientSampleId :

SSTDICC3.2

Tot Ion:	252	Resp:	179418
Ion Ratio	Lower	Upper	
252	100		
253	21.5	17.6	26.4
125	10.3	9.9	14.9

Manual Integrations
APPROVED

Sohil
4/21/2016 10:35:23 AM



#38

Benzo(a)pyrene

Concen: 3.46 ng




RT: 23.63 min Scan# 1415


Delta R.T. -0.00 min

Lab File: BE091687.D

Acq: 20 Apr 2016 16:18

Tgt Ion:	252	Resp:	163422
Ion Ratio	Lower	Upper	
252	100		
253	21.7	18.1	27.1
125	11.7	11.0	16.4









#39
Dibenzo(a,h)anthracene
Concen: 3.55 ng
RT: 26.33 min Scan# 1646
Delta R.T. -0.01 min
Lab File: BE091687.D
Acq: 20 Apr 2016 16:18

Instrument :
BNA_E
ClientSampleId :
SSTDICC3.2

Tot Ion	Ratio	Lower	Upper
278	100		
139	16.6	16.0	24.0
279	23.4	19.4	29.2

Manual Integrations
APPROVED

Sohil
4/21/2016 10:35:23 AM



#40
Benzo(g,h,i)perylene
Concen: 3.43 ng
RT: 27.09 min Scan# 1712
Delta R.T. -0.00 min
Lab File: BE091687.D
Acq: 20 Apr 2016 16:18

Tgt Ion	Ratio	Lower	Upper
276	100		
277	23.7	18.6	28.0
138	22.3	20.4	30.6

