

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE042017\
 Data File : BE092899.D
 Acq On : 20 Apr 2017 14:33
 Operator : SJ/MA
 Sample : SSTDICC0.2
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.2

Manual Integrations
 APPROVED

mohammad
 4/21/2017 11:44:23 AM

Quant Time: Apr 20 16:24:16 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE042017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 20 16:20:01 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.77	152	676	0.40	ng	0.00
7) Naphthalene-d8	10.53	136	2797	0.40	ng	0.00
13) Acenaphthene-d10	14.39	164	1278	0.40	ng	0.00
19) Phenanthrene-d10	17.13	188	3249	0.40	ng	0.00
26) Chrysene-d12	21.28	240	3089	0.40	ng	0.00
33) Perylene-d12	23.70	264	2164	0.40	ng	0.01

System Monitoring Compounds						
4) 2-Fluorophenol	5.35	112	290	0.20	ng	0.00
5) Phenol-d6	6.94	99	398	0.20	ng	0.00
8) Nitrobenzene-d5	8.90	82	358	0.20	ng	0.00
11) 2-Methylnaphthalene-d10	12.13	152	852	0.21	ng	0.00
14) 2,4,6-Tribromophenol	15.87	330	42	0.20	ng	0.00
15) 2-Fluorobiphenyl	13.01	172	1215	0.20	ng	0.00
24) Fluoranthene-d10	19.15	212	1521	0.20	ng	0.00
28) Terphenyl-d14	19.74	244	1076	0.19	ng	0.00

Target Compounds				Qvalue		
2) 1,4-Dioxane	3.31	88	264	0.20	ng	99
3) n-Nitrosodimethylamine	3.61	42	181	0.20	ng	# 86
6) bis(2-Chloroethyl)ether	7.20	93	520	0.21	ng	98
9) Naphthalene	10.58	128	1516	0.20	ng	92
10) Hexachlorobutadiene	10.89	225	214	0.21	ng	98
12) 2-Methylnaphthalene	12.19	142	928	0.21	ng	97
16) Acenaphthylene	14.10	152	3781	0.20	ng	99
17) Acenaphthene	14.46	154	806	0.20	ng	95
18) Fluorene	15.43	166	969	0.20	ng	99
20) Hexachlorobenzene	16.47	284	278	0.20	ng	# 100
21) Pentachlorophenol	16.79	266	44	0.17	ng	93
22) Phenanthrene	17.18	178	1830	0.20	ng	99
23) Anthracene	17.27	178	1355	0.20	ng	99
25) Fluoranthene	19.17	202	7587	0.20	ng	# 98
27) Pyrene	19.53	202	8081	0.20	ng	# 100
29) Benzo(a)anthracene	21.27	228	1358	0.19	ng	# 85
30) Chrysene	21.32	228	1821	0.20	ng	# 81
31) Bis(2-ethylhexyl)phthalate	21.20	149	386	0.20	ng	# 99
32) Indeno(1,2,3-cd)pyrene	26.22	276	5711	0.19	ng	99
34) Benzo(b)fluoranthene	22.95	252	1476m	0.20	ng	
35) Benzo(k)fluoranthene	22.99	252	1444	0.20	ng	# 88
36) Benzo(a)pyrene	23.58	252	1174	0.20	ng	# 86
37) Dibenzo(a,h)anthracene	26.25	278	1249	0.20	ng	# 87
38) Benzo(g,h,i)perylene	27.00	276	5764	0.20	ng	# 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA_E\DATA\BE042017\
Data File : BE092899.D
Acq On : 20 Apr 2017 14:33
Operator : SJ/MA
Sample : SSTDICC0.2
Misc :
ALS Vial : 3 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
SSTDICC0.2

Manual Integrations
APPROVED

mohammad
4/21/2017 11:44:23 AM

Quant Time: Apr 20 16:24:16 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE042017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 20 16:20:01 2017
Response via : Initial Calibration





#1


1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.77 min Scan# 555
Delta R.T. -0.00 min
Lab File: BE092899.D
Acq: 20 Apr 2017 14:33

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.2


Tgt Ion	Ratio	Lower	Upper
152	100		
150	166.7	129.9	194.9
115	84.3	62.6	93.8

Manual Integrations
APPROVED



mohammad
4/21/2017 11:44:23 AM



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E




Time-->




#2

1,4-Dioxane
Concen: 0.20 ng
RT: 3.31 min Scan# 62
Delta R.T. 0.00 min
Lab File: BE092899.D
Acq: 20 Apr 2017 14:33


Tgt Ion	Ratio	Lower	Upper
88	100		
58	80.7	64.9	97.3
43	31.1	26.6	40.0




Abundance Ion 88.00 (87.70 to 88.70): BE0
Ion 58.00 (57.70 to 58.70): BE0
Ion 43.00 (42.70 to 43.70): BE0



Time-->





#3





n-Nitrosodimethylamine
Concen: 0.20 ng
RT: 3.61 min Scan# 88
Delta R.T. 0.00 min
Lab File: BE092899.D
Acq: 20 Apr 2017 14:33

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.2

Tgt Ion: 42 Resp: 181
Ion Ratio Lower Upper
42 100
74 120.4 106.6 160.0
44 17.1 25.7 38.5#

Manual Integrations
APPROVED




mohammad
4/21/2017 11:44:23 AM




#4

2-Fluorophenol
Concen: 0.20 ng
RT: 5.35 min Scan# 266
Delta R.T. -0.00 min
Lab File: BE092899.D
Acq: 20 Apr 2017 14:33

Tgt Ion: 112 Resp: 290
Ion Ratio Lower Upper
112 100
64 69.7 52.6 79.0
63 35.2 29.0 43.4





#5

Phenol-d6

Concen: 0.20 ng

RT: 6.94 min Scan# 519

Delta R.T. -0.00 min

Lab File: BE092899.D

Acq: 20 Apr 2017 14:33

Instrument :

BNA_E





ClientSampleId :

SSTDIC0.2

Tgt Ion:	99	Resp:	398
Ion	Ratio	Lower	Upper
99	100		
42	0.0	0.0	0.0
71	0.0	0.0	0.0

Manual Integrations
APPROVED

mohammad
4/21/2017 11:44:23 AM



#6

bis(2-Chloroethyl)ether

Concen: 0.21 ng




RT: 7.20 min Scan# 530


Delta R.T. -0.00 min

Lab File: BE092899.D

Acq: 20 Apr 2017 14:33

Tgt Ion:	93	Resp:	520
Ion	Ratio	Lower	Upper
93	100		
63	75.0	61.2	91.8
95	30.4	25.5	38.3





#7

Naphthalene-d8

Concen: 0.40 ng

RT: 10.53 min Scan# 678

Delta R.T. -0.00 min

Lab File: BE092899.D

Acq: 20 Apr 2017 14:33

Instrument :

BNA_E


ClientSampleId :

SSTDIC0.2

Tgt Ion:	136	Resp:	2797
Ion	Ratio	Lower	Upper
136	100		
137	12.9	10.4	15.6
54	15.2	12.6	18.8
68	12.8	10.7	16.1

Manual Integrations
APPROVED

mohammad
4/21/2017 11:44:23 AM






Abundance Ion 136.00 (135.70 to 136.70): BE092899.D (-668) (-)

Ion 137.00 (136.70 to 137.70): BE092899.D (-668) (-)

Ion 54.00 (53.70 to 54.70): BE092899.D (-668) (-)

Ion 68.00 (67.70 to 68.70): BE092899.D (-668) (-)



#8

Nitrobenzene-d5

Concen: 0.20 ng


RT: 8.90 min Scan# 604

Delta R.T. -0.00 min

Lab File: BE092899.D

Acq: 20 Apr 2017 14:33



Tgt Ion:	82	Resp:	358
Ion	Ratio	Lower	Upper
82	100		
128	31.9	35.3	52.9#
54	80.5	65.1	97.7




Abundance Ion 82.00 (81.70 to 82.70): BE092899.D (-595) (-)

Ion 128.00 (127.70 to 128.70): BE092899.D (-595) (-)

Ion 54.00 (53.70 to 54.70): BE092899.D (-595) (-)





#9

Naphthalene

Concen: 0.20 ng

RT: 10.58 min Scan# 681

Delta R.T. -0.00 min

Lab File: BE092899.D

Acq: 20 Apr 2017 14:33

Instrument :

BNA_E





ClientSampleId :

SSTDIC0.2

Tgt Ion:	128	Resp:	1516
Ion Ratio	Lower	Upper	
128	100		
129	16.2	15.8	23.8
127	16.4	16.0	24.0

Manual Integrations
APPROVED

mohammad
4/21/2017 11:44:23 AM



#10

Hexachlorobutadiene

Concen: 0.21 ng




RT: 10.89 min Scan# 697


Delta R.T. -0.00 min

Lab File: BE092899.D

Acq: 20 Apr 2017 14:33

Tgt Ion:	225	Resp:	214
Ion Ratio	Lower	Upper	
225	100		
223	66.8	55.4	83.2
227	64.0	52.2	78.4









#11
2-Methylnaphthalene-d10
Concen: 0.21 ng
RT: 12.13 min Scan# 761
Delta R.T. 0.00 min
Lab File: BE092899.D
Acq: 20 Apr 2017 14:33

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.2

Tgt Ion:152 Resp: 852
Ion Ratio Lower Upper
152 100
151 19.4 15.7 23.5




Manual Integrations
APPROVED


mohammad
4/21/2017 11:44:23 AM



#12
2-Methylnaphthalene
Concen: 0.21 ng
RT: 12.19 min Scan# 765
Delta R.T. 0.00 min
Lab File: BE092899.D
Acq: 20 Apr 2017 14:33

Tgt Ion:142 Resp: 928
Ion Ratio Lower Upper
142 100
141 95.1 76.0 114.0
115 45.2 41.6 62.4





#13

Acenaphthene-d10

Concen: 0.40 ng

RT: 14.39 min Scan# 901

Delta R.T. 0.00 min

Lab File: BE092899.D

Acq: 20 Apr 2017 14:33

Instrument :

BNA_E





ClientSampleId :

SSTDIC0.2

Tgt Ion:	164	Resp:	1278
Ion	Ratio	Lower	Upper
164	100		
162	107.0	86.6	130.0
160	52.0	44.5	66.7

Manual Integrations
APPROVED

mohammad
4/21/2017 11:44:23 AM



#14

2,4,6-Tribromophenol

Concen: 0.20 ng




RT: 15.87 min Scan# 1010


Delta R.T. 0.00 min

Lab File: BE092899.D

Acq: 20 Apr 2017 14:33

Tgt Ion:	330	Resp:	42
Ion	Ratio	Lower	Upper
330	100		
332	88.1	69.1	103.7
141	57.1	54.6	81.8









#15
2-Fluorobiphenyl
Concen: 0.20 ng
RT: 13.01 min Scan# 818
Delta R.T. 0.00 min
Lab File: BE092899.D
Acq: 20 Apr 2017 14:33

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.2

Tgt Ion	Ratio	Lower	Upper
172	100		
171	41.6	35.8	53.6
170	30.1	28.5	42.7




Manual Integrations
APPROVED


mohammad
4/21/2017 11:44:23 AM



#16
Acenaphthylene
Concen: 0.20 ng
RT: 14.10 min Scan# 884
Delta R.T. 0.00 min
Lab File: BE092899.D
Acq: 20 Apr 2017 14:33

Tgt Ion	Ratio	Lower	Upper
152	100		
151	4.9	3.8	5.8
153	12.7	10.0	15.0









#17
Acenaphthene
Concen: 0.20 ng
RT: 14.46 min Scan# 905
Delta R.T. 0.00 min
Lab File: BE092899.D
Acq: 20 Apr 2017 14:33

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.2

Tgt Ion	Ratio	Lower	Upper
154	100		
153	457.9	355.6	533.4
152	230.8	177.8	266.8




Manual Integrations
APPROVED


mohammad
4/21/2017 11:44:23 AM



#18
Fluorene
Concen: 0.20 ng
RT: 15.43 min Scan# 972
Delta R.T. -0.01 min
Lab File: BE092899.D
Acq: 20 Apr 2017 14:33

Tgt Ion	Ratio	Lower	Upper
166	100		
165	101.7	80.7	121.1
167	13.6	11.3	16.9









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.13 min Scan# 1078
Delta R.T. 0.00 min
Lab File: BE092899.D
Acq: 20 Apr 2017 14:33

Instrument :
BNA_E
ClientSampleId :
SSTDICC0.2

Tgt Ion	Ratio	Resp	Lower	Upper
188	100	3249		
94	13.4	10.2	15.4	
80	13.4	10.6	15.8	




Manual Integrations
APPROVED


mohammad
4/21/2017 11:44:23 AM



#20
Hexachlorobenzene
Concen: 0.20 ng
RT: 16.47 min Scan# 1049
Delta R.T. 0.00 min
Lab File: BE092899.D
Acq: 20 Apr 2017 14:33

Tgt Ion	Ratio	Resp	Lower	Upper
284	100	278		
142	0.0	0.0	0.0	
249	0.0	0.0	0.0	





#21

Pentachlorophenol

Concen: 0.17 ng

RT: 16.79 min Scan# 1063

Delta R.T. 0.00 min

Lab File: BE092899.D


Acq: 20 Apr 2017 14:33

Instrument :

BNA_E

ClientSampleId :




SSTDIC0.2



Tgt Ion:	266	Resp:	44
Ion	Ratio	Lower	Upper
266	100		
264	89.1	75.4	113.2
268	85.9	75.4	113.2

Manual Integrations
APPROVED

mohammad
4/21/2017 11:44:23 AM



#22

Phenanthrene

Concen: 0.20 ng




RT: 17.18 min Scan# 1080


Delta R.T. 0.00 min

Lab File: BE092899.D

Acq: 20 Apr 2017 14:33

Tgt Ion:	178	Resp:	1830
Ion	Ratio	Lower	Upper
178	100		
176	18.5	14.2	21.4
179	16.6	13.6	20.4





#23

Anthracene

Concen: 0.20 ng

RT: 17.27 min Scan# 1084

Delta R.T. 0.00 min

Lab File: BE092899.D

Acq: 20 Apr 2017 14:33

Instrument :

BNA_E

ClientSampleId :





SSTDIC0.2

Tgt Ion:178 Resp: 1355

Ion	Ratio	Lower	Upper
178	100		
176	17.6	14.7	22.1
179	15.8	12.8	19.2

Manual Integrations
APPROVED

mohammad
4/21/2017 11:44:23 AM



#24

Fluoranthene-d10

Concen: 0.20 ng

RT: 19.15 min Scan# 1165




Delta R.T. -0.00 min


Lab File: BE092899.D

Acq: 20 Apr 2017 14:33

Tgt Ion:212 Resp: 1521

Ion	Ratio	Lower	Upper
212	100		
106	0.0	0.0	0.0
104	0.0	0.0	0.0





#25

Fluoranthene

Concen: 0.20 ng

RT: 19.17 min Scan# 1166

Delta R.T. -0.00 min

Lab File: BE092899.D

Acq: 20 Apr 2017 14:33

Instrument :

BNA_E





ClientSampleId :

SSTDIC0.2

Tgt Ion:	202	Resp:	7587
Ion Ratio	Lower	Upper	
202	100		
101	0.0	0.0	0.0
203	18.3	15.3	22.9

Manual Integrations
APPROVED

mohammad
4/21/2017 11:44:23 AM



#26

Chrysene-d12

Concen: 0.40 ng




RT: 21.28 min Scan# 1268


Delta R.T. 0.00 min

Lab File: BE092899.D

Acq: 20 Apr 2017 14:33

Tgt Ion:	240	Resp:	3089
Ion Ratio	Lower	Upper	
240	100		
120	4.6	4.6	6.8
236	24.3	19.8	29.8









#27
Pyrene
Concen: 0.20 ng
RT: 19.53 min Scan# 1183
Delta R.T. -0.00 min
Lab File: BE092899.D
Acq: 20 Apr 2017 14:33

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.2

Tgt Ion	Ratio	Lower	Upper
202	100		
200	0.0	0.0	0.0
203	17.1	13.5	20.3




Manual Integrations
APPROVED


mohammad
4/21/2017 11:44:23 AM



#28
Terphenyl-d14
Concen: 0.19 ng
RT: 19.74 min Scan# 1193
Delta R.T. -0.00 min
Lab File: BE092899.D
Acq: 20 Apr 2017 14:33

Tgt Ion	Ratio	Lower	Upper
244	100		
212	16.9	18.4	27.6#
122	27.5	23.3	34.9





#29

Benzo(a)anthracene

Concen: 0.19 ng

RT: 21.27 min Scan# 1267

Delta R.T. 0.00 min

Lab File: BE092899.D

Acq: 20 Apr 2017 14:33

Instrument :

BNA_E





ClientSampleId :

SSTDICC0.2

Tgt Ion:	228	Resp:	1358
Ion Ratio	Lower	Upper	
228	100		
226	26.0	32.0	48.0#
229	28.4	21.4	32.0

Manual Integrations
APPROVED

mohammad
4/21/2017 11:44:23 AM



#30

Chrysene

Concen: 0.20 ng




RT: 21.32 min Scan# 1270


Delta R.T. 0.00 min

Lab File: BE092899.D

Acq: 20 Apr 2017 14:33

Tgt Ion:	228	Resp:	1821
Ion Ratio	Lower	Upper	
228	100		
226	27.3	34.2	51.2#
229	27.3	18.6	28.0









#31
 Bis(2-ethylhexyl)phthalate
 Concen: 0.20 ng
 RT: 21.20 min Scan# 1263
 Delta R.T. 0.00 min
 Lab File: BE092899.D
 Acq: 20 Apr 2017 14:33

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.2

Tgt Ion	Ratio	Lower	Upper
149	100		
167	25.6	20.9	31.3
279	0.0	0.0	0.0




Manual Integrations
 APPROVED


mohammad
 4/21/2017 11:44:23 AM



#32
 Indeno(1,2,3-cd)pyrene
 Concen: 0.19 ng
 RT: 26.22 min Scan# 1650
 Delta R.T. -0.00 min
 Lab File: BE092899.D
 Acq: 20 Apr 2017 14:33

Tgt Ion	Ratio	Lower	Upper
276	100		
138	32.0	25.8	38.6
277	24.5	20.2	30.4









#33
 Perylene-d12
 Concen: 0.40 ng
 RT: 23.70 min Scan# 1440
 Delta R.T. 0.01 min
 Lab File: BE092899.D
 Acq: 20 Apr 2017 14:33

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.2

Tgt Ion	Ratio	Lower	Upper
264	100		
260	27.2	23.5	35.3
265	36.2	27.2	40.8




Manual Integrations
 APPROVED


mohammad
 4/21/2017 11:44:23 AM



#34
 Benzo(b)fluoranthene
 Concen: 0.20 ng m
 RT: 22.95 min Scan# 1375
 Delta R.T. -0.00 min
 Lab File: BE092899.D
 Acq: 20 Apr 2017 14:33

Tgt Ion	Ratio	Lower	Upper
252	100		
253	26.3	24.7	37.1
125	19.3	20.3	30.5#









#35
Benzo(k)fluoranthene
Concen: 0.20 ng
RT: 22.99 min Scan# 1379
Delta R.T. -0.00 min
Lab File: BE092899.D
Acq: 20 Apr 2017 14:33

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.2

Tgt Ion	Ratio	Lower	Upper
252	100		
253	26.5	25.8	38.8
125	21.5	22.9	34.3




Manual Integrations
APPROVED


mohammad
4/21/2017 11:44:23 AM



#36
Benzo(a)pyrene
Concen: 0.20 ng
RT: 23.58 min Scan# 1430
Delta R.T. -0.00 min
Lab File: BE092899.D
Acq: 20 Apr 2017 14:33

Tgt Ion	Ratio	Lower	Upper
252	100		
253	31.7	31.3	46.9
125	24.3	26.4	39.6









#37
 Dibenzo(a,h)anthracene
 Concen: 0.20 ng
 RT: 26.25 min Scan# 1652
 Delta R.T. -0.00 min
 Lab File: BE092899.D
 Acq: 20 Apr 2017 14:33

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.2

Tgt Ion	Ratio	Lower	Upper
278	100		
139	27.3	27.4	41.2#
279	33.0	33.3	49.9#

Manual Integrations
 APPROVED

mohammad
 4/21/2017 11:44:23 AM



#38
 Benzo(g,h,i)perylene
 Concen: 0.20 ng
 RT: 27.00 min Scan# 1696
 Delta R.T. 0.02 min
 Lab File: BE092899.D
 Acq: 20 Apr 2017 14:33

Tgt Ion	Ratio	Lower	Upper
276	100		
277	31.9	29.9	44.9
138	31.3	34.6	51.8#

