

Data Path : Z:\HPCHEM1\BNA E\DATA\BE042018\
 Data File : BE096074.D
 Acq On : 21 Apr 2018 11:21
 Operator : SJ/JU
 Sample : SSTDC00.4EC
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTD0.446

Quant Time: Apr 23 01:10:46 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE040618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 23 01:07:45 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.38	152	2458	0.40	ng/ul	0.00
2) Naphthalene-d8	11.20	136	7196	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.97	164	3994	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.68	188	9080	0.40	ng/ul	0.00
16) Chrysene-d12	21.77	240	11006	0.40	ng/ul	0.00
20) Perylene-d12	24.38	264	8907	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.76	152	4953	0.37	ng/ul	0.00
14) Fluoranthene-d10	19.66	212	12858	0.34	ng/ul	0.00
Target Compounds					Ovalue	
3) Naphthalene	11.25	128	7384	0.367	ng/ul#	89
5) 2-Methylnaphthalene	12.83	142	5112	0.375	ng/ul	99
7) Acenaphthylene	14.70	152	7735	0.374	ng/ul	96
8) Acenaphthene	15.03	153	5485	0.369	ng/ul	96
9) Fluorene	15.99	166	6088	0.363	ng/ul#	90
11) Pentachlorophenol	17.34	266	527	0.274	ng/ul	95
12) Phenanthrene	17.72	178	9719	0.363	ng/ul#	89
13) Anthracene	17.81	178	9808	0.368	ng/ul#	91
15) Fluoranthene	19.68	202	12376	0.340	ng/ul	84
17) Pyrene	20.04	202	12860	0.344	ng/ul#	79
18) Benzo(a)anthracene	21.75	228	15556	0.350	ng/ul#	86
19) Chrysene	21.81	228	10710	0.335	ng/ul	97
21) Benzo(b)fluoranthene	23.57	252	10406	0.361	ng/ul	88
22) Benzo(k)fluoranthene	23.62	252	10542	0.367	ng/ul#	89
23) Benzo(a)pyrene	24.26	252	10064	0.362	ng/ul#	83
24) Indeno(1,2,3-cd)pyrene	27.18	276	11693	0.378	ng/ul#	78
25) Dibenzo(a,h)anthracene	27.19	278	9368	0.376	ng/ul	92
26) Benzo(g,h,i)perylene	28.04	276	9573	0.371	ng/ul	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE042018\
Data File : BE096074.D
Acq On : 21 Apr 2018 11:21
Operator : SJ/JU
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTD0.446

Quant Time: Apr 23 01:10:46 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE040618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 23 01:07:45 2018
Response via : Initial Calibration

