

Data Path : Z:\HPCHEM1\BNA_E\Data\BE042117\
 Data File : BE092911.D
 Acq On : 21 Apr 2017 17:03
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 22 05:52:22 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE042017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 20 17:35:10 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	100	0.00
2	1,4-Dioxane	0.400	0.399	0.3	107	0.00
3	n-Nitrosodimethylamine	0.400	0.426	-6.5	112	0.00
4 S	2-Fluorophenol	0.400	0.382	4.5	103	0.00
5 S	Phenol-d6	0.400	0.392	2.0	108	0.00
6	bis(2-Chloroethyl)ether	0.400	0.407	-1.7	108	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	103	0.02
8 S	Nitrobenzene-d5	0.400	0.387	3.3	106	0.02
9	Naphthalene	0.400	0.379	5.3	103	0.00
10	Hexachlorobutadiene	0.400	0.381	4.8	102	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.376	6.0	104	0.00
12	2-Methylnaphthalene	0.400	0.374	6.5	104	0.02
13 I	Acenaphthene-d10	0.400	0.400	0.0	106	0.00
14 S	2,4,6-Tribromophenol	0.400	0.356	11.0	110	0.01
15 S	2-Fluorobiphenyl	0.400	0.382	4.5	102	0.02
16	Acenaphthylene	0.400	0.357	10.8	102	0.00
17	Acenaphthene	0.400	0.377	5.8	103	0.00
18	Fluorene	0.400	0.369	7.8	102	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	99	0.00
20	Hexachlorobenzene	0.400	0.428	-7.0	110	0.00
21	Pentachlorophenol	0.400	0.437	-9.2	112	0.00
22	Phenanthrene	0.400	0.423	-5.7	107	0.00
23	Anthracene	0.400	0.428	-7.0	113	0.00
24 SURR	Fluoranthene-d10	0.400	0.415	-3.7	111	0.00
25	Fluoranthene	0.400	0.398	0.5	107	0.00
26 I	Chrysene-d12	0.400	0.400	0.0	104	0.00
27	Pyrene	0.400	0.377	5.8	102	0.00
28 S	Terphenyl-d14	0.400	0.370	7.5	97	0.02
29	Benzo(a)anthracene	0.400	0.360	10.0	102	0.00
30	Chrysene	0.400	0.383	4.3	104	0.00
31	Bis(2-ethylhexyl)phthalate	0.400	0.377	5.8	107	0.00
32	Indeno(1,2,3-cd)pyrene	0.400	0.405	-1.3	114	0.00
33 I	Perylene-d12	0.400	0.400	0.0	130	0.01
34	Benzo(b)fluoranthene	0.400	0.303	24.3	103	0.00
35	Benzo(k)fluoranthene	0.400	0.332	17.0	117	0.01
36 C	Benzo(a)pyrene	0.400	0.313	21.8#	110	0.00
37	Dibenzo(a,h)anthracene	0.400	0.324	19.0	115	0.00
38	Benzo(g,h,i)perylene	0.400	0.328	18.0	112	0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 1