

Data Path : Z:\HPCHEM1\BNA E\DATA\BE042318\
 Data File : BE096081.D
 Acq On : 23 Apr 2018 12:48
 Operator : SJ/JU
 Sample : J2355-05
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 A3Y33

Manual Integrations
 APPROVED

Sohil
 4/24/2018 7:09:48 PM

Quant Time: Apr 24 06:17:27 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE040618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 24 05:21:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.38	152	1852	0.40	ng/ul	0.00
2) Naphthalene-d8	11.20	136	5346	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.97	164	3039	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.68	188	6545m	0.40	ng/ul	0.00
16) Chrysene-d12	21.77	240	6185m	0.40	ng/ul	0.00
20) Perylene-d12	24.38	264	7794	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.76	152	2651	0.27	ng/ul	0.00
14) Fluoranthene-d10	19.66	212	6579m	0.24	ng/ul	0.00
Target Compounds					Ovalue	
3) Naphthalene	11.25	128	1702	0.114	ng/ul#	89
5) 2-Methylnaphthalene	12.83	142	424	0.042	ng/ul	99
7) Acenaphthylene	14.70	152	2452	0.156	ng/ul	95
8) Acenaphthene	15.03	153	859	0.076	ng/ul#	94
9) Fluorene	15.99	166	1153	0.090	ng/ul	91
12) Phenanthrene	17.72	178	22840m	1.184	ng/ul	
13) Anthracene	17.80	178	6321m	0.329	ng/ul	
15) Fluoranthene	19.68	202	59042m	2.249	ng/ul	
17) Pyrene	20.04	202	58409	2.779	ng/ul#	79
18) Benzo(a)anthracene	21.74	228	32359m	1.295	ng/ul	
19) Chrysene	21.81	228	25915	1.441	ng/ul	96
21) Benzo(b)fluoranthene	23.57	252	34876m	1.383	ng/ul	
22) Benzo(k)fluoranthene	23.61	252	10092m	0.402	ng/ul	
23) Benzo(a)pyrene	24.26	252	26316	1.081	ng/ul#	80
24) Indeno(1,2,3-cd)pyrene	27.17	276	16879	0.624	ng/ul#	74
25) Dibenzo(a,h)anthracene	27.17	278	4297	0.197	ng/ul	95
26) Benzo(a,h,i)perylene	28.04	276	13056	0.578	ng/ul	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE042318\
Data File : BE096081.D
Acq On : 23 Apr 2018 12:48
Operator : SJ/JU
Sample : J2355-05
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
A3Y33

Manual Integrations
APPROVED

Sohil
4/24/2018 7:09:48 PM

Quant Time: Apr 24 06:17:27 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE040618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 24 05:21:12 2018
Response via : Initial Calibration

