

Data Path : Z:\HPCHEM1\BNA E\DATA\BE042318\
 Data File : BE096086.D
 Acq On : 23 Apr 2018 15:47
 Operator : SJ/JU
 Sample : J2355-04
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 A3Y32

Manual Integrations
 APPROVED

Sohil
 4/24/2018 7:09:57 PM

Quant Time: Apr 24 06:44:12 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE040618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 24 05:21:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.38	152	1950	0.40	ng/ul	0.00
2) Naphthalene-d8	11.20	136	5282	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.97	164	3136	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.68	188	6818m	0.40	ng/ul	0.00
16) Chrysene-d12	21.77	240	6442m	0.40	ng/ul	0.00
20) Perylene-d12	24.38	264	7456	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.76	152	2327	0.24	ng/ul	0.00
14) Fluoranthene-d10	19.66	212	6108m	0.22	ng/ul	0.00
Target Compounds						
8) Acenaphthene	15.02	153	248	0.021	ng/ul#	86
12) Phenanthrene	17.71	178	1235m	0.061	ng/ul	
15) Fluoranthene	19.68	202	3225m	0.118	ng/ul	
17) Pyrene	20.04	202	3802	0.174	ng/ul#	86
18) Benzo(a)anthracene	21.75	228	1833m	0.070	ng/ul	
19) Chrysene	21.80	228	1526m	0.081	ng/ul	
21) Benzo(b)fluoranthene	23.57	252	1896m	0.079	ng/ul	
22) Benzo(k)fluoranthene	23.62	252	719m	0.030	ng/ul	
23) Benzo(a)pyrene	24.26	252	1210	0.052	ng/ul#	78
24) Indeno(1,2,3-cd)pyrene	27.17	276	893	0.035	ng/ul#	78
26) Benzo(g,h,i)perylene	28.04	276	676	0.031	ng/ul#	58

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE042318\
Data File : BE096086.D
Acq On : 23 Apr 2018 15:47
Operator : SJ/JU
Sample : J2355-04
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
A3Y32

Manual Integrations
APPROVED

Sohil
4/24/2018 7:09:57 PM

Quant Time: Apr 24 06:44:12 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE040618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 24 05:21:12 2018
Response via : Initial Calibration

