

Data Path : Z:\HPCHEM1\BNA E\DATA\BE042318\
 Data File : BE096103.D
 Acq On : 24 Apr 2018 2:01
 Operator : SJ/JU
 Sample : J2434-16
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 C0AD1

Manual Integrations
 APPROVED

Sohil
 4/24/2018 7:11:04 PM

Quant Time: Apr 24 16:36:26 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE040618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 24 06:59:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.38	152	2096	0.40	ng/ul	0.00
2) Naphthalene-d8	11.20	136	5664	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.97	164	3262	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.68	188	7344m	0.40	ng/ul	0.00
16) Chrysene-d12	21.77	240	6982m	0.40	ng/ul	0.00
20) Perylene-d12	24.40	264	6792m	0.40	ng/ul	0.02
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.76	152	1780	0.17	ng/ul	0.00
14) Fluoranthene-d10	19.66	212	4491m	0.15	ng/ul	0.00
Target Compounds						
					Ovalue	
3) Naphthalene	11.25	128	4075	0.258	ng/ul#	91
5) 2-Methylnaphthalene	12.83	142	2629	0.245	ng/ul	100
7) Acenaphthylene	14.70	152	1558	0.092	ng/ul#	88
8) Acenaphthene	15.03	153	8566	0.706	ng/ul	95
9) Fluorene	15.99	166	4315	0.315	ng/ul	90
12) Phenanthrene	17.72	178	92706m	4.283	ng/ul	
13) Anthracene	17.81	178	21242m	0.986	ng/ul	
15) Fluoranthene	19.69	202	516006m	17.520	ng/ul	
17) Pyrene	20.04	202	603958	25.457	ng/ul#	77
18) Benzo(a)anthracene	21.75	228	585960m	20.773	ng/ul	
19) Chrysene	21.82	228	505091	24.880	ng/ul	97
21) Benzo(b)fluoranthene	23.59	252	949069m	43.183	ng/ul	
22) Benzo(k)fluoranthene	23.63	252	296250m	13.531	ng/ul	
23) Benzo(a)pyrene	24.28	252	716702m	33.794	ng/ul	
24) Indeno(1,2,3-cd)pyrene	27.21	276	520742m	22.094	ng/ul	
25) Dibenzo(a,h)anthracene	27.21	278	130758m	6.889	ng/ul	
26) Benzo(a,h,i)perylene	28.09	276	528262m	26.833	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE042318\
Data File : BE096103.D
Acq On : 24 Apr 2018 2:01
Operator : SJ/JU
Sample : J2434-16
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
C0AD1

Manual Integrations
APPROVED

Sohil
4/24/2018 7:11:04 PM

Quant Time: Apr 24 16:36:26 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE040618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 24 06:59:54 2018
Response via : Initial Calibration

