

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE042418\
 Data File : BE096129.D
 Acq On : 24 Apr 2018 20:54
 Operator : SJ/JU
 Sample : PB108470BL
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SBLK70

Manual Integrations
 APPROVED

Sohil
 4/25/2018 4:28:27 PM

Quant Time: Apr 25 06:00:41 2018
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE042418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 25 05:27:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.36	152	867	0.40	ng/ul	0.00
2) Naphthalene-d8	11.18	136	3371	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.95	164	2192	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.65	188	5973m	0.40	ng/ul	0.00
16) Chrysene-d12	21.75	240	8042m	0.40	ng/ul	0.00
20) Perylene-d12	24.35	264	6319m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.74	152	1892	0.30	ng/ul	0.00
14) Fluoranthene-d10	19.64	212	7945m	0.29	ng/ul	0.00
Target Compounds						
9) Fluorene	15.98	166	600	0.056	ng/ul	Qvalue 91
12) Phenanthrene	17.70	178	4357m	0.248	ng/ul	
13) Anthracene	17.79	178	802m	0.045	ng/ul	
15) Fluoranthene	19.67	202	3286m	0.122	ng/ul	
17) Pyrene	20.02	202	2440	0.151	ng/ul#	87
18) Benzo(a)anthracene	21.73	228	727m	0.030	ng/ul	
19) Chrysene	21.79	228	673m	0.030	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE042418\
Data File : BE096129.D
Acq On : 24 Apr 2018 20:54
Operator : SJ/JU
Sample : PB108470BL
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SBLK70

Manual Integrations
APPROVED

Sohil
4/25/2018 4:28:27 PM

Quant Time: Apr 25 06:00:41 2018
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE042418.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 25 05:27:03 2018
Response via : Initial Calibration

