

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE042418\
 Data File : BE096122.D
 Acq On : 24 Apr 2018 16:43
 Operator : SJ/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Quant Time: Apr 24 17:53:32 2018
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE042418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 24 17:40:03 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
3	Naphthalene	1.114	1.096	1.6	100	0.00
4 SURR	2-Methylnaphthalene-d10	0.739	0.714	3.4	98	0.00
5	2-Methylnaphthalene	0.746	0.722	3.2	98	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	98	0.00
7	Acenaphthylene	2.062	2.024	1.8	97	0.00
8 C	Acenaphthene	1.487	1.446	2.8	95	0.00
9	Fluorene	1.967	1.667	15.3	96	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
11	Pentachlorophenol	0.096	0.081	15.6	86	0.00
12	Phenanthrene	1.178	1.136	3.6	94	0.00
13	Anthracene	1.182	1.129	4.5	93	0.00
14 SURR	Fluoranthene-d10	1.845	1.759	4.7	93	0.00
15 C	Fluoranthene	1.800	1.734	3.7	94	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	94	0.00
17	Pyrene	0.802	0.779	2.9	92	0.00
18	Benzo(a)anthracene	1.211	1.193	1.5	99	0.00
19	Chrysene	1.113	1.083	2.7	97	0.00
20 I	Perylene-d12	1.000	1.000	0.0	98	0.00
21	Benzo(b)fluoranthene	1.515	1.353	10.7	91	0.00
22	Benzo(k)fluoranthene	1.448	1.332	8.0	95	0.00
23 C	Benzo(a)pyrene	1.316	1.225	6.9	94	0.00
24	Indeno(1,2,3-cd)pyrene	1.773	1.491	15.9	86	0.00
25	Dibenzo(a,h)anthracene	1.485	1.224	17.6	84	0.00
26	Benzo(g,h,i)perylene	1.543	1.266	18.0	85	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0