

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE042418\
 Data File : BE096123.D
 Acq On : 24 Apr 2018 17:19
 Operator : SJ/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Quant Time: Apr 25 05:24:31 2018
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE042418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 24 17:40:03 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
3	Naphthalene	1.114	1.097	1.5	100	0.00
4 SURR	2-Methylnaphthalene-d10	0.739	0.719	2.7	98	0.00
5	2-Methylnaphthalene	0.746	0.748	-0.3	102	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
7	Acenaphthylene	2.062	2.043	0.9	101	0.00
8 C	Acenaphthene	1.487	1.496	-0.6	101	0.00
9	Fluorene	1.967	2.041	-3.8	121	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	104	0.00
11	Pentachlorophenol	0.096	0.078	18.8	88	0.00
12	Phenanthrene	1.178	1.150	2.4	102	0.00
13	Anthracene	1.182	1.135	4.0	101	0.00
14 SURR	Fluoranthene-d10	1.845	1.783	3.4	101	0.00
15 C	Fluoranthene	1.800	1.763	2.1	103	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	103	0.00
17	Pyrene	0.802	0.752	6.2	97	0.00
18	Benzo(a)anthracene	1.211	1.162	4.0	105	0.00
19	Chrysene	1.113	1.043	6.3	102	0.00
20 I	Perylene-d12	1.000	1.000	0.0	102	0.00
21	Benzo(b)fluoranthene	1.515	1.358	10.4	95	0.00
22	Benzo(k)fluoranthene	1.448	1.306	9.8	97	0.00
23 C	Benzo(a)pyrene	1.316	1.232	6.4	98	0.00
24	Indeno(1,2,3-cd)pyrene	1.773	1.497	15.6	89	0.00
25	Dibenzo(a,h)anthracene	1.485	1.232	17.0	88	0.00
26	Benzo(g,h,i)perylene	1.543	1.255	18.7	87	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0