

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE042418\
 Data File : BE096139.D
 Acq On : 25 Apr 2018 2:55
 Operator : SJ/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Quant Time: Apr 25 05:30:14 2018
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE042418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 25 05:27:03 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	21	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	16#	0.00
3	Naphthalene	1.114	1.154	-3.6	17#	0.00
4 SURR	2-Methylnaphthalene-d10	0.739	0.705	4.6	15#	0.00
5	2-Methylnaphthalene	0.746	0.708	5.1	15#	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	12#	0.00
7	Acenaphthylene	2.062	2.073	-0.5	12#	0.00
8 C	Acenaphthene	1.487	1.570	-5.6	13#	0.00
9	Fluorene	1.967	1.886	4.1	14#	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	10#	0.00
11	Pentachlorophenol	0.096	0.089	7.3	10#	0.00
12	Phenanthrene	1.178	1.163	1.3	10#	0.00
13	Anthracene	1.182	1.171	0.9	10#	0.00
14 SURR	Fluoranthene-d10	1.845	1.337	27.5#	7#	0.00
15 C	Fluoranthene	1.800	1.353	24.8#	8#	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	7#	0.00
17	Pyrene	0.802	0.766	4.5	7#	0.00
18	Benzo(a)anthracene	1.211	1.077	11.1	7#	0.00
19	Chrysene	1.113	1.190	-6.9	8#	0.00
20 I	Perylene-d12	1.000	1.000	0.0	9#	0.00
21	Benzo(b)fluoranthene	1.515	1.537	-1.5	10#	0.00
22	Benzo(k)fluoranthene	1.448	1.411	2.6	9#	0.00
23 C	Benzo(a)pyrene	1.316	1.180	10.3	8#	0.00
24	Indeno(1,2,3-cd)pyrene	1.773	1.627	8.2	9#	0.00
25	Dibenzo(a,h)anthracene	1.485	1.343	9.6	9#	0.00
26	Benzo(g,h,i)perylene	1.543	1.389	10.0	9#	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1