

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE042519\
 Data File : BE099353.D
 Acq On : 25 Apr 2019 10:10
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Apr 26 01:07:48 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE041819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 19 06:51:38 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
2	1,4-Dioxane	0.560	0.569	-1.6	92	0.00
3	n-Nitrosodimethylamine	0.301	0.293	2.7	103	0.00
4 S	2-Fluorophenol	0.977	1.003	-2.7	92	0.00
5 S	Phenol-d6	1.395	1.398	-0.2	90	-0.01
6	bis(2-Chloroethyl)ether	1.301	1.325	-1.8	88	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	93	0.00
8 S	Nitrobenzene-d5	0.311	0.277	10.9	88	0.00
9	Naphthalene	4.364	5.655	-29.6#	120	0.00
10	Hexachlorobutadiene	0.277	0.276	0.4	91	-0.01
11 SURR	2-Methylnaphthalene-d10	0.706	0.694	1.7	92	0.00
12	2-Methylnaphthalene	0.743	0.905	-21.8	114	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	93	0.00
14 S	2,4,6-Tribromophenol	0.133	0.128	3.8	89	0.00
15 S	2-Fluorobiphenyl	1.572	1.631	-3.8	94	0.00
16	Acenaphthylene	1.808	1.797	0.6	95	0.00
17	Acenaphthene	1.114	1.146	-2.9	95	0.00
18	Fluorene	1.601	1.698	-6.1	97	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
20	4-Bromophenyl-phenylether	0.222	0.216	2.7	93	0.00
21	Hexachlorobenzene	0.213	0.226	-6.1	99	0.00
22	Pentachlorophenol	0.057	0.051	10.5	104	0.00
23	Phenanthrene	1.059	1.157	-9.3	104	0.00
24	Anthracene	0.942	0.942	0.0	97	0.00
25 SURR	Fluoranthene-d10	4.806	4.997	-4.0	102	0.00
26	Fluoranthene	1.418	1.499	-5.7	103	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	101	0.00
28	Pvrene	1.178	1.229	-4.3	103	0.00
29 S	Terphenyl-d14	0.787	0.812	-3.2	103	0.00
30	Benzo(a)anthracene	1.176	1.198	-1.9	108	0.00
31	Chrysene	1.251	1.279	-2.2	100	0.00
32	Bis(2-ethylhexyl)phthalate	1.070	0.920	14.0	102	0.00
33	Indeno(1,2,3-cd)pyrene	1.236	1.338	-8.3	110	0.00
34 I	Perylene-d12	1.000	1.000	0.0	107	0.00
35	Benzo(b)fluoranthene	1.225	1.227	-0.2	108	0.00
36	Benzo(k)fluoranthene	1.221	1.236	-1.2	107	0.00
37 C	Benzo(a)pyrene	1.097	1.100	-0.3	108	0.00
38	Dibenzo(a,h)anthracene	1.121	1.147	-2.3	110	0.00
39	Benzo(a,h,i)perylene	1.137	1.130	0.6	106	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE042519\
Data File : BE099353.D
Acq On : 25 Apr 2019 10:10
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 26 01:07:48 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE041819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 19 06:51:38 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			