

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE042519\
 Data File : BE099368.D
 Acq On : 25 Apr 2019 19:13
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 26 02:28:34 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE041819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 19 06:51:38 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	124	0.00
2	1,4-Dioxane	0.560	0.430	23.2	95	0.00
3	n-Nitrosodimethylamine	0.301	0.304	-1.0	147	-0.01
4 S	2-Fluorophenol	0.977	1.156	-18.3	146	0.01
5 S	Phenol-d6	1.395	1.810	-29.7#	161#	0.00
6	bis(2-Chloroethyl)ether	1.301	1.302	-0.1	119	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	128	0.00
8 S	Nitrobenzene-d5	0.311	0.372	-19.6	163#	0.00
9	Naphthalene	4.364	5.681	-30.2#	166#	0.00
10	Hexachlorobutadiene	0.277	0.304	-9.7	139	0.00
11 SURR	2-Methylnaphthalene-d10	0.706	0.728	-3.1	133	0.00
12	2-Methylnaphthalene	0.743	0.931	-25.3#	162#	0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	130	0.00
14 S	2,4,6-Tribromophenol	0.133	0.215	-61.7#	208#	0.00
15 S	2-Fluorobiphenyl	1.572	1.600	-1.8	127	0.00
16	Acenaphthylene	1.808	2.056	-13.7	150#	0.00
17	Acenaphthene	1.114	1.154	-3.6	132	0.01
18	Fluorene	1.601	1.634	-2.1	129	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	116	0.00
20	4-Bromophenyl-phenylether	0.222	0.246	-10.8	128	0.00
21	Hexachlorobenzene	0.213	0.232	-8.9	122	0.00
22	Pentachlorophenol	0.057	0.109	-91.2#	267#	0.00
23	Phenanthrene	1.059	1.125	-6.2	121	0.00
24	Anthracene	0.942	1.025	-8.8	126	0.00
25 SURR	Fluoranthene-d10	4.806	5.054	-5.2	123	0.00
26	Fluoranthene	1.418	1.442	-1.7	119	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	119	0.00
28	Pvrene	1.178	1.241	-5.3	123	0.00
29 S	Terphenyl-d14	0.787	0.826	-5.0	123	0.00
30	Benzo(a)anthracene	1.176	1.328	-12.9	141	0.00
31	Chrysene	1.251	1.290	-3.1	119	0.00
32	Bis(2-ethylhexyl)phthalate	1.070	2.307	-115.6#	301#	0.00
33	Indeno(1,2,3-cd)pyrene	1.236	1.420	-14.9	137	0.02
34 I	Perylene-d12	1.000	1.000	0.0	133	0.00
35	Benzo(b)fluoranthene	1.225	1.202	1.9	132	0.00
36	Benzo(k)fluoranthene	1.221	1.158	5.2	124	0.01
37 C	Benzo(a)pyrene	1.097	1.134	-3.4	138	0.00
38	Dibenzo(a,h)anthracene	1.121	1.135	-1.2	135	0.02
39	Benzo(a,h,i)perylene	1.137	1.110	2.4	130	0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE042519\
Data File : BE099368.D
Acq On : 25 Apr 2019 19:13
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 26 02:28:34 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE041819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 19 06:51:38 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			