

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE042519\
 Data File : BE099368.D
 Acq On : 25 Apr 2019 19:13
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 26 02:28:34 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE041819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 19 06:51:38 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	124	0.00
2	1,4-Dioxane	0.400	0.307	23.3	95	0.00
3	n-Nitrosodimethylamine	0.400	0.404	-1.0	147	-0.01
4 S	2-Fluorophenol	0.400	0.473	-18.2	146	0.01
5 S	Phenol-d6	0.400	0.519	-29.8#	161	0.00
6	bis(2-Chloroethyl)ether	0.400	0.400	0.0	119	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	128	0.00
8 S	Nitrobenzene-d5	0.400	0.480	-20.0	163	0.00
9	Naphthalene	0.400	0.521	-30.3#	166	0.00
10	Hexachlorobutadiene	0.400	0.439	-9.7	139	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.413	-3.2	133	0.00
12	2-Methylnaphthalene	0.400	0.501	-25.2#	162	0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	130	0.00
14 S	2,4,6-Tribromophenol	0.400	0.648	-62.0#	208	0.00
15 S	2-Fluorobiphenyl	0.400	0.407	-1.7	127	0.00
16	Acenaphthylene	0.400	0.455	-13.7	150	0.00
17	Acenaphthene	0.400	0.414	-3.5	132	0.01
18	Fluorene	0.400	0.408	-2.0	129	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	116	0.00
20	4-Bromophenyl-phenylether	0.400	0.444	-11.0	128	0.00
21	Hexachlorobenzene	0.400	0.436	-9.0	122	0.00
22	Pentachlorophenol	0.400	0.726	-81.5#	267	0.00
23	Phenanthrene	0.400	0.425	-6.2	121	0.00
24	Anthracene	0.400	0.435	-8.7	126	0.00
25 SURR	Fluoranthene-d10	0.400	0.421	-5.2	123	0.00
26	Fluoranthene	0.400	0.407	-1.7	119	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	119	0.00
28	Pvrene	0.400	0.421	-5.2	123	0.00
29 S	Terphenyl-d14	0.400	0.420	-5.0	123	0.00
30	Benzo(a)anthracene	0.400	0.452	-13.0	141	0.00
31	Chrysene	0.400	0.412	-3.0	119	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.863	-115.7#	301	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.460	-15.0	137	0.02
34 I	Perylene-d12	0.400	0.400	0.0	133	0.00
35	Benzo(b)fluoranthene	0.400	0.393	1.8	132	0.00
36	Benzo(k)fluoranthene	0.400	0.379	5.3	124	0.01
37 C	Benzo(a)pyrene	0.400	0.413	-3.2	138	0.00
38	Dibenzo(a,h)anthracene	0.400	0.405	-1.3	135	0.02
39	Benzo(a,h,i)perylene	0.400	0.391	2.3	130	0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE042519\
Data File : BE099368.D
Acq On : 25 Apr 2019 19:13
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 26 02:28:34 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE041819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 19 06:51:38 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				