


Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE042519\
Data File : BE099352.D
Acq On : 25 Apr 2019 9:34
Operator : JU/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE041819.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Fri Apr 19 06:51:38 2019



AutoFind: Scans 2558, 2559, 2560; Background Corrected with Scan 2548

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	27.0	15499	PASS
68	69	0.00	2	2.0	439	PASS
69	198	0.00	100	38.3	21954	PASS
70	69	0.00	2	0.5	110	PASS
127	198	10	80	36.4	20870	PASS
197	198	0.00	2	0.1	76	PASS
198	198	100	100	100.0	57347	PASS
199	198	5	9	6.9	3946	PASS
275	198	10	60	26.5	15201	PASS
365	198	1	100	3.3	1864	PASS
441	198	0.01	100	10.7	6141	PASS
442	198	50	100	65.8	37711	PASS
443	442	15	24	19.7	7427	PASS