

Data Path : Z:\HPCHEM1\BNA E\DATA\BE042616\
 Data File : BE091705.D
 Acq On : 26 Apr 2016 15:38
 Operator : UM/SJ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE042616

Quant Time: Apr 26 16:11:55 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE042616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 26 16:09:46 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
2	1,4-Dioxane	0.752	0.658	12.5	90	0.00
3	n-Nitrosodimethylamine	0.908	0.849	6.5	99	0.02
4 S	2-Fluorophenol	1.228	1.149	6.4	100	0.00
5 S	Phenol-d6	1.629	1.494	8.3	101	0.00
6	bis(2-Chloroethyl)ether	1.433	1.426	0.5	102	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	106	0.00
8 S	Nitrobenzene-d5	0.323	0.297	8.0	104	0.00
9	Nitrobenzene	0.338	0.305	9.8	104	0.00
10	Naphthalene	1.010	1.012	-0.2	104	0.00
11	Hexachlorobutadiene	0.149	0.148	0.7	107	0.00
12	2-Methylnaphthalene	0.649	0.641	1.2	104	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	107	0.00
14 S	2,4,6-Tribromophenol	0.157	0.137	12.7	103	0.00
15 S	2-Fluorobiphenyl	1.414	1.387	1.9	106	0.00
16	Acenaphthylene	1.923	1.925	-0.1	122	0.00
17	Acenaphthene	1.247	1.220	2.2	105	0.00
18	Fluorene	1.550	1.541	0.6	106	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	107	0.00
20	4-Bromophenyl-phenylether	0.180	0.173	3.9	101	0.00
21	Hexachlorobenzene	0.180	0.178	1.1	106	0.00
22	Pentachlorophenol	0.085	0.069	18.8	103	0.00
23	Phenanthrene	1.136	1.150	-1.2	106	0.00
24	Anthracene	1.025	0.994	3.0	105	0.00
25	Fluoranthene	1.188	1.197	-0.8	107	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	104	0.00
27	Benzydine	0.286	0.262	8.4	118	0.00
28	Pvrene	1.130	1.104	2.3	105	0.00
29 S	Terphenyl-d14	0.780	0.748	4.1	102	0.00
30	Benzo(a)anthracene	1.221	1.134	7.1	99	0.00
31	3,3'-Dichlorobenzidine	0.623	0.543	12.8	104	-0.02
32	Chrysene	1.269	1.374	-8.3	106	0.00
33	Bis(2-ethylhexyl)phthalate	0.370	0.299	19.2	109	-0.02
34	Indeno(1,2,3-cd)pyrene	1.130	1.131	-0.1	102	0.00
35 I	Perylene-d12	1.000	1.000	0.0	104	0.00
36	Benzo(b)fluoranthene	1.172	1.140	2.7	104	0.00
37	Benzo(k)fluoranthene	1.191	1.080	9.3	98	0.00
38 C	Benzo(a)pyrene	1.103	1.039	5.8	103	0.00
39	Dibenzo(a,h)anthracene	1.045	0.989	5.4	101	-0.01
40	Benzo(a,h,i)perylene	1.062	0.986	7.2	100	-0.01

Data Path : Z:\HPCHEM1\BNA E\DATA\BE042616\
Data File : BE091705.D
Acq On : 26 Apr 2016 15:38
Operator : UM/SJ
Sample : SSTDICV0.4
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE042616

Quant Time: Apr 26 16:11:55 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE042616.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 26 16:09:46 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0