

Data Path : Z:\HPCHEM1\BNA E\DATA\BE042616\
 Data File : BE091705.D
 Acq On : 26 Apr 2016 15:38
 Operator : UM/SJ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE042616

Quant Time: Apr 26 16:11:55 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE042616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 26 16:09:46 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	104	0.00
2	1,4-Dioxane	0.400	0.358	10.5	90	0.00
3	n-Nitrosodimethylamine	0.400	0.374	6.5	99	0.02
4 S	2-Fluorophenol	0.400	0.374	6.5	100	0.00
5 S	Phenol-d6	0.400	0.367	8.3	101	0.00
6	bis(2-Chloroethyl)ether	0.400	0.398	0.5	102	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	106	0.00
8 S	Nitrobenzene-d5	0.400	0.369	7.8	104	0.00
9	Nitrobenzene	0.400	0.362	9.5	104	0.00
10	Naphthalene	0.400	0.401	-0.3	104	0.00
11	Hexachlorobutadiene	0.400	0.398	0.5	107	0.00
12	2-Methylnaphthalene	0.400	0.395	1.3	104	0.00
13 I	Acenaphthene-d10	5.000	5.000	0.0	107	0.00
14 S	2,4,6-Tribromophenol	0.400	0.450	-12.5	103	0.00
15 S	2-Fluorobiphenyl	0.400	0.392	2.0	106	0.00
16	Acenaphthylene	0.400	0.400	0.0	122	0.00
17	Acenaphthene	0.400	0.391	2.3	105	0.00
18	Fluorene	0.400	0.398	0.5	106	0.00
19 I	Phenanthrene-d10	5.000	5.000	0.0	107	0.00
20	4-Bromophenyl-phenylether	0.400	0.384	4.0	101	0.00
21	Hexachlorobenzene	0.400	0.395	1.3	106	0.00
22	Pentachlorophenol	0.400	0.377	5.8	103	0.00
23	Phenanthrene	0.400	0.405	-1.3	106	0.00
24	Anthracene	0.400	0.388	3.0	105	0.00
25	Fluoranthene	0.400	0.403	-0.8	107	0.00
26 I	Chrysene-d12	5.000	5.000	0.0	104	0.00
27	Benzydine	0.400	0.419	-4.7	118	0.00
28	Pvrene	0.400	0.391	2.3	105	0.00
29 S	Terphenyl-d14	0.400	0.384	4.0	102	0.00
30	Benzo(a)anthracene	0.400	0.371	7.3	99	0.00
31	3,3'-Dichlorobenzidine	0.400	0.349	12.8	104	-0.02
32	Chrysene	0.400	0.433	-8.2	106	0.00
33	Bis(2-ethylhexyl)phthalate	0.400	0.323	19.3	109	-0.02
34	Indeno(1,2,3-cd)pyrene	0.400	0.400	0.0	102	0.00
35 I	Perylene-d12	5.000	5.000	0.0	104	0.00
36	Benzo(b)fluoranthene	0.400	0.389	2.8	104	0.00
37	Benzo(k)fluoranthene	0.400	0.363	9.3	98	0.00
38 C	Benzo(a)pyrene	0.400	0.377	5.8	103	0.00
39	Dibenzo(a,h)anthracene	0.400	0.378	5.5	101	-0.01
40	Benzo(a,h,i)perylene	0.400	0.371	7.3	100	-0.01

Data Path : Z:\HPCHEM1\BNA E\DATA\BE042616\
Data File : BE091705.D
Acq On : 26 Apr 2016 15:38
Operator : UM/SJ
Sample : SSTDICV0.4
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE042616

Quant Time: Apr 26 16:11:55 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE042616.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 26 16:09:46 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)
----------	--------	-------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0