

Data Path : Z:\HPCHEM1\BNA E\DATA\BE042616\
 Data File : BE091703.D
 Acq On : 26 Apr 2016 14:23
 Operator : UM/SJ
 Sample : SSTDICC10
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC10

Manual Integrations
 APPROVED

sohil
 4/27/2016 7:03:42 PM

Quant Time: Apr 26 16:10:40 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE042616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 26 13:13:58 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.80	152	39066	5.00	ng	0.00
7) Naphthalene-d8	10.57	136	188172	5.00	ng	0.00
13) Acenaphthene-d10	14.42	164	113602	5.00	ng	0.00
19) Phenanthrene-d10	17.15	188	265030	5.00	ng	0.00
26) Chrysene-d12	21.31	240	340059	5.00	ng	0.00
35) Perylene-d12	23.75	264	295152	5.00	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.39	112	98579	11.41	ng	0.00
5) Phenol-d6	6.96	99	139840	11.96	ng	0.00
8) Nitrobenzene-d5	8.95	82	133893	13.34	ng	0.00
14) 2,4,6-Tribromophenol	15.90	330	42947m	16.17	ng	0.00
15) 2-Fluorobiphenyl	13.03	172	302031	9.61	ng	0.00
29) Terphenyl-d14	19.76	244	492913	9.14	ng	0.00

Target Compounds						Qvalue
3) n-Nitrosodimethylamine	3.63	42	72986	13.01	ng	# 89
6) bis(2-Chloroethyl)ether	7.21	93	113872	10.85	ng	# 76
9) Nitrobenzene	8.99	77	143789	14.31	ng	# 92
10) Naphthalene	10.61	128	367727	9.59	ng	96
11) Hexachlorobutadiene	10.92	225	53075m	9.68	ng	
12) 2-Methylnaphthalene	12.23	142	247649	10.07	ng	93
16) Acenaphthylene	14.12	152	491991	10.83	ng	# 87
17) Acenaphthene	14.47	154	285303	10.20	ng	95
18) Fluorene	15.46	166	346809	10.00	ng	97
20) 4-Bromophenyl-phenylether	16.33	248	98659	10.91	ng	96
21) Hexachlorobenzene	16.45	284	95046	9.78	ng	96
23) Phenanthrene	17.18	178	586083	9.85	ng	99
24) Anthracene	17.27	178	585871	11.03	ng	98
25) Fluoranthene	19.19	202	665929	11.31	ng	98
28) Pyrene	19.55	202	731032	9.30	ng	98
30) Benzo(a)anthracene	21.29	228	784148m	9.49	ng	
31) 3,3'-Dichlorobenzidine	21.22	252	430975	12.42	ng	# 80
32) Chrysene	21.34	228	751659m	8.34	ng	
33) Bis(2-ethylhexyl)phthalate	21.20	149	293103	18.47	ng	# 57
34) Indeno(1,2,3-cd)pyrene	26.31	276	766194	9.51	ng	97
36) Benzo(b)fluoranthene	23.00	252	695857	9.83	ng	98
37) Benzo(k)fluoranthene	23.04	252	716391	10.29	ng	98
38) Benzo(a)pyrene	23.63	252	677993	11.07	ng	96
39) Dibenzo(a,h)anthracene	26.34	278	639956	10.20	ng	# 96
40) Benzo(g,h,i)perylene	27.10	276	642784	9.99	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA E\DATA\BE042616\
Data File : BE091703.D
Acq On : 26 Apr 2016 14:23
Operator : UM/SJ
Sample : SSTDICC10
Misc :
ALS Vial : 10 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
SSTDICC10

Manual Integrations
APPROVED

sohil
4/27/2016 7:03:42 PM

Quant Time: Apr 26 16:10:40 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE042616.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 26 13:13:58 2016
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 7.80 min Scan# 300

Delta R.T. 0.00 min

Lab File: BE091703.D


Acq: 26 Apr 2016 14:23

Instrument :

BNA_E

ClientSampleId :




SSTDICC10



Tot Ion:	152	Resp:	39066
Ion	Ratio	Lower	Upper
152	100		
150	148.9	122.6	183.8
115	55.5	52.8	79.2

Manual Integrations
APPROVED

sohil
4/27/2016 7:03:42 PM



#3

n-Nitrosodimethylamine


Concen: 13.01 ng

RT: 3.63 min Scan# 60



Delta R.T. 0.00 min


Lab File: BE091703.D

Acq: 26 Apr 2016 14:23



Tgt Ion:	42	Resp:	72986
Ion	Ratio	Lower	Upper
42	100		
74	97.8	87.9	131.9
44	5.7	6.6	9.8#





#4

2-Fluorophenol

Concen: 11.41 ng

RT: 5.39 min Scan# 162

Delta R.T. 0.00 min

Lab File: BE091703.D

Acq: 26 Apr 2016 14:23

Instrument :

BNA_E





ClientSampleId :

SSTDICC10

Tot Ion:	112	Resp:	98579
Ion Ratio	Lower	Upper	
112	100		
64	67.6	30.9	46.3#
63	36.4	16.7	25.1#

Manual Integrations
APPROVED

sohil
4/27/2016 7:03:42 PM



#5

Phenol-d6

Concen: 11.96 ng




RT: 6.96 min Scan# 251




Delta R.T. 0.00 min

Lab File: BE091703.D

Acq: 26 Apr 2016 14:23

Tgt Ion:	99	Resp:	139840
Ion Ratio	Lower	Upper	
99	100		
42	26.8	16.2	24.2#
71	36.0	29.0	43.4





#6

bis(2-Chloroethyl)ether

Concen: 10.85 ng

RT: 7.21 min Scan# 266

Delta R.T. 0.00 min

Lab File: BE091703.D

Acq: 26 Apr 2016 14:23

Tot Ion:	93	Resp:	113872
Ion Ratio	Lower	Upper	
93	100		
63	85.7	49.5	74.3#
95	32.5	22.2	33.4

Instrument :





BNA_E

ClientSampleId :

SSTDICC10

Manual Integrations
APPROVED

sohil
4/27/2016 7:03:42 PM



#7

Naphthalene-d8

Concen: 5.00 ng


RT: 10.57 min Scan# 511


Delta R.T. -0.00 min

Lab File: BE091703.D

Acq: 26 Apr 2016 14:23

Tgt Ion:	136	Resp:	188172
Ion Ratio	Lower	Upper	
136	100		
137	10.9	8.9	13.3
54	7.8	8.2	12.4#
68	5.7	6.2	9.4#





#8

Nitrobenzene-d5

Concen: 13.34 ng

RT: 8.95 min Scan# 391

Delta R.T. -0.00 min

Lab File: BE091703.D

Acq: 26 Apr 2016 14:23

Instrument :

BNA_E





ClientSampleId :

SSTDICC10

Tot Ion:	82	Resp:	133893
Ion Ratio	Lower	Upper	
82	100		
128	33.3	25.4	38.0
54	58.0	48.4	72.6

Manual Integrations
APPROVED

sohil
4/27/2016 7:03:42 PM



#9

Nitrobenzene

Concen: 14.31 ng




RT: 8.99 min Scan# 395


Delta R.T. -0.00 min

Lab File: BE091703.D

Acq: 26 Apr 2016 14:23

Tgt Ion:	77	Resp:	143789
Ion Ratio	Lower	Upper	
77	100		
123	38.6	26.9	40.3
65	14.2	9.4	14.2#









#10
Naphthalene
Concen: 9.59 ng
RT: 10.61 min Scan# 513
Delta R.T. -0.00 min
Lab File: BE091703.D
Acq: 26 Apr 2016 14:23

Instrument :
BNA_E
ClientSampleId :
SSTDICC10

Tot Ion	Ratio	Lower	Upper
128	100		
129	9.9	9.3	13.9
127	14.8	10.7	16.1




Manual Integrations
APPROVED


sohil
4/27/2016 7:03:42 PM



#11
Hexachlorobutadiene
Concen: 9.68 ng m
RT: 10.92 min Scan# 529
Delta R.T. -0.00 min
Lab File: BE091703.D
Acq: 26 Apr 2016 14:23

Tgt Ion	Ratio	Lower	Upper
225	100		
223	68.7	49.0	73.6
227	69.1	50.3	75.5





#12

2-Methylnaphthalene

Concen: 10.07 ng

RT: 12.23 min Scan# 637

Delta R.T. -0.01 min

Lab File: BE091703.D

Acq: 26 Apr 2016 14:23

Instrument :

BNA_E





ClientSampleId :

SSTDICC10

Tot Ion	Ratio	Lower	Upper
142	100		
141	96.0	71.4	107.0
115	41.4	30.3	45.5

Manual Integrations
APPROVED

sohil
4/27/2016 7:03:42 PM



#13

Acenaphthene-d10

Concen: 5.00 ng




RT: 14.42 min Scan# 820


Delta R.T. 0.00 min

Lab File: BE091703.D

Acq: 26 Apr 2016 14:23


Tgt Ion	Ratio	Lower	Upper
164	100		
162	92.6	85.3	127.9
160	36.4	46.2	69.2#





#14
 2,4,6-Tribromophenol
 Concen: 16.17 ng/ml
 RT: 15.90 min Scan# 939
 Delta R.T. 0.00 min
 Lab File: BE091703.D
 Acq: 26 Apr 2016 14:23




Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC10




Tot Ion	Ratio	Lower	Upper
330	100		
332	110.6	79.1	118.7
141	44.8	125.4	188.0

Manual Integrations
 APPROVED



sohil
 4/27/2016 7:03:42 PM




#15
 2-Fluorobiphenyl
 Concen: 9.61 ng
 RT: 13.03 min Scan# 707
 Delta R.T. 0.00 min
 Lab File: BE091703.D
 Acq: 26 Apr 2016 14:23



Tgt Ion	Ratio	Lower	Upper
172	100		
171	35.7	28.8	43.2
170	24.2	19.3	28.9









#16
Acenaphthylene
Concen: 10.83 ng
RT: 14.12 min Scan# 798
Delta R.T. 0.00 min
Lab File: BE091703.D
Acq: 26 Apr 2016 14:23

Instrument :
BNA_E
ClientSampleId :
SSTDICC10

Tot Ion	Ratio	Resp	Lower	Upper
152	100	491991		
151	20.0		3.9	5.9#
153	13.3		10.3	15.5




Manual Integrations
APPROVED


sohil
4/27/2016 7:03:42 PM



#17
Acenaphthene
Concen: 10.20 ng
RT: 14.47 min Scan# 824
Delta R.T. 0.00 min
Lab File: BE091703.D
Acq: 26 Apr 2016 14:23

Tgt Ion	Ratio	Resp	Lower	Upper
154	100	285303		
153	105.9		80.0	120.0
152	52.1		40.0	60.0





#18

Fluorene

Concen: 10.00 ng

RT: 15.46 min Scan# 901

Delta R.T. 0.00 min

Lab File: BE091703.D

Acq: 26 Apr 2016 14:23

Instrument :

BNA_E





ClientSampleId :

SSTDICC10

Manual Integrations
APPROVED

sohil
4/27/2016 7:03:42 PM

Tot Ion:	166	Resp:	346809
Ion Ratio	Lower	Upper	
166	100		
165	98.2	80.8	121.2
167	13.5	10.9	16.3



#19

Phenanthrene-d10

Concen: 5.00 ng




RT: 17.15 min Scan# 1015


Delta R.T. -0.00 min

Lab File: BE091703.D

Acq: 26 Apr 2016 14:23

Tgt Ion:	188	Resp:	265030
Ion Ratio	Lower	Upper	
188	100		
94	8.0	8.7	13.1#
80	7.5	9.4	14.0#









#20
4-Bromophenyl-phenylether
Concen: 10.91 ng
RT: 16.33 min Scan# 966
Delta R.T. 0.00 min
Lab File: BE091703.D
Acq: 26 Apr 2016 14:23

Instrument :
BNA_E
ClientSampleId :
SSTDICC10

Tot Ion	Ratio	Lower	Upper
248	100		
250	96.4	80.5	120.7
141	85.5	72.0	108.0




Manual Integrations
APPROVED


sohil
4/27/2016 7:03:42 PM



#21
Hexachlorobenzene
Concen: 9.78 ng
RT: 16.45 min Scan# 973
Delta R.T. 0.00 min
Lab File: BE091703.D
Acq: 26 Apr 2016 14:23

Tgt Ion	Ratio	Lower	Upper
284	100		
142	42.5	32.2	48.2
249	33.8	25.4	38.2





#23

Phenanthrene

Concen: 9.85 ng

RT: 17.18 min Scan# 1018

Delta R.T. -0.00 min

Lab File: BE091703.D

Acq: 26 Apr 2016 14:23

Instrument :

BNA_E





ClientSampleId :

SSTDICC10

Tot Ion:178	Resp:	586083
Ion Ratio	Lower	Upper
178	100	
176	19.8	15.3 22.9
179	15.5	12.6 18.8

Manual Integrations
APPROVED

sohil
4/27/2016 7:03:42 PM



#24

Anthracene

Concen: 11.03 ng




RT: 17.27 min Scan# 1026


Delta R.T. -0.00 min

Lab File: BE091703.D

Acq: 26 Apr 2016 14:23

Tgt Ion:178	Resp:	585871
Ion Ratio	Lower	Upper
178	100	
176	19.2	14.6 21.8
179	15.7	11.8 17.6





#25

Fluoranthene

Concen: 11.31 ng

RT: 19.19 min Scan# 1149

Delta R.T. -0.00 min

Lab File: BE091703.D

Acq: 26 Apr 2016 14:23

Instrument :

BNA_E





ClientSampleId :

SSTDICC10

Tot Ion:	202	Resp:	665929
Ion Ratio	Lower	Upper	
202	100		
101	11.8	11.0	16.6
203	17.3	13.9	20.9

Manual Integrations
APPROVED

sohil
4/27/2016 7:03:42 PM



#26

Chrysene-d12

Concen: 5.00 ng




RT: 21.31 min Scan# 1247


Delta R.T. 0.00 min

Lab File: BE091703.D

Acq: 26 Apr 2016 14:23

Tgt Ion:	240	Resp:	340059
Ion Ratio	Lower	Upper	
240	100		
120	1.2	11.0	16.4#
236	24.5	21.7	32.5









#28
Pvrene
Concen: 9.30 ng
RT: 19.55 min Scan# 1168
Delta R.T. -0.00 min
Lab File: BE091703.D
Acq: 26 Apr 2016 14:23

Instrument :
BNA_E
ClientSampleId :
SSTDICC10

Tot Ion	Ratio	Lower	Upper
202	100		
200	22.3	16.8	25.2
203	17.7	14.0	21.0




Manual Integrations
APPROVED


sohil
4/27/2016 7:03:42 PM



#29
Terphenyl-d14
Concen: 9.14 ng
RT: 19.76 min Scan# 1179
Delta R.T. -0.00 min
Lab File: BE091703.D
Acq: 26 Apr 2016 14:23

Tgt Ion	Ratio	Lower	Upper
244	100		
212	8.1	6.9	10.3
122	2.2	11.0	16.4#









#30
Benzo(a)anthracene
Concen: 9.49 ng/mL
RT: 21.29 min Scan# 1246
Delta R.T. 0.00 min
Lab File: BE091703.D
Acq: 26 Apr 2016 14:23

Instrument :
BNA_E
ClientSampleId :
SSTDICC10

Tot Ion	Ratio	Lower	Upper
228	100		
226	25.6	21.0	31.4
229	21.2	15.7	23.5




Manual Integrations
APPROVED


sohil
4/27/2016 7:03:42 PM



#31
3,3'-Dichlorobenzidine
Concen: 12.42 ng/mL
RT: 21.22 min Scan# 1243
Delta R.T. -0.02 min
Lab File: BE091703.D
Acq: 26 Apr 2016 14:23

Tgt Ion	Ratio	Lower	Upper
252	100		
254	46.7	51.1	76.7#
126	21.0	12.6	18.8#









#32
Chrysene
Concen: 8.34 ng/mL
RT: 21.34 min Scan# 1248
Delta R.T. 0.00 min
Lab File: BE091703.D
Acq: 26 Apr 2016 14:23

Instrument :
BNA_E
ClientSampleId :
SSTDICC10

Tot Ion	Ratio	Lower	Upper
228	100		
226	33.8	24.0	36.0
229	17.7	15.9	23.9




Manual Integrations
APPROVED


sohil
4/27/2016 7:03:42 PM



#33
Bis(2-ethylhexyl)phthalate
Concen: 18.47 ng/mL
RT: 21.20 min Scan# 1242
Delta R.T. -0.02 min
Lab File: BE091703.D
Acq: 26 Apr 2016 14:23

Tgt Ion	Ratio	Lower	Upper
149	100		
167	0.0	19.5	29.3#
279	0.0	10.0	15.0#









#34
Indeno(1,2,3-cd)pyrene
Concen: 9.51 ng
RT: 26.31 min Scan# 1644
Delta R.T. 0.00 min
Lab File: BE091703.D
Acq: 26 Apr 2016 14:23

Instrument :
BNA_E
ClientSampleId :
SSTDICC10

Tot Ion	Ratio	Resp	Lower	Upper
276	100	766194		
138	26.4		23.4	35.2
277	25.1		19.8	29.8




Manual Integrations
APPROVED


sohil
4/27/2016 7:03:42 PM



#35
Perylene-d12
Concen: 5.00 ng
RT: 23.75 min Scan# 1425
Delta R.T. 0.00 min
Lab File: BE091703.D
Acq: 26 Apr 2016 14:23

Tgt Ion	Ratio	Resp	Lower	Upper
264	100	295152		
260	25.8		20.0	30.0
265	20.9		17.5	26.3









#36
Benzo(b)fluoranthene
Concen: 9.83 ng
RT: 23.00 min Scan# 1360
Delta R.T. 0.00 min
Lab File: BE091703.D
Acq: 26 Apr 2016 14:23

Instrument :
BNA_E
ClientSampleId :
SSTDICC10

Tot Ion	Ratio	Resp	Lower	Upper
252	100	695857		
253	21.6		17.4	26.0
125	10.5		10.2	15.2




Manual Integrations
APPROVED


sohil
4/27/2016 7:03:42 PM



#37
Benzo(k)fluoranthene
Concen: 10.29 ng
RT: 23.04 min Scan# 1364
Delta R.T. 0.00 min
Lab File: BE091703.D
Acq: 26 Apr 2016 14:23

Tgt Ion	Ratio	Resp	Lower	Upper
252	100	716391		
253	21.5		17.6	26.4
125	10.7		9.9	14.9









#38
Benzo(a)pyrene
Concen: 11.07 ng
RT: 23.63 min Scan# 1415
Delta R.T. 0.00 min
Lab File: BE091703.D
Acq: 26 Apr 2016 14:23

Instrument :
BNA_E
ClientSampleId :
SSTDICC10

Tot Ion	Ratio	Lower	Upper
252	100		
253	21.2	18.1	27.1
125	11.8	11.0	16.4




Manual Integrations
APPROVED




sohil
4/27/2016 7:03:42 PM



#39
Dibenzo(a,h)anthracene
Concen: 10.20 ng
RT: 26.34 min Scan# 1647
Delta R.T. 0.00 min
Lab File: BE091703.D
Acq: 26 Apr 2016 14:23

Tgt Ion	Ratio	Lower	Upper
278	100		
139	15.7	16.0	24.0#
279	24.3	19.4	29.2





#40

Benzo(a,h,i)perylene

Concen: 9.99 ng

RT: 27.10 min Scan# 1713

Delta R.T. 0.00 min

Lab File: BE091703.D

Acq: 26 Apr 2016 14:23

Instrument :

BNA_E

ClientSampleId :

SSTDICC10

Tot Ion:	276	Resp:	642784
Ion Ratio	Lower	Upper	
276	100		
277	24.0	18.6	28.0
138	22.1	20.4	30.6

Manual Integrations
APPROVED

sohil
4/27/2016 7:03:42 PM

