

Data Path : Z:\HPCHEM1\BNA E\DATA\BE042618\
 Data File : BE096178.D
 Acq On : 26 Apr 2018 13:22
 Operator : SJ/JU
 Sample : J2431-08DL 10X
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 DASN9DL

Manual Integrations
 APPROVED

Sohil
 4/27/2018 7:28:45 PM

Quant Time: Apr 27 03:01:59 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE042418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 27 02:29:47 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.36	152	693	0.40	ng/ul	0.00
2) Naphthalene-d8	11.18	136	2736	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.95	164	1618	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.66	188	3977m	0.40	ng/ul	0.00
16) Chrysene-d12	21.75	240	5576m	0.40	ng/ul	0.00
20) Perylene-d12	24.35	264	3924m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.74	152	169	0.03	ng/ul	0.00
14) Fluoranthene-d10	19.64	212	488m	0.03	ng/ul	0.00
Target Compounds					Ovalue	
3) Naphthalene	11.23	128	1941	0.255	ng/ul#	78
5) 2-Methylnaphthalene	12.81	142	7465	1.463	ng/ul	100
7) Acenaphthylene	14.67	152	3083	0.370	ng/ul#	83
8) Acenaphthene	15.01	153	65147	10.833	ng/ul	93
9) Fluorene	15.98	166	179362	22.541	ng/ul#	92
12) Phenanthrene	17.71	178	1232289m	105.242	ng/ul	
13) Anthracene	17.79	178	193164m	16.443	ng/ul	
15) Fluoranthene	19.67	202	759388m	42.429	ng/ul	
17) Pyrene	20.02	202	377209	33.721	ng/ul#	77
18) Benzo(a)anthracene	21.73	228	132576m	7.853	ng/ul	
19) Chrysene	21.79	228	125979	8.121	ng/ul	97
21) Benzo(b)fluoranthene	23.55	252	88765m	5.974	ng/ul	
22) Benzo(k)fluoranthene	23.59	252	28568m	2.011	ng/ul	
23) Benzo(a)pyrene	24.24	252	57189m	4.430	ng/ul	
24) Indeno(1,2,3-cd)pyrene	27.13	276	19346m	1.112	ng/ul	
25) Dibenzo(a,h)anthracene	27.13	278	5649m	0.388	ng/ul	
26) Benzo(a,h,i)perylene	28.00	276	13845m	0.914	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE042618\
Data File : BE096178.D
Acq On : 26 Apr 2018 13:22
Operator : SJ/JU
Sample : J2431-08DL 10X
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
DASN9DL

Manual Integrations
APPROVED

Sohil
4/27/2018 7:28:45 PM

Quant Time: Apr 27 03:01:59 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE042418.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 27 02:29:47 2018
Response via : Initial Calibration

