

Data File : BE092917.D
Acq On : 27 Apr 2017 16:39
Operator : SJ/MA
Sample : I2837-02
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
KY038PS027-170424

Quant Time: Apr 28 04:05:29 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE042017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 20 17:35:10 2017
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.77	152	943	0.40	ng	0.00
7) Naphthalene-d8	10.53	136	3743	0.40	ng	0.00
13) Acenaphthene-d10	14.39	164	1573	0.40	ng	0.00
19) Phenanthrene-d10	17.13	188	3773	0.40	ng	0.00
26) Chrysene-d12	21.28	240	3013	0.40	ng	0.00
33) Perylene-d12	23.69	264	2461	0.40	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.36	112	145	0.07	ng	0.00
5) Phenol-d6	6.94	99	120	0.04	ng	0.00
8) Nitrobenzene-d5	8.90	82	798	0.32	ng	0.00
11) 2-Methylnaphthalene-d10	12.13	152	2092	0.38	ng	0.00
14) 2,4,6-Tribromophenol	15.87	330	27	0.10	ng	0.00
15) 2-Fluorobiphenyl	13.01	172	3170	0.44	ng	0.00
24) Fluoranthene-d10	19.15	212	3358	0.37	ng	0.00
28) Terphenyl-d14	19.74	244	2490	0.48	ng	0.00

Target Compounds				Qvalue		
2) 1,4-Dioxane	3.32	88	71	0.04	ng	# 1
6) bis(2-Chloroethyl)ether	7.11	93	108	0.03	ng	# 20
18) Fluorene	15.43	166	123	0.02	ng	# 91
20) Hexachlorobenzene	16.46	284	50	0.03	ng	# 100
21) Pentachlorophenol	16.79	266	4	0.15	ng	96
22) Phenanthrene	17.27	178	254	0.02	ng	98
23) Anthracene	17.27	178	283	0.03	ng	99
25) Fluoranthene	19.17	202	1181	0.03	ng	# 98
27) Pyrene	19.53	202	1159	0.03	ng	# 95
29) Benzo(a)anthracene	21.27	228	225	0.03	ng	# 80
30) Chrysene	21.32	228	232	0.03	ng	# 81
31) Bis(2-ethylhexyl)phthalate	21.20	149	179	0.10	ng	# 96
32) Indeno(1,2,3-cd)pyrene	26.22	276	916	0.03	ng	94
34) Benzo(b)fluoranthene	22.95	252	213	0.02	ng	# 62
35) Benzo(k)fluoranthene	22.99	252	198	0.02	ng	# 58
36) Benzo(a)pyrene	23.58	252	175	0.03	ng	# 56
37) Dibenzo(a,h)anthracene	26.25	278	192	0.02	ng	# 65
38) Benzo(g,h,i)perylene	27.00	276	863	0.03	ng	# 72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BE092917.D

Acq On : 27 Apr 2017 16:39

Operator : SJ/MA

Sample : I2837-02

Misc :

ALS Vial : 6 Sample Multiplier: 1

Instrument :

BNA_E

ClientSampleId :

KY038PS027-170424


Quant Time: Apr 28 04:05:29 2017


Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE042017.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Thu Apr 20 17:35:10 2017

Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 0.40 ng

RT: 7.77 min Scan# 555

Delta R.T. 0.00 min

Lab File: BE092917.D

Acq: 27 Apr 2017 16:39

Instrument :

BNA_E

ClientSampleId :

KY038PS027-170424


Tgt Ion: 152 Resp: 943

Ion Ratio Lower Upper

152 100

150 171.8 129.9 194.9




115 86.3 62.6 93.8



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 0.04 ng

RT: 3.32 min Scan# 63

Delta R.T. 0.01 min

Lab File: BE092917.D

Acq: 27 Apr 2017 16:39


Tgt Ion: 88 Resp: 71

Ion Ratio Lower Upper

88 100

58 109.9 64.9 97.3#



43 271.8 26.6 40.0#




Abundance Ion 88.00 (87.70 to 88.70): BE0

Ion 58.00 (57.70 to 58.70): BE0

Ion 43.00 (42.70 to 43.70): BE0





#4

2-Fluorophenol

Concen: 0.07 ng

RT: 5.36 min Scan# 267

Delta R.T. 0.01 min

Lab File: BE092917.D

Acq: 27 Apr 2017 16:39

Instrument :

BNA_E

ClientSampleId :

KY038PS027-170424


Tgt Ion: 112 Resp: 145

Ion Ratio Lower Upper

112 100

64 69.0 52.6 79.0

63 37.2 29.0 43.4



Abundance Ion 112.00 (111.70 to 112.70): BE0

Ion 64.00 (63.70 to 64.70): BE0

Ion 63.00 (62.70 to 63.70): BE0

5.36

150



100

50

0

5.30 5.35 5.40 5.45

Time-->



#5

Phenol-d6

Concen: 0.04 ng

RT: 6.94 min Scan# 519

Delta R.T. 0.00 min

Lab File: BE092917.D

Acq: 27 Apr 2017 16:39


Tgt Ion: 99 Resp: 120

Ion Ratio Lower Upper

99 100

42 50.8 0.0 0.0#

71 0.0 0.0 0.0



Abundance Ion 99.00 (98.70 to 99.70): BE0

Ion 42.00 (41.70 to 42.70): BE0

Ion 71.00 (70.70 to 71.70): BE0

6.94


100


50

0

6.85 6.90 6.95 7.00

Time-->





#6

bis(2-Chloroethyl)ether

Concen: 0.03 ng

RT: 7.11 min Scan# 526

Delta R.T. -0.09 min

Lab File: BE092917.D

Acq: 27 Apr 2017 16:39

Instrument :

BNA_E

ClientSampleId :

KY038PS027-170424


Tgt Ion: 93 Resp: 108

Ion Ratio Lower Upper

93 100

63 0.0 61.2 91.8#




95 0.0 25.5 38.3#



Abundance Ion 93.00 (92.70 to 93.70): BE0

Ion 63.00 (62.70 to 63.70): BE0

Ion 95.00 (94.70 to 95.70): BE0



#7

Naphthalene-d8

Concen: 0.40 ng

RT: 10.53 min Scan# 678

Delta R.T. 0.00 min

Lab File: BE092917.D

Acq: 27 Apr 2017 16:39

Tgt Ion: 136 Resp: 3743


Ion Ratio Lower Upper

136 100

137 11.9 10.4 15.6

54 13.4 12.6 18.8

68 11.9 10.7 16.1






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BE0

Ion 68.00 (67.70 to 68.70): BE0





#8

Nitrobenzene-d5

Concen: 0.32 ng

RT: 8.90 min Scan# 604

Delta R.T. 0.00 min

Lab File: BE092917.D

Acq: 27 Apr 2017 16:39

Instrument :

BNA_E

ClientSampleId :

KY038PS027-170424


Tgt Ion: 82 Resp: 798

Ion Ratio Lower Upper

82 100

128 20.7 35.3 52.9#




54 75.9 65.1 97.7



Abundance Ion 82.00 (81.70 to 82.70): BE0

Ion 128.00 (127.70 to 128.70): BE0

Ion 54.00 (53.70 to 54.70): BE0



#11

2-Methylnaphthalene-d10

Concen: 0.38 ng

RT: 12.13 min Scan# 761

Delta R.T. 0.00 min

Lab File: BE092917.D


Acq: 27 Apr 2017 16:39

Tgt Ion: 152 Resp: 2092

Ion Ratio Lower Upper



152 100


151 19.2 15.7 23.5



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E





#13

Acenaphthene-d10

Concen: 0.40 ng

RT: 14.39 min Scan# 901

Delta R.T. 0.00 min

Lab File: BE092917.D

Acq: 27 Apr 2017 16:39

Instrument :

BNA_E

ClientSampleId :

KY038PS027-170424





Tgt Ion:164 Resp: 1573

Ion Ratio Lower Upper

164 100

162 106.5 86.6 130.0

160 52.6 44.5 66.7



#14

2,4,6-Tribromophenol

Concen: 0.10 ng

RT: 15.87 min Scan# 1010

Delta R.T. 0.00 min

Lab File: BE092917.D

Acq: 27 Apr 2017 16:39




Tgt Ion:330 Resp: 27


Ion Ratio Lower Upper

330 100

332 100.0 69.1 103.7

141 81.5 54.6 81.8









#15
2-Fluorobiphenyl
Concen: 0.44 ng
RT: 13.01 min Scan# 818
Delta R.T. 0.00 min
Lab File: BE092917.D
Acq: 27 Apr 2017 16:39




Instrument :
BNA_E
ClientSampleId :
KY038PS027-170424


Tgt Ion:172 Resp: 3170
Ion Ratio Lower Upper
172 100
171 39.3 35.8 53.6
170 28.7 28.5 42.7



#18
Fluorene
Concen: 0.02 ng
RT: 15.43 min Scan# 972
Delta R.T. -0.01 min
Lab File: BE092917.D
Acq: 27 Apr 2017 16:39

Tgt Ion:166 Resp: 123
Ion Ratio Lower Upper
166 100
165 108.9 80.7 121.1
167 19.5 11.3 16.9#





#19

Phenanthrene-d10

Concen: 0.40 ng

RT: 17.13 min Scan# 1077

Delta R.T. 0.00 min

Lab File: BE092917.D

Acq: 27 Apr 2017 16:39

Instrument :

BNA_E

ClientSampleId :

KY038PS027-170424


Tgt Ion:188 Resp: 3773

Ion Ratio Lower Upper

188 100

94 13.4 10.2 15.4




80 13.7 10.6 15.8



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BE0

Ion 80.00 (79.70 to 80.70): BE0



#20

Hexachlorobenzene

Concen: 0.03 ng

RT: 16.46 min Scan# 1048

Delta R.T. 0.00 min

Lab File: BE092917.D

Acq: 27 Apr 2017 16:39


Tgt Ion:284 Resp: 50

Ion Ratio Lower Upper

284 100

142 0.0 0.0 0.0



249 0.0 0.0 0.0




Abundance Ion 283.80 (283.50 to 284.50): E

Ion 142.00 (141.70 to 142.70): E

Ion 249.00 (248.70 to 249.70): E









#21
 Pentachlorophenol
 Concen: 0.15 ng
 RT: 16.79 min Scan# 1062
 Delta R.T. 0.00 min
 Lab File: BE092917.D
 Acq: 27 Apr 2017 16:39




Instrument :
 BNA_E
 ClientSampleId :
 KY038PS027-170424


Tgt Ion: 266 Resp: 4
 Ion Ratio Lower Upper
 266 100
 264 95.7 75.4 113.2
 268 100.0 75.4 113.2



#22
 Phenanthrene
 Concen: 0.02 ng
 RT: 17.27 min Scan# 1083
 Delta R.T. 0.09 min
 Lab File: BE092917.D
 Acq: 27 Apr 2017 16:39

Tgt Ion: 178 Resp: 254
 Ion Ratio Lower Upper
 178 100
 176 18.5 14.2 21.4
 179 16.1 13.6 20.4





#23

Anthracene

Concen: 0.03 ng

RT: 17.27 min Scan# 1083

Delta R.T. 0.00 min

Lab File: BE092917.D

Acq: 27 Apr 2017 16:39

Instrument :

BNA_E

ClientSampleId :

KY038PS027-170424


Tgt Ion:178 Resp: 283

Ion Ratio Lower Upper

178 100

176 19.1 14.7 22.1




179 16.3 12.8 19.2



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#24

Fluoranthene-d10

Concen: 0.37 ng

RT: 19.15 min Scan# 1165

Delta R.T. 0.00 min

Lab File: BE092917.D

Acq: 27 Apr 2017 16:39


Tgt Ion:212 Resp: 3358

Ion Ratio Lower Upper

212 100

106 0.0 0.0 0.0



104 0.0 0.0 0.0




Abundance Ion 212.00 (211.70 to 212.70): E

Ion 106.00 (105.70 to 106.70): E

Ion 104.00 (103.70 to 104.70): E





#25

Fluoranthene

Concen: 0.03 ng

RT: 19.17 min Scan# 1166

Delta R.T. 0.00 min

Lab File: BE092917.D

Acq: 27 Apr 2017 16:39

Instrument :

BNA_E

ClientSampleId :

KY038PS027-170424


Tgt Ion:202 Resp: 1181

Ion Ratio Lower Upper

202 100

101 0.0 0.0 0.0




203 18.2 15.3 22.9



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#26

Chrysene-d12

Concen: 0.40 ng

RT: 21.28 min Scan# 1268

Delta R.T. 0.00 min

Lab File: BE092917.D

Acq: 27 Apr 2017 16:39


Tgt Ion:240 Resp: 3013

Ion Ratio Lower Upper

240 100

120 4.7 4.6 6.8



236 24.5 19.8 29.8




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E





#27

Pyrene

Concen: 0.03 ng

RT: 19.53 min Scan# 1183

Delta R.T. 0.00 min

Lab File: BE092917.D

Acq: 27 Apr 2017 16:39

Instrument :

BNA_E

ClientSampleId :

KY038PS027-170424


Tgt Ion:202 Resp: 1159

Ion Ratio Lower Upper

202 100

200 0.0 0.0 0.0




203 19.2 13.5 20.3



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E



#28

Terphenyl-d14

Concen: 0.48 ng

RT: 19.74 min Scan# 1193

Delta R.T. 0.00 min

Lab File: BE092917.D

Acq: 27 Apr 2017 16:39


Tgt Ion:244 Resp: 2490

Ion Ratio Lower Upper

244 100

212 12.5 18.4 27.6#



122 16.7 23.3 34.9#




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E





#29

Benzo(a)anthracene

Concen: 0.03 ng

RT: 21.27 min Scan# 1267

Delta R.T. 0.00 min

Lab File: BE092917.D

Acq: 27 Apr 2017 16:39


Instrument :

BNA_E


ClientSampleId :

KY038PS027-170424



Tgt Ion:	228	Resp:	225
Ion	Ratio	Lower	Upper
228	100		
226	45.0	32.0	48.0
229	46.3	21.4	32.0#



Abundance Ion 228.00 (227.70 to 228.70): E
Ion 226.00 (225.70 to 226.70): E
Ion 229.00 (228.70 to 229.70): E



Time-->



#30

Chrysene

Concen: 0.03 ng


RT: 21.32 min Scan# 1270

Delta R.T. 0.00 min


Lab File: BE092917.D

Acq: 27 Apr 2017 16:39


Tgt Ion:	228	Resp:	232
Ion	Ratio	Lower	Upper
228	100		
226	46.2	34.2	51.2
229	44.9	18.6	28.0#




Abundance Ion 228.00 (227.70 to 228.70): E
Ion 226.00 (225.70 to 226.70): E
Ion 229.00 (228.70 to 229.70): E



Time-->









#31
 Bis(2-ethylhexyl)phthalate
 Concen: 0.10 ng
 RT: 21.20 min Scan# 1263
 Delta R.T. 0.00 min
 Lab File: BE092917.D
 Acq: 27 Apr 2017 16:39




Instrument :
 BNA_E
 ClientSampleId :
 KY038PS027-170424


Tgt Ion:149 Resp: 179
 Ion Ratio Lower Upper
 149 100
 167 24.0 20.9 31.3
 279 0.0 0.0 0.0



#32
 Indeno(1,2,3-cd)pyrene
 Concen: 0.03 ng
 RT: 26.22 min Scan# 1650
 Delta R.T. -0.00 min
 Lab File: BE092917.D
 Acq: 27 Apr 2017 16:39

Tgt Ion:276 Resp: 916
 Ion Ratio Lower Upper
 276 100
 138 29.5 25.8 38.6
 277 28.6 20.2 30.4









#33
Perylene-d12
Concen: 0.40 ng
RT: 23.69 min Scan# 1439
Delta R.T. 0.00 min
Lab File: BE092917.D
Acq: 27 Apr 2017 16:39




Instrument :
BNA_E
ClientSampleId :
KY038PS027-170424


Tgt Ion: 264 Resp: 2461
Ion Ratio Lower Upper
264 100
260 28.8 23.5 35.3
265 31.5 27.2 40.8



#34
Benzo(b)fluoranthene
Concen: 0.02 ng
RT: 22.95 min Scan# 1375
Delta R.T. 0.00 min
Lab File: BE092917.D
Acq: 27 Apr 2017 16:39

Tgt Ion: 252 Resp: 213
Ion Ratio Lower Upper
252 100
253 50.0 24.7 37.1#
125 46.5 20.3 30.5#





#35

Benzo(k)fluoranthene

Concen: 0.02 ng

RT: 22.99 min Scan# 1379

Delta R.T. 0.00 min

Lab File: BE092917.D

Acq: 27 Apr 2017 16:39

Instrument :

BNA_E

ClientSampleId :

KY038PS027-170424


Tgt Ion:252 Resp: 198

Ion Ratio Lower Upper

252 100

253 52.9 25.8 38.8#




125 53.5 22.9 34.3#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#36

Benzo(a)pyrene

Concen: 0.03 ng

RT: 23.58 min Scan# 1430

Delta R.T. 0.00 min

Lab File: BE092917.D

Acq: 27 Apr 2017 16:39


Tgt Ion:252 Resp: 175

Ion Ratio Lower Upper

252 100

253 62.5 31.3 46.9#



125 61.7 26.4 39.6#




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#37

Dibenzo(a,h)anthracene

Concen: 0.02 ng

RT: 26.25 min Scan# 1652

Delta R.T. -0.00 min

Lab File: BE092917.D

Acq: 27 Apr 2017 16:39

Instrument :

BNA_E

ClientSampleId :

KY038PS027-170424


Tgt Ion: 278 Resp: 192

Ion Ratio Lower Upper

278 100

139 57.4 27.4 41.2#




279 61.4 33.3 49.9#



Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E



#38

Benzo(g,h,i)perylene

Concen: 0.03 ng

RT: 27.00 min Scan# 1696

Delta R.T. 0.02 min

Lab File: BE092917.D

Acq: 27 Apr 2017 16:39


Tgt Ion: 276 Resp: 863

Ion Ratio Lower Upper

276 100

277 57.1 29.9 44.9#

138 58.5 34.6 51.8#



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

