

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE042717\
 Data File : BE092913.D
 Acq On : 27 Apr 2017 14:13
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 28 04:04:34 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE042017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 20 17:35:10 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
2	1,4-Dioxane	0.768	0.725	5.6	107	0.00
3	n-Nitrosodimethylamine	0.582	0.562	3.4	107	0.00
4 S	2-Fluorophenol	0.870	0.860	1.1	113	0.00
5 S	Phenol-d6	1.213	1.158	4.5	111	0.00
6	bis(2-Chloroethyl)ether	1.474	1.356	8.0	103	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
8 S	Nitrobenzene-d5	0.267	0.253	5.2	104	0.00
9	Naphthalene	1.066	1.032	3.2	105	0.00
10	Hexachlorobutadiene	0.148	0.139	6.1	101	0.00
11 SURR	2-Methylnaphthalene-d10	0.590	0.563	4.6	105	0.00
12	2-Methylnaphthalene	0.658	0.612	7.0	103	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	105	0.00
14 S	2,4,6-Tribromophenol	0.066	0.063	4.5	116	0.00
15 S	2-Fluorobiphenyl	1.838	1.764	4.0	101	0.00
16	Acenaphthylene	6.430	5.521	14.1	97	0.00
17	Acenaphthene	1.263	1.203	4.8	103	-0.02
18	Fluorene	1.541	1.431	7.1	102	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	105	0.00
20	Hexachlorobenzene	0.177	0.162	8.5	99	0.00
21	Pentachlorophenol	0.032	0.050	-56.3#	193#	0.00
22	Phenanthrene	1.159	1.079	6.9	100	0.00
23	Anthracene	0.886	0.762	14.0	96	0.00
24 SURR	Fluoranthene-d10	0.958	0.846	11.7	99	0.00
25	Fluoranthene	4.994	4.430	11.3	101	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	106	0.00
27	Pyrene	5.224	4.788	8.3	101	0.00
28 S	Terphenyl-d14	0.694	0.647	6.8	99	0.00
29	Benzo(a)anthracene	0.931	0.849	8.8	105	0.00
30	Chrysene	1.193	1.144	4.1	107	-0.02
31	Bis(2-ethylhexyl)phthalate	0.241	0.225	6.6	109	0.00
32	Indeno(1,2,3-cd)pyrene	3.910	4.223	-8.0	124	0.00
33 I	Perylene-d12	1.000	1.000	0.0	126	0.00
34	Benzo(b)fluoranthene	1.401	1.177	16.0	110	0.00
35	Benzo(k)fluoranthene	1.427	1.204	15.6	115	0.00
36 C	Benzo(a)pyrene	1.133	1.027	9.4	124	0.00
37	Dibenzo(a,h)anthracene	1.264	1.124	11.1	122	0.00
38	Benzo(g,h,i)perylene	5.484	4.794	12.6	115	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0