

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE042717\
 Data File : BE092920.D
 Acq On : 27 Apr 2017 18:29
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 28 04:06:03 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE042017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 20 17:35:10 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	102	0.00
2	1,4-Dioxane	0.400	0.375	6.3	102	0.00
3	n-Nitrosodimethylamine	0.400	0.391	2.3	105	0.00
4 S	2-Fluorophenol	0.400	0.385	3.8	106	0.00
5 S	Phenol-d6	0.400	0.382	4.5	108	0.00
6	bis(2-Chloroethyl)ether	0.400	0.373	6.8	101	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	102	0.00
8 S	Nitrobenzene-d5	0.400	0.388	3.0	105	0.00
9	Naphthalene	0.400	0.373	6.8	100	0.00
10	Hexachlorobutadiene	0.400	0.376	6.0	99	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.377	5.8	102	0.00
12	2-Methylnaphthalene	0.400	0.372	7.0	102	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	107	0.00
14 S	2,4,6-Tribromophenol	0.400	0.296	26.0#	93	0.00
15 S	2-Fluorobiphenyl	0.400	0.374	6.5	101	0.00
16	Acenaphthylene	0.400	0.347	13.3	100	0.00
17	Acenaphthene	0.400	0.363	9.3	100	-0.02
18	Fluorene	0.400	0.361	9.8	102	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	107	0.00
20	Hexachlorobenzene	0.400	0.350	12.5	97	0.00
21	Pentachlorophenol	0.400	0.454	-13.5	128	0.00
22	Phenanthrene	0.400	0.340	15.0	93	0.00
23	Anthracene	0.400	0.339	15.3	96	0.00
24 SURR	Fluoranthene-d10	0.400	0.362	9.5	104	0.00
25	Fluoranthene	0.400	0.348	13.0	101	0.00
26 I	Chrysene-d12	0.400	0.400	0.0	110	0.00
27	Pyrene	0.400	0.357	10.8	102	0.00
28 S	Terphenyl-d14	0.400	0.375	6.3	104	0.00
29	Benzo(a)anthracene	0.400	0.364	9.0	109	0.00
30	Chrysene	0.400	0.370	7.5	107	0.00
31	Bis(2-ethylhexyl)phthalate	0.400	0.384	4.0	116	0.00
32	Indeno(1,2,3-cd)pyrene	0.400	0.457	-14.2	136	0.00
33 I	Perylene-d12	0.400	0.400	0.0	154	0.00
34	Benzo(b)fluoranthene	0.400	0.291	27.3#	117	0.00
35	Benzo(k)fluoranthene	0.400	0.316	21.0	131	0.00
36 C	Benzo(a)pyrene	0.400	0.313	21.8#	131	0.00
37	Dibenzo(a,h)anthracene	0.400	0.328	18.0	137	0.00
38	Benzo(g,h,i)perylene	0.400	0.326	18.5	131	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1