

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE042717\
 Data File : BE092920.D
 Acq On : 27 Apr 2017 18:29
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 28 04:06:03 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE042017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 20 17:35:10 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2	1,4-Dioxane	0.768	0.720	6.3	102	0.00
3	n-Nitrosodimethylamine	0.582	0.569	2.2	105	0.00
4 S	2-Fluorophenol	0.870	0.838	3.7	106	0.00
5 S	Phenol-d6	1.213	1.159	4.5	108	0.00
6	bis(2-Chloroethyl)ether	1.474	1.374	6.8	101	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
8 S	Nitrobenzene-d5	0.267	0.259	3.0	105	0.00
9	Naphthalene	1.066	0.994	6.8	100	0.00
10	Hexachlorobutadiene	0.148	0.139	6.1	99	0.00
11 SURR	2-Methylnaphthalene-d10	0.590	0.557	5.6	102	0.00
12	2-Methylnaphthalene	0.658	0.612	7.0	102	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	107	0.00
14 S	2,4,6-Tribromophenol	0.066	0.049	25.8#	93	0.00
15 S	2-Fluorobiphenyl	1.838	1.718	6.5	101	0.00
16	Acenaphthylene	6.430	5.572	13.3	100	0.00
17	Acenaphthene	1.263	1.146	9.3	100	-0.02
18	Fluorene	1.541	1.392	9.7	102	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	107	0.00
20	Hexachlorobenzene	0.177	0.155	12.4	97	0.00
21	Pentachlorophenol	0.032	0.032	0.0	128	0.00
22	Phenanthrene	1.159	0.985	15.0	93	0.00
23	Anthracene	0.886	0.750	15.3	96	0.00
24 SURR	Fluoranthene-d10	0.958	0.866	9.6	104	0.00
25	Fluoranthene	4.994	4.345	13.0	101	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	110	0.00
27	Pyrene	5.224	4.668	10.6	102	0.00
28 S	Terphenyl-d14	0.694	0.650	6.3	104	0.00
29	Benzo(a)anthracene	0.931	0.847	9.0	109	0.00
30	Chrysene	1.193	1.105	7.4	107	0.00
31	Bis(2-ethylhexyl)phthalate	0.241	0.231	4.1	116	0.00
32	Indeno(1,2,3-cd)pyrene	3.910	4.470	-14.3	136	0.00
33 I	Perylene-d12	1.000	1.000	0.0	154#	0.00
34	Benzo(b)fluoranthene	1.401	1.019	27.3#	117	0.00
35	Benzo(k)fluoranthene	1.427	1.126	21.1	131	0.00
36 C	Benzo(a)pyrene	1.133	0.887	21.7#	131	0.00
37	Dibenzo(a,h)anthracene	1.264	1.035	18.1	137	0.00
38	Benzo(g,h,i)perylene	5.484	4.465	18.6	131	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1