

Data Path : Z:\HPCHEM1\BNA E\DATA\BE043018\
 Data File : BE096225.D
 Acq On : 30 Apr 2018 17:09
 Operator : SJ/JU
 Sample : J2463-10RE
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 LF001MW001-180418RE

Quant Time: May 01 04:49:41 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE041118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 30 14:27:29 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.36	152	801	0.40	ng	0.00
7) Naphthalene-d8	11.19	136	2835	0.40	ng	0.00
13) Acenaphthene-d10	14.94	164	8993	0.40	ng	0.00
19) Phenanthrene-d10	17.65	188	3809	0.40	ng	0.00
27) Chrysene-d12	21.75	240	4049	0.40	ng	0.01
34) Perylene-d12	24.36	264	3854	0.40	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.87	112	558	0.23	ng	0.01
5) Phenol-d6	7.53	99	599	0.18	ng	0.02
8) Nitrobenzene-d5	9.54	82	2518	0.76	ng	0.00
11) 2-Methylnaphthalene-d10	12.74	152	1782	0.40	ng	0.00
14) 2,4,6-Tribromophenol	16.43	330	372	0.09	ng	0.01
15) 2-Fluorobiphenyl	13.59	172	2499	0.06	ng	0.01
25) Fluoranthene-d10	19.64	212	17804	0.34	ng	0.00
29) Terphenyl-d14	20.19	244	2595	0.29	ng	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) Naphthalene	11.26	128	3165229	105.905	ng	99
12) 2-Methylnaphthalene	12.81	142	77383	15.643	ng	94
18) Fluorene	15.97	166	1660	0.046	ng	# 81
23) Phenanthrene	17.70	178	792	0.074	ng	# 61
28) Pyrene	19.98	202	484	0.061	ng	# 75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE043018\
Data File : BE096225.D
Acq On : 30 Apr 2018 17:09
Operator : SJ/JU
Sample : J2463-10RE
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
LF001MW001-180418RE

Quant Time: May 01 04:49:41 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE041118.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 30 14:27:29 2018
Response via : Initial Calibration

