

Data Path : Z:\HPCHEM1\BNA E\Data\BE043018\
 Data File : BE096231.D
 Acq On : 30 Apr 2018 20:47
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 01 05:29:40 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE041118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 30 14:27:29 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	54	0.00
2	1,4-Dioxane	0.400	0.345	13.8	50	0.00
3	n-Nitrosodimethylamine	0.400	0.419	-4.7	62	0.00
4 S	2-Fluorophenol	0.400	0.379	5.3	54	0.00
5 S	Phenol-d6	0.400	0.386	3.5	55	0.00
6	bis(2-Chloroethyl)ether	0.400	0.367	8.3	52	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	51	0.00
8 S	Nitrobenzene-d5	0.400	0.405	-1.3	56	0.00
9	Naphthalene	0.400	0.385	3.8	52	0.00
10	Hexachlorobutadiene	0.400	0.307	23.3	43	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.376	6.0	51	0.00
12	2-Methylnaphthalene	0.400	0.379	5.3	52	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	53	0.00
14 S	2,4,6-Tribromophenol	0.400	0.393	1.8	57	0.00
15 S	2-Fluorobiphenyl	0.400	0.378	5.5	52	0.00
16	Acenaphthylene	0.400	0.378	5.5	54	0.00
17	Acenaphthene	0.400	0.375	6.3	54	0.00
18	Fluorene	0.400	0.470	-17.5	67	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	56	0.00
20	4-Bromophenyl-phenylether	0.400	0.363	9.3	54	0.00
21	Hexachlorobenzene	0.400	0.365	8.8	53	0.00
22	Pentachlorophenol	0.400	0.347	13.3	51	0.00
23	Phenanthrene	0.400	0.378	5.5	56	0.00
24	Anthracene	0.400	0.378	5.5	57	0.00
25 SURR	Fluoranthene-d10	0.400	0.377	5.8	57	0.00
26	Fluoranthene	0.400	0.384	4.0	58	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	52	0.01
28	Pvrene	0.400	0.033	91.8#	5	0.00
29 S	Terphenyl-d14	0.400	0.392	2.0	56	0.00
30	Benzo(a)anthracene	0.400	0.391	2.3	55	0.00
31	Chrysene	0.400	0.393	1.8	54	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.414	-3.5	58	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.417	-4.2	57	0.00
34 I	Perylene-d12	0.400	0.400	0.0	52	0.00
35	Benzo(b)fluoranthene	0.400	0.390	2.5	53	0.00
36	Benzo(k)fluoranthene	0.400	0.393	1.8	55	0.00
37 C	Benzo(a)pyrene	0.400	0.385	3.8	53	0.00
38	Dibenzo(a,h)anthracene	0.400	0.417	-4.2	59	0.00
39	Benzo(a,h,i)perylene	0.400	0.399	0.3	55	0.00

Data Path : Z:\HPCHEM1\BNA E\Data\BE043018\
Data File : BE096231.D
Acq On : 30 Apr 2018 20:47
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 01 05:29:40 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE041118.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 30 14:27:29 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				