

Data Path : Z:\HPCHEM1\BNA E\Data\BE043018\
 Data File : BE096231.D
 Acq On : 30 Apr 2018 20:47
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 01 05:29:40 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE041118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 30 14:27:29 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	54	0.00
2	1,4-Dioxane	0.618	0.533	13.8	50	0.00
3	n-Nitrosodimethylamine	0.757	0.793	-4.8	62	0.00
4 S	2-Fluorophenol	1.237	1.173	5.2	54	0.00
5 S	Phenol-d6	1.707	1.648	3.5	55	0.00
6	bis(2-Chloroethyl)ether	1.502	1.377	8.3	52	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	51	0.00
8 S	Nitrobenzene-d5	0.469	0.475	-1.3	56	0.00
9	Naphthalene	4.217	4.055	3.8	52	0.00
10	Hexachlorobutadiene	0.187	0.143	23.5	43#	0.00
11 SURR	2-Methylnaphthalene-d10	0.634	0.596	6.0	51	0.00
12	2-Methylnaphthalene	0.698	0.660	5.4	52	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	53	0.00
14 S	2,4,6-Tribromophenol	0.193	0.190	1.6	57	0.00
15 S	2-Fluorobiphenyl	1.711	1.618	5.4	52	0.00
16	Acenaphthylene	2.136	2.018	5.5	54	0.00
17	Acenaphthene	1.284	1.204	6.2	54	0.00
18	Fluorene	1.588	1.868	-17.6	67	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	56	0.00
20	4-Bromophenyl-phenylether	0.238	0.216	9.2	54	0.00
21	Hexachlorobenzene	0.237	0.216	8.9	53	0.00
22	Pentachlorophenol	0.078	0.058	25.6#	51	0.00
23	Phenanthrene	1.124	1.061	5.6	56	0.00
24	Anthracene	1.074	1.014	5.6	57	0.00
25 SURR	Fluoranthene-d10	5.430	5.115	5.8	57	0.00
26	Fluoranthene	1.477	1.417	4.1	58	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	52	0.01
28	Pvrene	0.788	0.064	91.9#	5#	0.00
29 S	Terphenyl-d14	0.884	0.867	1.9	56	0.00
30	Benzo(a)anthracene	1.283	1.253	2.3	55	0.00
31	Chrysene	1.239	1.217	1.8	54	0.00
32	Bis(2-ethylhexyl)phthalate	3.311	3.427	-3.5	58	0.00
33	Indeno(1,2,3-cd)pyrene	1.201	1.253	-4.3	57	0.00
34 I	Perylene-d12	1.000	1.000	0.0	52	0.00
35	Benzo(b)fluoranthene	1.200	1.168	2.7	53	0.00
36	Benzo(k)fluoranthene	1.234	1.214	1.6	55	0.00
37 C	Benzo(a)pyrene	1.153	1.111	3.6	53	0.00
38	Dibenzo(a,h)anthracene	1.017	1.060	-4.2	59	0.00
39	Benzo(a,h,i)perylene	1.076	1.072	0.4	55	0.00

Data Path : Z:\HPCHEM1\BNA E\Data\BE043018\
Data File : BE096231.D
Acq On : 30 Apr 2018 20:47
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 01 05:29:40 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE041118.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 30 14:27:29 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			