

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE043019\
 Data File : BE099433.D
 Acq On : 30 Apr 2019 19:11
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 30 22:45:17 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE042919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 29 18:36:43 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	-0.01
2	1,4-Dioxane	0.550	0.547	0.5	121	0.00
3	n-Nitrosodimethylamine	0.332	0.294	11.4	105	-0.01
4 S	2-Fluorophenol	1.106	1.138	-2.9	113	0.00
5 S	Phenol-d6	1.471	1.625	-10.5	118	-0.01
6	bis(2-Chloroethyl)ether	1.269	1.235	2.7	104	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	111	0.00
8 S	Nitrobenzene-d5	0.375	0.384	-2.4	113	0.00
9	Naphthalene	4.359	4.510	-3.5	113	0.00
10	Hexachlorobutadiene	0.303	0.313	-3.3	112	0.00
11 SURR	2-Methylnaphthalene-d10	0.721	0.775	-7.5	120	0.00
12	2-Methylnaphthalene	0.758	0.823	-8.6	121	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	126	0.00
14 S	2,4,6-Tribromophenol	0.252	0.264	-4.8	136	-0.01
15 S	2-Fluorobiphenyl	1.526	1.491	2.3	120	-0.02
16	Acenaphthylene	1.957	1.924	1.7	123	-0.02
17	Acenaphthene	1.118	1.132	-1.3	128	0.00
18	Fluorene	1.649	1.696	-2.9	128	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	131	0.00
20	4-Bromophenyl-phenylether	0.233	0.230	1.3	128	0.00
21	Hexachlorobenzene	0.222	0.226	-1.8	132	0.00
22	Pentachlorophenol	0.100	0.109	-9.0	169#	0.00
23	Phenanthrene	1.032	1.022	1.0	128	0.00
24	Anthracene	0.984	0.972	1.2	130	0.00
25 SURR	Fluoranthene-d10	5.391	5.172	4.1	125	0.00
26	Fluoranthene	1.546	1.506	2.6	126	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	114	0.00
28	Pyrene	1.043	1.292	-23.9	125	0.00
29 S	Terphenyl-d14	0.733	0.874	-19.2	121	0.00
30	Benzo(a)anthracene	1.274	1.396	-9.6	111	-0.01
31	Chrysene	1.247	1.409	-13.0	113	0.00
32	Bis(2-ethylhexyl)phthalate	2.216	2.151	2.9	102	-0.01
33	Indeno(1,2,3-cd)pyrene	1.474	1.534	-4.1	104	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	105	0.00
35	Benzo(b)fluoranthene	1.153	1.224	-6.2	110	-0.01
36	Benzo(k)fluoranthene	1.112	1.163	-4.6	104	-0.01
37 C	Benzo(a)pyrene	1.090	1.139	-4.5	106	-0.01
38	Dibenzo(a,h)anthracene	1.115	1.150	-3.1	105	-0.01
39	Benzo(g,h,i)perylene	1.118	1.138	-1.8	103	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE043019\
Data File : BE099433.D
Acq On : 30 Apr 2019 19:11
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 30 22:45:17 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE042919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 29 18:36:43 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			