

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE043019\
 Data File : BE099435.D
 Acq On : 30 Apr 2019 21:02
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 30 22:46:57 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE042919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 29 18:36:43 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	108	-0.01
2	1,4-Dioxane	0.400	0.377	5.8	118	0.00
3	n-Nitrosodimethylamine	0.400	0.352	12.0	107	-0.01
4 S	2-Fluorophenol	0.400	0.417	-4.2	118	0.00
5 S	Phenol-d6	0.400	0.424	-6.0	117	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.399	0.3	110	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	114	0.00
8 S	Nitrobenzene-d5	0.400	0.401	-0.3	114	0.00
9	Naphthalene	0.400	0.414	-3.5	116	0.00
10	Hexachlorobutadiene	0.400	0.404	-1.0	112	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.411	-2.7	118	0.00
12	2-Methylnaphthalene	0.400	0.424	-6.0	121	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	123	0.00
14 S	2,4,6-Tribromophenol	0.400	0.413	-3.2	131	-0.01
15 S	2-Fluorobiphenyl	0.400	0.399	0.3	120	-0.02
16	Acenaphthylene	0.400	0.400	0.0	122	-0.02
17	Acenaphthene	0.400	0.410	-2.5	126	0.00
18	Fluorene	0.400	0.419	-4.7	127	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	128	0.00
20	4-Bromophenyl-phenylether	0.400	0.396	1.0	125	0.00
21	Hexachlorobenzene	0.400	0.410	-2.5	130	0.00
22	Pentachlorophenol	0.400	0.516	-29.0#	152	0.00
23	Phenanthrene	0.400	0.407	-1.7	129	0.00
24	Anthracene	0.400	0.398	0.5	128	0.00
25 SURR	Fluoranthene-d10	0.400	0.379	5.3	121	0.00
26	Fluoranthene	0.400	0.387	3.3	123	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	110	0.00
28	Pyrene	0.400	0.493	-23.2	120	0.00
29 S	Terphenyl-d14	0.400	0.485	-21.2	119	0.00
30	Benzo(a)anthracene	0.400	0.438	-9.5	107	-0.01
31	Chrysene	0.400	0.468	-17.0	113	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.371	7.3	91	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.424	-6.0	103	-0.02
34 I	Perylene-d12	0.400	0.400	0.0	104	0.00
35	Benzo(b)fluoranthene	0.400	0.419	-4.7	107	0.00
36	Benzo(k)fluoranthene	0.400	0.425	-6.2	105	-0.01
37 C	Benzo(a)pyrene	0.400	0.417	-4.2	105	0.00
38	Dibenzo(a,h)anthracene	0.400	0.409	-2.2	103	-0.01
39	Benzo(g,h,i)perylene	0.400	0.411	-2.7	103	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE043019\
Data File : BE099435.D
Acq On : 30 Apr 2019 21:02
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 30 22:46:57 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE042919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 29 18:36:43 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				