

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE043019\
 Data File : BE099435.D
 Acq On : 30 Apr 2019 21:02
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 30 22:46:57 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE042919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 29 18:36:43 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	108	-0.01
2	1,4-Dioxane	0.550	0.519	5.6	118	0.00
3	n-Nitrosodimethylamine	0.332	0.292	12.0	107	-0.01
4 S	2-Fluorophenol	1.106	1.154	-4.3	118	0.00
5 S	Phenol-d6	1.471	1.560	-6.1	117	-0.01
6	bis(2-Chloroethyl)ether	1.269	1.265	0.3	110	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	114	0.00
8 S	Nitrobenzene-d5	0.375	0.375	0.0	114	0.00
9	Naphthalene	4.359	4.507	-3.4	116	0.00
10	Hexachlorobutadiene	0.303	0.306	-1.0	112	0.00
11 SURR	2-Methylnaphthalene-d10	0.721	0.741	-2.8	118	0.00
12	2-Methylnaphthalene	0.758	0.804	-6.1	121	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	123	0.00
14 S	2,4,6-Tribromophenol	0.252	0.260	-3.2	131	-0.01
15 S	2-Fluorobiphenyl	1.526	1.524	0.1	120	-0.02
16	Acenaphthylene	1.957	1.959	-0.1	122	-0.02
17	Acenaphthene	1.118	1.146	-2.5	126	0.00
18	Fluorene	1.649	1.726	-4.7	127	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	128	0.00
20	4-Bromophenyl-phenylether	0.233	0.231	0.9	125	0.00
21	Hexachlorobenzene	0.222	0.227	-2.3	130	0.00
22	Pentachlorophenol	0.100	0.100	0.0	152#	0.00
23	Phenanthrene	1.032	1.049	-1.6	129	0.00
24	Anthracene	0.984	0.980	0.4	128	0.00
25 SURR	Fluoranthene-d10	5.391	5.102	5.4	121	0.00
26	Fluoranthene	1.546	1.494	3.4	123	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	110	0.00
28	Pyrene	1.043	1.286	-23.3	120	0.00
29 S	Terphenyl-d14	0.733	0.888	-21.1	119	0.00
30	Benzo(a)anthracene	1.274	1.394	-9.4	107	-0.01
31	Chrysene	1.247	1.460	-17.1	113	0.00
32	Bis(2-ethylhexyl)phthalate	2.216	1.988	10.3	91	-0.01
33	Indeno(1,2,3-cd)pyrene	1.474	1.563	-6.0	103	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	104	0.00
35	Benzo(b)fluoranthene	1.153	1.208	-4.8	107	0.00
36	Benzo(k)fluoranthene	1.112	1.183	-6.4	105	-0.01
37 C	Benzo(a)pyrene	1.090	1.137	-4.3	105	0.00
38	Dibenzo(a,h)anthracene	1.115	1.142	-2.4	103	-0.01
39	Benzo(g,h,i)perylene	1.118	1.147	-2.6	103	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE043019\
Data File : BE099435.D
Acq On : 30 Apr 2019 21:02
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 30 22:46:57 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE042919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 29 18:36:43 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		