

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE043019\
 Data File : BE099452.D
 Acq On : 1 May 2019 7:14
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 01 07:56:51 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE042919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 29 18:36:43 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	114	0.00
2	1,4-Dioxane	0.550	0.506	8.0	121	0.00
3	n-Nitrosodimethylamine	0.332	0.320	3.6	124	-0.01
4 S	2-Fluorophenol	1.106	1.195	-8.0	128	0.00
5 S	Phenol-d6	1.471	1.821	-23.8	144	0.00
6	bis(2-Chloroethyl)ether	1.269	1.270	-0.1	116	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	135	0.00
8 S	Nitrobenzene-d5	0.375	0.384	-2.4	137	0.00
9	Naphthalene	4.359	4.501	-3.3	137	0.00
10	Hexachlorobutadiene	0.303	0.299	1.3	130	0.00
11 SURR	2-Methylnaphthalene-d10	0.721	0.750	-4.0	142	0.00
12	2-Methylnaphthalene	0.758	0.807	-6.5	144	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	145	0.00
14 S	2,4,6-Tribromophenol	0.252	0.268	-6.3	160#	0.00
15 S	2-Fluorobiphenyl	1.526	1.471	3.6	137	-0.01
16	Acenaphthylene	1.957	1.998	-2.1	147	0.00
17	Acenaphthene	1.118	1.150	-2.9	150	0.00
18	Fluorene	1.649	1.678	-1.8	146	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	138	0.00
20	4-Bromophenyl-phenylether	0.233	0.237	-1.7	138	0.00
21	Hexachlorobenzene	0.222	0.238	-7.2	147	0.00
22	Pentachlorophenol	0.100	0.113	-13.0	185#	0.00
23	Phenanthrene	1.032	1.042	-1.0	137	0.00
24	Anthracene	0.984	1.024	-4.1	144	0.00
25 SURR	Fluoranthene-d10	5.391	5.135	4.7	131	0.00
26	Fluoranthene	1.546	1.474	4.7	130	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	117	0.00
28	Pyrene	1.043	1.298	-24.4	129	0.00
29 S	Terphenyl-d14	0.733	0.879	-19.9	125	0.00
30	Benzo(a)anthracene	1.274	1.472	-15.5	120	-0.01
31	Chrysene	1.247	1.382	-10.8	114	0.00
32	Bis(2-ethylhexyl)phthalate	2.216	2.302	-3.9	112	0.00
33	Indeno(1,2,3-cd)pyrene	1.474	1.604	-8.8	112	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	111	0.00
35	Benzo(b)fluoranthene	1.153	1.222	-6.0	116	0.00
36	Benzo(k)fluoranthene	1.112	1.154	-3.8	109	0.00
37 C	Benzo(a)pyrene	1.090	1.135	-4.1	112	0.00
38	Dibenzo(a,h)anthracene	1.115	1.170	-4.9	113	0.00
39	Benzo(g,h,i)perylene	1.118	1.168	-4.5	112	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE043019\
Data File : BE099452.D
Acq On : 1 May 2019 7:14
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 01 07:56:51 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE042919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 29 18:36:43 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		