


Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE043019\
Data File : BE099415.D
Acq On : 30 Apr 2019 8:20
Operator : JU/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE042919.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Mon Apr 29 18:36:43 2019



AutoFind: Scans 2559, 2560, 2561; Background Corrected with Scan 2550

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	22.0	19572	PASS
68	69	0.00	2	1.0	295	PASS
69	198	0.00	100	32.4	28798	PASS
70	69	0.00	2	0.6	170	PASS
127	198	10	80	34.3	30503	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	88915	PASS
199	198	5	9	6.6	5839	PASS
275	198	10	60	31.0	27536	PASS
365	198	1	100	3.8	3384	PASS
441	198	0.01	100	14.1	12506	PASS
442	198	50	100	85.2	75712	PASS
443	442	15	24	20.3	15356	PASS