

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE050118\
 Data File : BE096282.D
 Acq On : 2 May 2018 6:12
 Operator : SJ/JU
 Sample : J2641-10
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 BPS1-DUP03-20180426

Quant Time: May 02 11:31:06 2018
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE041118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 30 14:27:29 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.36	152	909	0.40	ng	0.00
7) Naphthalene-d8	11.19	136	3427	0.40	ng	0.00
13) Acenaphthene-d10	14.94	164	1791	0.40	ng	0.00
19) Phenanthrene-d10	17.65	188	4038	0.40	ng	0.00
27) Chrysene-d12	21.74	240	3950	0.40	ng	0.00
34) Perylene-d12	24.35	264	3752	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.88	112	321	0.11	ng	0.01
5) Phenol-d6	7.53	99	258	0.07	ng	0.02
8) Nitrobenzene-d5	9.54	82	1856	0.46	ng	0.00
11) 2-Methylnaphthalene-d10	12.74	152	1707	0.31	ng	0.00
14) 2,4,6-Tribromophenol	16.41	330	180	0.21	ng	0.00
15) 2-Fluorobiphenyl	13.57	172	3796	0.50	ng	0.00
25) Fluoranthene-d10	19.63	212	17634	0.32	ng	0.00
29) Terphenyl-d14	20.19	244	4349	0.50	ng	0.00
Target Compounds						
26) Fluoranthene	19.66	202	321	0.022	ng	# 95
32) Bis(2-ethylhexyl)phthalate	21.60	149	938	0.029	ng	# 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE050118\
Data File : BE096282.D
Acq On : 2 May 2018 6:12
Operator : SJ/JU
Sample : J2641-10
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
BPS1-DUP03-20180426

Quant Time: May 02 11:31:06 2018
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE041118.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 30 14:27:29 2018
Response via : Initial Calibration

