

Data Path : Z:\HPCHEM1\BNA_E\Data\BE050118\
 Data File : BE096267.D
 Acq On : 1 May 2018 21:06
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 02 03:47:05 2018
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE041118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 30 14:27:29 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	66	0.00
2	1,4-Dioxane	0.618	0.504	18.4	58	0.00
3	n-Nitrosodimethylamine	0.757	0.785	-3.7	75	0.00
4 S	2-Fluorophenol	1.237	1.231	0.5	70	0.01
5 S	Phenol-d6	1.707	1.735	-1.6	71	0.01
6	bis(2-Chloroethyl)ether	1.502	1.481	1.4	68	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	66	0.00
8 S	Nitrobenzene-d5	0.469	0.475	-1.3	73	0.00
9	Naphthalene	4.217	3.974	5.8	66	0.00
10	Hexachlorobutadiene	0.187	0.148	20.9	58	0.00
11 SURR	2-Methylnaphthalene-d10	0.634	0.615	3.0	69	0.00
12	2-Methylnaphthalene	0.698	0.672	3.7	69	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	70	0.00
14 S	2,4,6-Tribromophenol	0.193	0.198	-2.6	78	0.00
15 S	2-Fluorobiphenyl	1.711	1.597	6.7	68	0.00
16	Acenaphthylene	2.136	2.038	4.6	72	0.00
17	Acenaphthene	1.284	1.219	5.1	72	0.00
18	Fluorene	1.588	1.893	-19.2	89	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	70	0.00
20	4-Bromophenyl-phenylether	0.238	0.227	4.6	71	0.00
21	Hexachlorobenzene	0.237	0.221	6.8	67	-0.01
22	Pentachlorophenol	0.078	0.096	-23.1	107	0.00
23	Phenanthrene	1.124	1.032	8.2	68	0.00
24	Anthracene	1.074	1.022	4.8	72	0.00
25 SURR	Fluoranthene-d10	5.430	4.804	11.5	67	0.00
26	Fluoranthene	1.477	1.300	12.0	66	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	60	0.00
28	Pyrene	0.788	0.084	89.3#	7#	0.00
29 S	Terphenyl-d14	0.884	0.894	-1.1	66	0.00
30	Benzo(a)anthracene	1.283	1.223	4.7	62	0.00
31	Chrysene	1.239	1.191	3.9	61	0.00
32	Bis(2-ethylhexyl)phthalate	3.311	3.809	-15.0	74	0.00
33	Indeno(1,2,3-cd)pyrene	1.201	1.221	-1.7	65	0.00
34 I	Perylene-d12	1.000	1.000	0.0	61	0.00
35	Benzo(b)fluoranthene	1.200	1.127	6.1	60	0.00
36	Benzo(k)fluoranthene	1.234	1.154	6.5	61	0.00
37 C	Benzo(a)pyrene	1.153	1.091	5.4	61	0.00
38	Dibenzo(a,h)anthracene	1.017	1.049	-3.1	68	0.00
39	Benzo(g,h,i)perylene	1.076	1.053	2.1	63	0.00

Data Path : Z:\HPCHEM1\BNA_E\Data\BE050118\
Data File : BE096267.D
Acq On : 1 May 2018 21:06
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 02 03:47:05 2018
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE041118.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 30 14:27:29 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			