

Data Path : Z:\HPCHEM1\BNA E\DATA\BE050218\
 Data File : BE096298.D
 Acq On : 2 May 2018 17:43
 Operator : SJ/JU
 Sample : J2507-06RE
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SS043MW009-180419RE

Manual Integrations
 APPROVED

Sohil
 5/3/2018 4:28:00 PM

Quant Time: May 03 02:26:17 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE041118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 30 14:27:29 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.36	152	1058	0.40	ng	0.00
7) Naphthalene-d8	11.19	136	3732	0.40	ng	0.00
13) Acenaphthene-d10	14.93	164	5576	0.40	ng	-0.01
19) Phenanthrene-d10	17.65	188	4913	0.40	ng	0.00
27) Chrysene-d12	21.75	240	5097	0.40	ng	0.01
34) Perylene-d12	24.35	264	4467	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.87	112	456	0.14	ng	0.00
5) Phenol-d6	7.53	99	114	0.03	ng	0.02
8) Nitrobenzene-d5	9.54	82	1005	0.23	ng	0.00
11) 2-Methylnaphthalene-d10	12.73	152	2472	0.42	ng	0.00
14) 2,4,6-Tribromophenol	16.41	330	507	0.19	ng	0.00
15) 2-Fluorobiphenyl	13.57	172	4120	0.17	ng	0.00
25) Fluoranthene-d10	19.64	212	21959	0.33	ng	0.00
29) Terphenyl-d14	20.19	244	5734	0.51	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.59	88	2741	1.677	ng	# 55
6) bis(2-Chloroethyl)ether	7.76	93	10294	2.590	ng	# 51
9) Naphthalene	11.24	128	67060	1.704	ng	# 86
18) Fluorene	15.97	166	2748	0.124	ng	# 37
22) Pentachlorophenol	17.31	266	169	0.262	ng	# 86
23) Phenanthrene	17.68	178	1844	0.134	ng	# 61
24) Anthracene	17.78	178	920	0.070	ng	# 79
26) Fluoranthene	19.66	202	972	0.054	ng	# 1
32) Bis(2-ethylhexyl)phthalate	21.60	149	5616m	0.133	ng	#
33) Indeno(1,2,3-cd)pyrene	27.13	276	521	0.034	ng	# 70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE050218\
Data File : BE096298.D
Acq On : 2 May 2018 17:43
Operator : SJ/JU
Sample : J2507-06RE
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SS043MW009-180419RE

Manual Integrations
APPROVED

Sohil
5/3/2018 4:28:00 PM

Quant Time: May 03 02:26:17 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE041118.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 30 14:27:29 2018
Response via : Initial Calibration

